

The most current laboratory reference ranges are included in the laboratory report from the LIS.
 Laboratory testing information is also available on The Ohio State Wexner Medical Center Labs Test Catalog Website. <https://theohiostatelabs.testcatalog.org/>

This document applies all laboratory testing at OSU/WHC Clinical Laboratories. This includes:

James Genetics Laboratories: 5001 Polaris Pkwy, Innovations Centre, Suite 1500, Columbus OH 43240

Microbiology Laboratory: 1st Fl Microbiology Medical Plaza Tower, 2650 Kenny Rd, Columbus OH 43221

Spelman Laboratory: 1145 Olentangy River Rd Rm 2030, Columbus, Oh 43212

Clinical Laboratories (LH): 410 West 10th Avenue, Columbus OH 43210

Clinical Laboratories (LH): 411 1st Taylor Avenue, Columbus OH 43203

James West Campus Laboratory (JWC): 2121 Kenny Rd, Columbus, OH 43221

Test Name	Synonym	Reference Values	Reference Values	Interpretation	Method Description	Source of Reference Range	Technical Range / AMR	Reportable Range / CRR
		Female	Male					
		PH 0-29 days: 7.29-7.45 30+ days: 7.35-7.45	PH 0-29 days: 7.29-7.45 30+ days: 7.35-7.45					
		PCO2 0-29 days: 27-40 mm Hg 30+ days: 32-48 mm Hg	PCO2 0-29 days: 27-40 mm Hg 30+ days: 32-48 mm Hg					
		PO2 83-108 mm Hg	PO2 83-108 mm Hg					
		HCO3 0-29 days: 17-24 mmol/L 30+ days: 22-28 mmol/L	HCO3 0-29 days: 17-24 mmol/L 30+ days: 22-28 mmol/L					
		OSAT 0-365 days: 40-90 % 1+ years: 94-98 %	OSAT 0-365 days: 40-90 % 1+ years: 94-98 %			pH: Clinical Guide to Laboratory Tests 3rd Edition, Tietz, 1995. HCO3 Arterial: Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications. 3rd Edition, 2021 (p-3) 2012068.		
		BASE -3.0-3.0 mmol/L	BASE -3.0-3.0 mmol/L			aO2 Arterial: Clinical Guide to Laboratory Tests 3rd Edition, Tietz, 1995.		
		GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL	GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL			aO2 Arterial: Clinical Guide to Laboratory Tests 3rd Edition, Tietz, 1995.		
		LACT 0.5-1.6 mmol/L	LACT 0.5-1.6 mmol/L			pO2 Arterial: Clinical Guide to Laboratory Tests, Tietz, 1995 & Fundamentals of Clinical Chem, 1987.		
		NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L	NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L			pCO2 Arterial: Clinical Guide to Laboratory Tests, Tietz, 1995 & Fundamentals of Clinical Chem, 1987.		
		K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+17 years: 3.4-6.0 mmol/L 18+ years: 3.5-5.0 mmol/L	K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+17 years: 3.4-6.0 mmol/L 18+ years: 3.5-5.0 mmol/L			Base excess: Contemporary Practice in Clinical Chemistry 3rd Edition 2016. Chapter 32. Table 32.1 p450.		
		ICA 4.60-5.30 mg/dL	ICA 4.60-5.30 mg/dL			Glucose: Clinical Guide to Laboratory Tests, 3rd Edition Tietz, 1995; Pediatric Reference Ranges, Solkin, 1999.		
		BGHGB 0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 11.4-15.2 g/dL	BGHGB 0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 11.4-16.8 g/dL			Lactate: ABL 800 Flex Reference Manual, 2008, verified by GEM Validation Study, 2021.		
		BGHCT (Calculated) - Radiometer 0-14 days: 40-57 % 15-30 days: 32-45 % 31-60 days: 28-35 % 61-179 days: 30-37 % 180-2 years: 31-38 % 2-5 years: 31-38 % 6-11 years: 32-40 % 12-17 years: 33-40 % 18+ years: 34-46 %	BGHCT (Calculated) - Radiometer 0-14 days: 40-54 % 15-30 days: 31-45 % 31-60 days: 27-38 % 61-179 days: 29-37 % 180-2 years: 31-38 % 2-5 years: 31-38 % 6-11 years: 32-40 % 12-17 years: 34-44 % 18+ years: 40-50 %			Na, Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications. 3rd Edition, 2021 (p.165). Clinical Guide to Laboratory Tests Third Edition, Tietz, 1995. Lower limit adjusted and verified with GEM Validation Study, 2021.		
		BGHCT (Calculated) - GEM 5000/7000 0-15 days: 40-60 % 15-30 days: 32-44 % 30-60 days: 28-34 % 60-180 days: 30-37 % 180-2 years: 31-38 % 2-6 years: 31-38 % 6-12 years: 32-40 % 12-18 years: 32-40 % 18+ years: 34-46 %	BGHCT (Calculated) - GEM 5000/7000 0-15 days: 42-57 % 15-30 days: 30-46 % 30-60 days: 27-38 % 60-180 days: 29-37 % 180-2 years: 30-38 % 2-6 years: 31-38 % 6-12 years: 32-40 % 12-18 years: 33-44 % 18+ years: 40-50 %			Total IGB: Co-oximetry: TIB + O2IB + COB + MetHb + IHB aO2: Co-oximetry: aO2 - 100 x O2IB% (O2IB% + IHB%) Base excess GEM: HCO3 - 24.8 + 16.2 x (pH - 7.4)		
		O2HGB 94-98 %	O2HGB 94-98 %			K: Clinical Guide to Laboratory Tests, Tietz 3rd Edition, 1995. Clinical Guide to Laboratory Tests, Tietz, 1995. Upper limit adjusted and verified with GEM Validation Study, 2021.		
		COHGB ≤1.5 %	COHGB ≤1.5 %			K: Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications. 3rd Edition, 2021 (p.102).		
		MTHGB ≤1.5 %	MTHGB ≤1.5 %			HCT: Derived from total hemoglobin reference interval. ABL 800 FLEX Reference Manual, 2008. Derived from total hemoglobin reference interval based on GEM calculation of HCT.		
						ICB: OSU Internal Normal Range Study, October 2018. Verified with GEM Validation Study, 2021.		
						O2HGB, COHGB, MTHGB: ABL 800 Flex Reference Manual 2008, verified by GEM Validation Study, 2021. Clinical Guide to Laboratory Tests Third Edition, Tietz, 1995. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics Sixth Edition, 2018.		

Arterial Blood Gas	GAS5	<p>PH 0-29 days: 7.29-7.45 30+ days: 7.35-7.45</p> <p>PCO2 0-29 days: 27-40 mm Hg 30+ days: 32-48 mm Hg</p> <p>PO2 83-108 mm Hg</p> <p>HCO3 0-29 days: 17-24 mmol/L 30+ days: 22-28 mmol/L</p> <p>OSAT 0-365 days: 40-90 % 1+ years: 94-98 %</p> <p>BASE -3.0-3.0 mmol/L</p>		<p>pO2: Amperometry</p> <p>pH, pCO2: Potentiometry</p> <p>Base excess GEM: $HCO_3 - 24.8 + 16.2 \times (pH - 7.4)$</p> <p>HCO3 GEM: $\text{Log}(HCO_3 (c)) - pH + \text{logpCO}_2 - 7.608$ mmol/L</p> <p>sO2: Co-oximetry; $sO_2 - 100 \times O_2Hb\% / (O_2Hb\% + HbB\%)$</p>	See Arterial Blood Gas (Full Panel)	See Arterial Blood Gas (Full Panel)	See Arterial Blood Gas (Full Panel)	
Arterial Blood Gas Plus Serial Lactate	GASSL 2HRLACT	<p>PH 0-29 days: 7.29-7.45 30+ days: 7.35-7.45</p> <p>PCO2 0-29 days: 27-40 mm Hg 30+ days: 32-48 mm Hg</p> <p>PO2 83-108 mm Hg</p> <p>HCO3 0-29 days: 17-24 mmol/L 30+ days: 22-28 mmol/L</p> <p>OSAT 0-365 days: 40-90 % 1+ years: 94-98 %</p> <p>BASE -3.0-3.0 mmol/L</p> <p>LACT 0.5-1.6 mmol/L</p>		<p>pO2, Lactate: Amperometry</p> <p>pH, pCO2: Potentiometry</p> <p>Base excess GEM: $HCO_3 - 24.8 + 16.2 \times (pH - 7.4)$</p> <p>HCO3 GEM: $\text{Log}(HCO_3 (c)) - pH + \text{logpCO}_2 - 7.608$ mmol/L</p> <p>sO2: Co-oximetry; $sO_2 - 100 \times O_2Hb\% / (O_2Hb\% + HbB\%)$</p>	See Arterial Blood Gas (Full Panel)	See Arterial Blood Gas (Full Panel)	See Arterial Blood Gas (Full Panel)	
Arterial Blood Gas, Umbilical Cord	GASCA	<p>PH 7.23-7.33</p> <p>PCO2 41-58 mm Hg</p> <p>PO2 12-24 mm Hg</p> <p>HCO3 20-25 mmol/L</p> <p>OSAT 3-69 %</p> <p>BASE -3.0-3.0 mmol/L</p>		<p>pO2: Amperometry</p> <p>pH, pCO2: Potentiometry</p> <p>Base: Calculation of the expression that approximates the amount of acid or base required to titrate one liter of blood back to a normal pH of 7.40</p> <p>HCO3: Calculation</p> <p>sO2: Cooximetry</p>	<p>pH, pCO2, pO2, HCO3: Clinical Guide to Laboratory Tests, Tietz 3rd Edition, 1995.</p> <p>sO2: Brit. Journ. Obst. Gyn. 8:2000 Vol 107 pp 987-994 Cord Blood O2 SAT in vigorous infants at birth: What is normal?</p> <p>Base: Email: Cord blood Gas RR Physician Established Base Excess, 11-2020</p>	<p>PH: 6.80-8.00 PCO2: 5-100 mm Hg PO2: 0-700 mm Hg HCO3: 0-100 mmol/L OSAT: 5-100 % BASE: -30.0-30.0 mmol/L</p>	<p>PH: 6.80-8.00 PCO2: 5-100 mm Hg PO2: 0-700 mm Hg HCO3: 0-100 mmol/L OSAT: 5-100 % BASE: -30.0-30.0 mmol/L</p>	
Carboxyhemoglobin	COHGB	≤1.5 %	Carboxyhemoglobin is elevated in patients exposed to high exogenous sources of carbon monoxide such as cigarette smoke, vehicle exhaust, etc. or increased endogenous production such as hemolysis.	Co-oximetry; $COHb\% = 100 \times (COHb/TIb)\%$	See Arterial Blood Gas (Full Panel)	<p>Radiometer ABL800: 0.0-50.0 %</p> <p>GEM 5000: 0.0-25.0%</p>		
Cooximetry, Whole Blood	COOXB	<p>O2HGB 94-98 %</p> <p>MtHGB ≤1.5 %</p> <p>COHGB ≤1.5 %</p> <p>BGHGB 0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 11.4-15.2 g/dL</p> <p>OSATA 0-365 days: 40-90 % 1+ years: 94-98 %</p> <p>OSATV 70-80 %</p> <p>OSATMV 60-80%</p>	<p>O2HGB 94-98 %</p> <p>MtHGB ≤1.5 %</p> <p>COHGB ≤1.5 %</p> <p>BGHGB 0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL</p> <p>OSATA 0-365 days: 40-90 % 1+ years: 94-98 %</p> <p>OSATV 70-80 %</p> <p>OSATMV 60-80%</p>		<p>COHGB: Co-oximetry; $COHb\% = 100 \times (COHb/TIb)\%$</p> <p>MtHGB: Co-oximetry; $MtHb\% = 100 \times (MtHb/TIb)\%$</p> <p>O2HGB: Co-oximetry; $O_2Hb\% = 100 \times (O_2Hb/TIb)\%$</p> <p>Total HGB: Co-oximetry; $TIb = O_2Hb + COHb + MtHb + HbB$</p> <p>sO2: Co-oximetry; $sO_2 - 100 \times O_2Hb\% / (O_2Hb\% + HbB\%)$</p>	See Arterial Blood Gas (Full Panel)	<p>Radiometer ABL800: O2HGB: 0-100 % MtHGB: 0.0-30.0 % COHGB: 0.0-50.0 % BGHGB: 4.8-23.5 g/dL OSAT: 5-100 %</p> <p>GEM 5000: BGHGB: 5.0-20.0 g/dL O2HGB: 0-100 % COHGB: 0.0-75.0 % MtHGB: 0.0-30.0 % sO2: 0-100 %</p>	<p>Radiometer ABL800: O2HGB: 0-100 % MtHGB: 0.0-50.0 % COHGB: 0.0-75.0 % BGHGB: 4.8-23.5 g/dL OSAT: 5-100 %</p> <p>GEM 5000: BGHGB: 5.0-20.0 g/dL O2HGB: 0-100 % COHGB: 0.0-75.0 % MtHGB: 0.0-30.0 % sO2: 0-100 %</p>
Ionized Calcium (CRRT)	ICACRT		Post filter CRRT. When using Citrate anticoagulant protocol, the preferred target range is 1.2 mmol/L.	Potentiometry	Email_ICA CRRT RR Physician Established 11-19-2020	1.00-13.00 mg/dL	1.00-13.00 mg/dL	
Ionized Calcium, Serum	ICASST	4.60-5.30 mg/dL		Potentiometry	Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications, 3rd Edition, 2021 (p.102)	1.00-13.00 mg/dL	1.00-13.00 mg/dL	
Ionized Calcium, Whole Blood	ICA	4.60-5.30 mg/dL		Potentiometry	See Arterial Blood Gas (Full Panel)	1.00-13.00 mg/dL	1.00-13.00 mg/dL	
Lactate, Whole Blood	BGLACT	0.5-1.6 mmol/L		Amperometry	See Arterial Blood Gas (Full Panel)	<p>Radiometer ABL800: 0.0-30.0 mmol/L</p> <p>GEM 5000: 0.3-17.0 mmol/L</p>	<p>Radiometer ABL800: 0.0-30.0 mmol/L</p> <p>GEM 5000: 0.3-17.0 mmol/L</p>	
Lactate, Whole Blood, Serial	HRLACT 2HRLACT	0.5-1.6 mmol/L		Amperometry		0.3-17.0 mmol/L	0.3-17.0 mmol/L	
Methemoglobin	MtHGB	≤1.5 %		Co-oximetry; $MtHb\% = 100 \times (MtHb/TIb)\%$	See Arterial Blood Gas (Full Panel)	0.0-30.0 %	0.0-30.0 %	
pH, Pleural Fluid	PHF			Potentiometry		6.80-8.00	6.80-8.00	

Potassium, Whole Blood	BGK	0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.4-6 mmol/L 6-6 days: 133-146 mmol/L	8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.4-6 mmol/L 6-6 days: 133-146 mmol/L	Potentiometry	See Arterial Blood Gas (Full Panel)	Radiometer ABL800: 1.0-14.0 mmol/L GEM 5000: 1.0-14.0 mmol/L Radiometer ABL 800: 80-175 mmol/L	Radiometer ABL800: 1.0-14.0 mmol/L GEM 5000: 1.0-14.0 mmol/L Radiometer ABL 800: 80-175 mmol/L	
Sodium, Whole Blood	BGNA	0-7 days: 134-200 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL	0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL	Potentiometry	See Arterial Blood Gas (Full Panel)	GEM 5000: 100-180 mmol/L	GEM 5000: 100-180 mmol/L	
Total Hemoglobin	BGHGB	0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL	0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL	Co-oximetry; TlB + O2lB + COB + MetlB + HlB	See Arterial Blood Gas (Full Panel)	5.0-20.0 g/dL	5.0-20.0 g/dL	
Venous Blood Gas (Full Panel)	GASALL	PH 7.32-7.43 PCO2 36-52 mm Hg HCO3 22-29 mmol/L OSAT 70-80 % BASE -3.0-3.0 mmol/L GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL LACT 0-5.6 mmol/L NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L ICA 4.60-5.30 mg/dL BGHGB 0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL BGHCT (Calculated) - Radiometer 0-14 days: 40-57 % 15-30 days: 32-45 % 31-60 days: 28-35 % 61-179 days: 30-37 % 180-2 years: 31-38 % 2-5 years: 31-38 % 6-11 years: 32-40 % 12-17 years: 33-40 % 18+ years: 34-46 % BGHCT (Calculated) - GEM 5000/7000 0-15 days: 40-60 % 15-30 days: 32-44 % 30-60 days: 28-34 % 60-180 days: 29-37 % 180-2 years: 31-38 % 2-6 years: 31-38 % 6-12 years: 32-40 % 12-18 years: 32-40 % 18+ years: 34-46 % O2HGB 94-98 % COHGB ≤1.5 % MTHGB ≤1.5 %	PH 7.32-7.43 PCO2 36-52 mm Hg HCO3 22-29 mmol/L OSAT 70-80 % BASE -3.0-3.0 mmol/L GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL LACT 0-5.6 mmol/L NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L ICA 4.60-5.30 mg/dL BGHGB 0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL BGHCT (Calculated) - Radiometer 0-14 days: 40-57 % 15-30 days: 32-45 % 31-60 days: 28-35 % 61-179 days: 30-37 % 180-2 years: 31-38 % 2-5 years: 31-38 % 6-11 years: 32-40 % 12-17 years: 33-40 % 18+ years: 34-46 % BGHCT (Calculated) - GEM 5000/7000 0-15 days: 42-57 % 15-30 days: 30-46 % 30-60 days: 27-38 % 60-180 days: 29-37 % 180-2 years: 30-38 % 2-6 years: 31-38 % 6-12 years: 32-40 % 12-18 years: 33-44 % 18+ years: 40-50 % O2HGB 94-98 % COHGB ≤1.5 % MTHGB ≤1.5 %	pO2: Amperometry pH, pCO2: Potentiometry Base excess GEM: HCO3 - 24.8 + 16.2 x (pH - 7.4) HCO3 GEM: Log (HCO3 (c)) = pH + logpCO2 - 7.608 mmol/L sO2: Co-oximetry; sO2 - 100 x O2lB/(O2lB% + HlB%)	See Venous Blood Gas (Full Panel)	pH: Clinical Guide to Laboratory Tests 3rd Edition, Tietz, 1995. HCO3 Venous: Clinical Guide to Laboratory Tests 3rd Edition, Tietz, 1995. sO2 Venous: Blood Gas O2Sat: Radiometer Bulletin No. 44 Compendium of reference intervals pO2 Venous: Respiratory, 2014 Feb;19(2):168-75, doi:10.1111/resp.122225. Pub 2014 Jan 3. © 2012 Radiometer Medical ApS. All rights reserved. 995-950, 201206B. pCO2 Venous: Respiratory, 2014 Feb;19(2):168-75, doi:10.1111/resp.122225. Pub 2014 Jan 3. Base excess: Contemporary Practice in Clinical Chemistry 3rd Edition 2016. Chapter 32. Table 32.1 p450. Glucose: Clinical Guide to Laboratory Tests, 3rd Edition, Tietz, 1995; Pediatric Reference Ranges, Sokkin, 1999. Lactate: ABL 800 Flex Reference Manual, 2008, verified by GEM Validation Study, 2021. No. Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications, 3rd Edition, 2021 (p.165). Clinical Guide to Laboratory Tests Third Edition, Tietz, 1995. Lower limit adjusted and verified with GEM Validation Study, 2021. K: Clinical Guide to Laboratory Tests, Tietz 3rd Edition, 1995. Clinical Guide to Laboratory Tests, Tietz, 1995. Upper limit adjusted and verified with GEM Validation Study, 2021. ICA: Blood Gases and Critical Care Testing Physiology, Clinical Interpretations, and Laboratory Applications, 3rd Edition, 2021 (p.102). HCT: Derived from total hemoglobin reference interval. ABL 800 FLEX Reference Manual, 2008. Derived from total hemoglobin reference interval based on GEM calculation of HCT. HGB: OSU Internal Normal Range Study, October 2018. Verified with GEM Validation Study, 2021. O2HGB, COHGB, MTHGB: ABL 800 Flex Reference Manual, 2008, verified by GEM Validation Study, 2021. Clinical Guide to Laboratory Tests Third Edition, Tietz, 1995. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics Sixth Edition, 2018.	Radiometer ABL800: PH: 6.80-8.00 PCO2: 5-100 mm Hg PO2: 0-700 mm Hg HCO3: 0-100 mmol/L OSAT: 5-100 % FIO2: 40-100 % BASE: -30.0-30.0 mmol/L GLUC: 1-1,000 mg/dL LACT: 0.0-30.0 mmol/L NA: 80-175 mmol/L K: 1.0-14.0 mmol/L ICA: 1.00-13.00 mg/dL BGHGB: 4.8-23.5 g/dL HCT: 14-71 % O2HGB: 0-100 % COHGB: 0.0-50.0 % MTHGB: 0.0-30.0 % GEM 5000: PH: 7.00-7.92 PCO2: 18-125 mm Hg PO2: 35-529 mm Hg HCO3: 0-100 mmol/L sO2: 0-100 % FIO2: 40-100 % BASE: -30.0-30.0 mmol/L GLUC: 16-685 mg/dL LACT: 0.3-17.0 mmol/L NA: 100-180 mmol/L K: 1.0-10.0 mmol/L ICA: 1.00-13.00 mg/dL BGHGB: 5.0-20.0 g/dL HCT: 15-60 % O2HGB: 0-100 % COHGB: 0.0-75.0 % MTHGB: 0.0-30.0 %	Radiometer ABL800: PH: 6.80-8.00 PCO2: 5-100 mm Hg PO2: 0-700 mm Hg HCO3: 0-100 mmol/L OSAT: 5-100 % FIO2: 40-100 % BASE: -30.0-30.0 mmol/L GLUC: 1-1,000 mg/dL LACT: 0.0-30.0 mmol/L NA: 80-175 mmol/L K: 1.0-14.0 mmol/L ICA: 1.00-13.00 mg/dL BGHGB: 4.8-23.5 g/dL HCT: 14-71 % O2HGB: 0-100 % COHGB: 0.0-50.0 % MTHGB: 0.0-30.0 % GEM 5000: PH: 7.00-7.92 PCO2: 18-125 mm Hg PO2: 35-529 mm Hg HCO3: 0-100 mmol/L sO2: 0-100 % FIO2: 40-100 % BASE: -30.0-30.0 mmol/L GLUC: 16-685 mg/dL LACT: 0.3-17.0 mmol/L NA: 100-180 mmol/L K: 1.0-10.0 mmol/L ICA: 1.00-13.00 mg/dL BGHGB: 5.0-20.0 g/dL HCT: 15-60 % O2HGB: 0-100 % COHGB: 0.0-75.0 % MTHGB: 0.0-30.0 %
Venous Blood Gas	GASV5	PH 7.32-7.43 PCO2 36-52 mm Hg HCO3 22-29 mmol/L OSAT 70-80 % BASE -3.0-3.0 mmol/L	PH 7.32-7.43 PCO2 36-52 mm Hg HCO3 22-29 mmol/L OSAT 70-80 % BASE -3.0-3.0 mmol/L	pO2: Amperometry pH, pCO2: Potentiometry Base excess GEM: HCO3 - 24.8 + 16.2 x (pH - 7.4) HCO3 GEM: Log (HCO3 (c)) = pH + logpCO2 - 7.608 mmol/L sO2: Co-oximetry; sO2 - 100 x O2lB/(O2lB% + HlB%)	See Venous Blood Gas (Full Panel)	See Venous Blood Gas (Full Panel)	See Venous Blood Gas (Full Panel)	

Venous Blood Gas Plus Serial Lactate	GASVLT 2HRLACT		<p>PH 7.32-7.43</p> <p>PCO2 36-52 mm Hg</p> <p>HC03 22-29 mmol/L</p> <p>OSAT 70-80 %</p> <p>BASE -3.0-3.0 mmol/L</p> <p>LACT 0.5-1.6 mmol/L</p>		<p>pO2, Lactate: Amperometry</p> <p>pH, pCO2: Potentiometry</p> <p>Base excess GEM: $\text{HCO}_3^- - 24.8 + 16.2 \times (\text{pH} - 7.4)$</p> <p>HC03 GEM: $\text{Log}(\text{HCO}_3^-) - \text{pH} + \text{log}(\text{PCO}_2) - 7.608$ mmol/L</p> <p>sO2: Co-oximetry: $\text{sO}_2 - 100 \times \text{O2IB}\% / (\text{O2IB}\% + 1\text{BB}\%)$</p>	See Venous Blood Gas (Full Panel)	See Venous Blood Gas (Full Panel)	See Venous Blood Gas (Full Panel)
Venous Blood Gas, Umbilical Cord	GASCV		<p>PH 7.30-7.40</p> <p>PCO2 33-44 mm Hg</p> <p>PO2 23-35 mm Hg</p> <p>HC03 16-25 mmol/L</p> <p>OSAT 16-83 %</p> <p>BASE -2.0-2.0 mmol/L</p>		<p>pO2: Amperometry</p> <p>pH, pCO2: Potentiometry</p> <p>Base: Calculation of the expression that approximates the amount of acid or base required to titrate one liter of blood back to a normal pH of 7.40</p> <p>HC03: Calculation</p> <p>sO2: Co-oximetry</p>	<p>pCO2, pH, HCO3: Clinical Guide to Laboratory Tests, Tietz, 3rd Edition, 1995.</p> <p>Base: Email Cord blood Gas RR Physician Established Base Excess, 11-2020.</p> <p>sO2: Brit Jasm Obst Gyn 8:2000 Vol 107 pp 987, 994 Cord Bld O2 SAT in vigorous infants at birth: What is normal?</p>	<p>PH: 6.80-8.00</p> <p>PCO2: 5-100 mm Hg</p> <p>PO2: 0-700 mm Hg</p> <p>HC03: 0-100 mmol/L</p> <p>OSAT: 5-100 %</p> <p>BASE: -30.0-30.0 mmol/L</p>	<p>PH: 6.80-8.00</p> <p>PCO2: 5-100 mm Hg</p> <p>PO2: 0-700 mm Hg</p> <p>HC03: 0-100 mmol/L</p> <p>OSAT: 5-100 %</p> <p>BASE: -30.0-30.0 mmol/L</p>
1 Hour Post Glucose	GLUG		70-134 mg/dL		Photometric rate with hexokinase	ADA Standards October 2012, Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Seldin, 1999	10-800 mg/dL	10-2,400 mg/dL
Acetaminophen Level			≤150.0 mcg/mL		Enzyme immunoassay	Applied Pharmacokinetics: Principles of Therapeutic Drug Monitoring, 2nd Edition 2002 Applied Therapeutics, Inc. and Micromedex On OSU Intranet.	10.0-200.0 mcg/mL	10.0-600.0 mcg/mL
Albumin-Adjusted Calcium	ACTM	<p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1-19 years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1-3 years: 8.6-10.5 mg/dL</p>	<p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 3.2-4.8 g/dL 1-19 years: 3.2-4.7 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1-3 years: 8.6-10.5 mg/dL</p>	<p>ALB Colometric; Bromocresol green</p> <p>CA Photometric; arsenazo</p> <p>AAC Total calcium mg/dL + 0.8*(4-albumin g/dL)</p>	See individual analytes	See individual analytes	See individual analytes	
Albumin	AAC	<p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1-19 years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p>	<p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 3.2-4.8 g/dL 1-19 years: 3.2-4.7 g/dL 19+ years: 3.5-5.0 g/dL</p>	Colometric; Bromocresol green	Tietz 2nd Edition referenced by Beckman Coulter IFU for recumbent adult and verified by OSUWMC Reference Interval Study 2021.	1.5-6.0 g/dL	1.5-18.0 g/dL	
Albumin, CSF	CFMAlB		10.0-30.0 mg/dL		Turbidimetry	CCLM Vol 54 issue 2 p285-292, Feb. 2016.	1.0-45.0 mg/dL	1.0-450.0 mg/dL
Albumin, Fluid	FAlB			<p>Pleural: Serum-pleural fluid albumin gradients of >1.2 g/dL are consistent with transudates.</p> <p>Peritoneal: Serum-ascites albumin gradient (SAAG) of 1.1 g/dL or greater suggests portal hypertension.</p> <p>Pericardial: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.</p>	Colometric; Bromocresol green	<p>Pleural: Roth, B.J., et al. Chest, Vol 98, 546-549, 1990.</p> <p>Peritoneal: Runyon, H.A. Ann Intern Med. 1992;117:215-220.</p>	0.5-6.0 g/dL	0.5-6.0 g/dL
Alcohol (Ethanol), Blood	ETOH		~10 mg/dL	<p>10-50 mg/dL - effects of alcohol may or may not be apparent</p> <p>10-120 mg/dL - mild euphoria and/or impairment</p> <p>180-300 mg/dL - disorientation, confusion, dizziness</p> <p>350-500 mg/dL - unconsciousness, coma, circulation and respiration impairment, possible death</p> <p>> 450 mg/dL - possible death due to respiratory impairment</p> <p>Concentrations of blood alcohol causing impairment and/or intoxication may vary by person.</p>	Enzymatic using alcohol dehydrogenase		10-600 mg/dL	10-600 mg/dL
Alk Phosphatase	ALP	<p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p>	<p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 62-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p>	<p>Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p>	OSUWMC Reference Range Study effective 12.11.2013; Verified by OSUWMC Reference Interval Study 2021; Pediatric Reference Ranges, Seldin, 1999; Synchro Performance Verification Manual A22219	5-1,500 U/L	5-15,000 U/L	
ALP ALT AST	ENZ3	<p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-38 U/L</p>	<p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 62-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p>	<p>ALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p>	See individual analytes	See individual analytes	See individual analytes	

ALP ALT AST LD		<p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-39 U/L</p> <p>LD 0-30 days: 145-765 U/L 31-365 days: 190-420 U/L 1-3 years: 165-395 U/L 4-6 years: 135-345 U/L 7-9 years: 140-280 U/L 10-12 years: 120-260 U/L 13-15 years: 100-275 U/L 16-18 years: 105-230 U/L 19+ years: 100-190 U/L</p>	<p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 62-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p> <p>LD 0-30 days: 125-735 U/L 31-365 days: 170-450 U/L 1-3 years: 155-345 U/L 4-6 years: 155-345 U/L 7-9 years: 145-300 U/L 10-12 years: 120-225 U/L 13-15 years: 120-290 U/L 16-18 years: 105-235 U/L 19+ years: 100-190 U/L</p>		<p>ALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>LD Photometric rate</p>	See individual analytes	See individual analytes	See individual analytes
EN24								
ALP ALT AST LD GGT BIL		<p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-39 U/L</p> <p>LD 0-30 days: 145-765 U/L 31-365 days: 190-420 U/L 1-3 years: 165-395 U/L 4-6 years: 135-345 U/L 7-9 years: 140-280 U/L 10-12 years: 120-260 U/L 13-15 years: 100-275 U/L 16-18 years: 105-230 U/L 19+ years: 100-190 U/L</p> <p>GGT 0-182 days: 15-132 U/L 183-365 days: 8-39 U/L 1-12 years: 8-22 U/L 13-18 years: 8-24 U/L 19+ years: 8-64 U/L</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p>	<p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 62-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p> <p>LD 0-30 days: 125-735 U/L 31-365 days: 170-450 U/L 1-3 years: 155-345 U/L 4-6 years: 155-345 U/L 7-9 years: 145-300 U/L 10-12 years: 120-225 U/L 13-15 years: 120-290 U/L 16-18 years: 105-235 U/L 19+ years: 100-190 U/L</p> <p>GGT 0-182 days: 12-122 U/L 183-365 days: 8-39 U/L 1-12 years: 8-22 U/L 13-18 years: 8-42 U/L 19+ years: 8-64 U/L</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>DBIL</p>		<p>ALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>LD Photometric rate</p> <p>GGT Photometric rate</p> <p>DBIL Photometric with diazonium salt, 3,5-dichloroaniline (DPD).</p> <p>TBIL Photometric with 3,5-dichlorophenyl diazonium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.</p>	See individual analytes	See individual analytes	See individual analytes
EN24GB								
Alpha 1 Antitrypsin	A1AT		84-218 mg/dL		Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	30-500 mg/dL	30-5,000 mg/dL
ALT		<p>0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 8-48 U/L</p>	<p>0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p>		Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.	OSUWMC Reference Range Study effective 12.11.2013; verified by OSUWMC Reference Interval Study 2021. Pediatric Reference Ranges, Soldin, 1999 (Lower end of reference range modified to agree with the linear limits.)	3-500 U/L	3-25,000 U/L
Ammonia	NH4		6-47 μmol/L		Photometric	Package insert	10-600 μmol/L	10-3,000 μmol/L
Ammonia, Arterial	NH4A		6-47 μmol/L		Photometric	Package insert	10-600 μmol/L	10-3,000 μmol/L
AMY LIPA	AMY/LIPA		AMY 0-30 days: 0-6 U/L 31-182 days: 1-17 U/L		Photometric rate	See individual analytes	See individual analytes	See individual analytes
Amylase	AMY		0-30 days: 0-6 U/L 31-182 days: 1-17 U/L		Photometric rate	Prior study verified by OSUWMC Reference Interval Study 2021.	10-2,000 U/L	10-10,000 U/L

Amylase, Body Fluid	FAMY			<p>Pleural: Pleural fluid rich in amylase (fluid amylase to serum/plasma amylase ratio > 1) is associated with acute and chronic pancreatitis, esophageal leakage, malignancy, carcinoma, or pneumonia.</p> <p>Pancreatic cyst: Very low pancreatic cyst fluid amylase concentrations (< 250 U/L) exclude a pseudocyst in the majority of cases.</p> <p>Peritoneal Drainage, PFU: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.</p>	Photometric rate	Pleural: State of the art. The pleura Sabu SA Am Rev Respir Dis. 1988;138(1):184. Pancreatic cyst: The GH, et al. Am J Gastroenterol. 208:113-464-479.	10-2,000 U/L	10-10,000 U/L
Amylase, Urine Random	LIAMYR			The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.	Photometric rate	Prior study verified by OSUWMC Reference Interval Study 2021.	10-1,500 U/L	10-75,000 U/L
Amylase, 24 Hour Urine	LIAMY		0-400 U/24 hrs		Photometric rate	Prior study verified by OSUWMC Reference Interval Study 2021.		
Angiotensin Converting Enzyme	ACE		19-123 U/L		Furylacetoxy-Phenylalanyl-Glycylglycine	Calculated reference range based on correlation study with Mayo Clinic.	14-145 U/L	14-1,450 U/L
Anti Streptolysin O	ASO		<250 IU/mL		Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	100-1,000 IU/mL	100-10,000 IU/mL
AST	AST	<p>0-30 days: 0-49 U/L 31-365 days: 0-49 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-30 U/L</p>	<p>0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-30 U/L</p>		Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.	Verified by OSUWMC Reference Interval Study 2021.	3-1,000 U/L	3-50,000 U/L
Basic Metabolic Panel	BMP	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+7 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.80-1.20 mg/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO CALC 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+7 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO CALC 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p>		<p>NA/K/CL Indirect ion-selective electrode</p> <p>CO2 Photometric</p> <p>GLUC Photometric rate with hexokinase</p> <p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p> <p>CA Photometric, arsenazo</p> <p>ANION GAP (Na + K) - (Cl + CO2)</p> <p>eGFR eGFR = 142 x min(Ser/Cr, 1) x max(Ser/Cr, 1) - 1.200 x 0.9938Age x 1.012 [if female]</p> <p>Where x = 0.7 (females) or 0.9 (males) a = -0.241 (female) or -0.302 (male) Ser = serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L Age (years)</p> <p>The "min/Ser/Cr, 1" factor indicates the minimum of Ser/Cr or 1.0 and "max/Ser/Cr, 1" indicates the maximum of Ser/Cr or 1.0.</p> <p>OSMO CALC (1.86 (Na + K) + 1.15 (Glucose/18) + (Urea/2.8) + 14 where Na and K are in mmol/L; Glucose and Urea are in mg/dL</p>	See individual analytes	See individual analytes	See individual analytes
Beta hCG, Qual. Blood	SPREG			Negative	Lateral-flow test using a monoclonal antibody specific to the beta subunit of hCG.	Package insert		Negative, Positive
Beta hCG, Quant. Blood	QBCCB			<p>Non-pregnant: <10 mIU/mL Postmenopausal: <10 mIU/mL Male: <10 mIU/mL</p> <p>FEMALE GESTATIONAL AGE 2-4 Weeks: 39.1-8,388 mIU/mL 5-6 Weeks: 861-88,769 mIU/mL 6-8 Weeks: 8,636-218,085 mIU/mL 8-10 Weeks: 18,700-244,467 mIU/mL 10-12 Weeks: 23,143-181,899 mIU/mL 13-27 Weeks: 6,303-97,171 mIU/mL 24-40 Weeks: 4,360-74,883 mIU/mL</p>	Two-site sandwich immunoassay chemiluminescent.	Advia Centaur hCG Package Insert 10634917_EN Rev. F, 2011-04	2.6-1,000.0 mIU/mL	2.6-128,000,000.0 mIU/mL
Beta-Hydroxybutyrate, Serum	BHBB		<0.27 mmol/L	Test results cannot be interpreted as absolute evidence for the presence or absence of exogenous glucose.	Photometric	Saabho Package Insert and verified by OSUWMC Reference Interval Study 2021.	0.10-8.00 mmol/L	0.10-24.00 mmol/L
Bicarbonate, Fluid	FCO2			<p>Stool: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.</p> <p>PFU: Measurement of secretin stimulated electrolytes and enzymes in pancreatic fluid may correlate with the degree of pancreas function. Peak bicarbonate values > 75 mmol/L for screening (time 15 minutes) and >80 mmol/L for confirmatory (time 60 minutes) tests suggest normal pancreas function.</p>	Photometric		5-45 mmol/L	5-90 mmol/L

Bilirubin - Baby		<p>DBIL <0.3 mg/dL</p> <p>TBIL 0-1 day: 14-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p>			<p>DBIL Photometric with diazotium salt, 3,5-dichloroaniline (DPD).</p> <p>TBIL Photometric with 3,5-dichlorophenyldiazotium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.</p>	See individual analytes	See individual analytes	See individual analytes
Bilirubin Direct	BIL.D		<0.3 mg/dL		Photometric with diazotium salt, 3,5-dichloroaniline (DPD).	Clinical Guide to Laboratory Tests, Tietz 1995, verified by OSUWMC Reference Interval Study 2021.	>10.0 mg/dL	0.1-20.0 mg/dL
Bilirubin Total			<p>0-1 day: 14-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p>		Photometric with 3,5-dichlorophenyldiazotium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.	Clinical Guide to Laboratory Tests, Tietz 1995, verified by OSUWMC Reference Interval Study 2021.	0.1-30.0 mg/dL	0.1-90.0 mg/dL
Bilirubin, Total and Direct			<p>DBIL <0.3 mg/dL</p> <p>TBIL 0-1 day: 14-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p>		<p>DBIL Photometric with diazotium salt, 3,5-dichloroaniline (DPD).</p> <p>TBIL Photometric with 3,5-dichlorophenyldiazotium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.</p>	See individual analytes	See individual analytes	See individual analytes
Bilirubin, Total, Fluid				<p>Peritoneal: Peritoneal bilirubin concentrations greater than that of serum/plasma may suggest bile within the abdomen.</p> <p>Drainage: Drain fluid bilirubin concentration-to-serum/plasma bilirubin concentration ratios exceeding 5 indicates bile leakage.</p>	Photometric with 3,5-dichlorophenyldiazotium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.	Peritoneal: Runyon BA, J Clin Gastroenterol. 1987;9(5):543. Drain: Darwin. Gastrointestinal Endosc; 2010 Jan;71(1):99-104.	0.1-30.0 mg/dL	0.1-90.0 mg/dL
BMP without Glucose		<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-146 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+7 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m2</p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-146 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+7 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m2</p>	<p>NA/K/CL Indirect ion-selective electrode</p> <p>CO2 Photometric</p> <p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p> <p>CA Photometric, arsenazo</p> <p>ANION GAP (Na + K) - (Cl + CO2)</p> <p>eGFR eGFR = $142 \times \text{minSerCr}_i / 1.73 \times \text{maxSerCr}_i$, 1]-1.200 x 0.993^{Age} x 1.012 [if female]</p> <p>Where: $x = 0.7$ (female) or 0.9 (male) $i = 0.241$ (female) or 0.302 (male) $\text{Ser} =$ serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L Age (years) The "minSerCr_i, 1]" factor indicates the minimum of SerCr_i or 1.0 and "maxSerCr_i, 1]" indicates the maximum of SerCr_i or 1.0.</p>	See individual analytes	See individual analytes	See individual analytes	
BUN	BMPNG		7-25 mg/dL		Photometric rate	Bockman Coulter IFU for serum verified by OSUWMC Reference Interval Study 2021.	2-130 mg/dL	2-650 mg/dL
BUN CREA		<p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p>	<p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p>		<p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p>	See individual analytes	See individual analytes	See individual analytes
C Reactive Protein	BCR		<10.00 mg/L		Measurement of the rate of decrease in light intensity transmitted (increase in absorbance) through particles suspended in solution is the result of complexes formed during the immunological reaction between the CRP of the patient serum and rabbit anti-CRP-antibodies coated on latex particles.	Clinical Guide to Laboratory Tests, Tietz, 2005; Verified by OSUWMC Reference Interval Study 2021.	0.20-80.00 mg/L	0.20-480.00 mg/L
C Reactive Protein for Cardiac Risk	CRP		<2.00 mg/dL		Measurement of the rate of decrease in light intensity transmitted (increase in absorbance) through particles suspended in solution is the result of complexes formed during the immunological reaction between the CRP of the patient serum and rabbit anti-CRP-antibodies coated on latex particles.	Clinical Guide to Laboratory Tests, Tietz, 2005; Verified by OSUWMC Reference Interval Study 2021.	0.20-80.00 mg/L	0.20-480.00 mg/L
C3 Complement	C3		87-200 mg/dL		Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	15-500 mg/dL	15-1,500 mg/dL
C3, C4			<p>C3 87-200 mg/dL</p> <p>C4 8-52 mg/dL</p>		Turbidimetry	See individual analytes	See individual analytes	See individual analytes
C4 Complement	COMP		8-52 mg/dL		Turbidimetry	Historic Reference Range. Verified by OSUWMC Reference Interval Study 2021.	8-150 mg/dL	8-450 mg/dL

CA 125	CA125		≤30 U/mL		Two-site sandwich immunoassay using direct chemiluminometric technology.	Advia Centaur CA 125H Package Insert 12R516 Rev. 11, 2009-02	3-600 U/mL	3-360,000 U/mL
CA 15-3N	CA153		0.0-32.4 U/mL	0.0-32.4 U/mL (use not defined)	Two-site sandwich immunoassay chemiluminescent	Abbott IM CA 15.3 Package Insert 11206285 EN Rev. 04, 2020-03	3.0-200.0 U/mL	3.0-200,000.0 U/mL
CA 19-9	CA199		≤37.00 U/mL		Two-site sandwich immunoassay using direct chemiluminometric technology.	Clinical Guide to Laboratory Tests, Tietz, 1995; see Source link for additional Reference Range information	15.00-700.00 mg/dL	15.00-33,600,000.00 U/mL
Calcium	CA	0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL		0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL	Photometric; arsenazo	Established by OSUWMC Reference Interval Study 2013, verified by OSUWMC Reference Interval study 2021.	4.0-18.0 mg/dL	4.0-18.0 mg/dL
Calcium, Urine 24HR	UCA		100.0-1000 mg/24HR		Photometric; arsenazo			
Calcium/Creat Ratio, Random Urine	CALCR		0-299 days: <0.86 Ca mg/Crea mg 210-569 days: <0.60 Ca mg/Crea mg 570 days-2 years: <0.42 Ca mg/Crea mg 3+ years: <0.22 Ca mg/Crea mg		CALCIUM Photometric; arsenazo CREA Kinetic Jaffe	See individual analytes	See individual analytes	See individual analytes
Carbamazepine Total Level	CARB		Therapeutic Range: 4.0-12.0 mcg/mL		Competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change.	Applied Clinical Pharmacokinetics, 2001 Micromedex, OSU Intranet	2.0-20.0 mcg/mL	2.0-100.0 mcg/mL
CEA	CEA		≤5.0 ng/mL		Two-site Sandwich Immunoassay Chemiluminescent	Clinical Guide to Laboratory Tests, Tietz, 1995. See source link for additional Reference Range information.	2.0-100.0 ng/mL	2.0-8,000,000.0 ng/mL
Ceruloplasmin	CERP		20-60 mg/dL		Turbidimetry	Verified by OSUWMC Reference Interval Study 2021.	6-200 mg/dL	6-4,000 mg/dL
Chem 6 (Lytes, BUN, CREA)		NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L BUN 7-25 mg/dL CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.90-1.20 mg/dL ANION GAP 7-17 mmol/L eGFR ≥60 mL/min/1.73m2	NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L BUN 7-25 mg/dL CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL ANION GAP 7-17 mmol/L eGFR ≥60 mL/min/1.73m2	NA/K/CL Indirect ion-selective electrode CO2 Photometric BUN Photometric rate CREA Kinetic Jaffe ANION GAP (Na + K) - (Cl + CO2) BUN/CREA RATIO BUN/Creatinine eGFR eGFR = 142 x min(Ser/cr, 1) x max(Ser/cr, 1) - 1.200 x 0.9938Age x 1.012 [if female] Where: c = 0.7 (females) or 0.9 (males) a = -0.241 (female) or -0.302 (male) Ser = serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L Age (years) The "min/Ser/cr, 1" factor indicates the minimum of Ser/cr or 1.0 and "max/Ser/cr, 1" indicates the maximum of Ser/cr or 1.0.	See individual analytes	See individual analytes	See individual analytes	
Chem 7 (LYTES, BUN, CREA, GLUC)	CHM6	NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L BUN 7-25 mg/dL	NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L BUN 7-25 mg/dL	NA/K/CL Indirect ion-selective electrode CO2 Photometric BUN Photometric rate CREA Kinetic Jaffe GLUC Photometric rate with hexokinase	See individual analytes	See individual analytes	See individual analytes	

		<p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>OSMO (CALC) 278-305 mOsm/kg</p> <p>eGFR ≥60 mL/min/1.73m²</p>	<p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>OSMO (CALC) 278-305 mOsm/kg</p> <p>eGFR ≥60 mL/min/1.73m²</p>		<p>BUN/CREA RATIO BUN/Creatinine</p> <p>eGFR $eGFR = 142 \times \min(Ser/Cr, 1) \times \max(Ser/Cr, 1) - 1.200 \times 0.9938Age \times 1.012$ [if female]</p> <p>Where $x = -0.7$ (females) or 0.9 (males) $a = -0.241$ (female) or -0.302 (male) $Ser =$ serum creatinine in mg/dL; divide by 88.4 for creatinine in μmol/L</p> <p>Age (years) The "min/SerCr, 1" factor indicates the minimum of Ser/Cr or 1.0 and "max/SerCr, 1" indicates the maximum of Ser/Cr or 1.0.</p>				
Chloride	CL		0-365 days: 96-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L		Indirect ion-selective electrode	Established by OSUWMC Reference Interval Study 2013, verified by OSUWMC Reference Interval study 2021	50-200 mmol/L	50-200 mmol/L	
Chloride, 24 HR Urine	UCL		110-250 mmol/24 hrs		Indirect ion-selective electrode	Clinical Guide to Laboratory Tests, Tietz, 1995			
Chloride, Random Urine	UCLR			The reference range has not been established for random urine specimens. The test results should be integrated into the clinical context for interpretation.	Indirect ion-selective electrode		15-400 mmol/L	15-400 mmol/L	
Cholesterol, Body Fluid				Pleural: Pleural fluid cholesterol concentrations > 200 mg/dL are associated with pseudochylous effusions. Peritoneal: Peritoneal fluid cholesterol concentrations greater than 32-70 mg/dL may suggest malignant ascites.	Enzymatic colorimetric	Pleural: Hooper C, et al. Thorax. 2010 Aug;65:Suppl2:i66-17. McGrath, et al. Int J Clin Pract. 2009 Nov;63(11):1653-9. Peritoneal: Block, et al. Crit Rev Clin Lab Sci. 2013;50:107-124.	25-700 mg/dL	25-700 mg/dL	
Cholesterol Total	TCHOL	0-30 days: 62-155 mg/dL 31-182 days: 62-141 mg/dL 183-365 days: 76-216 mg/dL 1-3 years: 108-193 mg/dL 4-6 years: 106-193 mg/dL 7-9 years: 104-210 mg/dL 10-12 years: 105-218 mg/dL 13-15 years: 108-205 mg/dL 16-18 years: 92-234 mg/dL 19+ years: <200 mg/dL	0-30 days: 54-151 mg/dL 31-182 days: 81-147 mg/dL 183-365 days: 76-179 mg/dL 1-3 years: 85-182 mg/dL 4-6 years: 110-217 mg/dL 7-9 years: 116-211 mg/dL 10-12 years: 105-223 mg/dL 13-15 years: 91-204 mg/dL 16-18 years: 82-192 mg/dL 19+ years: <200 mg/dL		Enzymatic colorimetric	National Cholesterol Education Project (NCEP) Adult Treatment Protocol (ATP-III) (Circulation. 2002;106:3143-3421)	25-700 mg/dL	25-2100 mg/dL	
Cholesterol Triglyceride	TCHOL	0-30 days: 62-155 mg/dL 31-182 days: 62-141 mg/dL 183-365 days: 76-216 mg/dL 1-3 years: 108-193 mg/dL 4-6 years: 106-193 mg/dL 7-9 years: 104-210 mg/dL 10-12 years: 105-218 mg/dL 13-15 years: 108-205 mg/dL 16-18 years: 92-234 mg/dL 19+ years: <200 mg/dL	0-30 days: 54-151 mg/dL 31-182 days: 81-147 mg/dL 183-365 days: 76-179 mg/dL 1-3 years: 85-182 mg/dL 4-6 years: 110-217 mg/dL 7-9 years: 116-211 mg/dL 10-12 years: 105-223 mg/dL 13-15 years: 91-204 mg/dL 16-18 years: 82-192 mg/dL 19+ years: <200 mg/dL		Enzymatic colorimetric	See individual analytes	See individual analytes	See individual analytes	
	TRIG	0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 30-140 mg/dL 12-13 years: 37-130 mg/dL 14-15 years: 38-135 mg/dL 16-19 years: 37-140 mg/dL 20+ years: 25-150 mg/dL	0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 24-137 mg/dL 12-13 years: 24-145 mg/dL 14-15 years: 34-165 mg/dL 16-19 years: 34-140 mg/dL 20+ years: 25-150 mg/dL						
CK	CHTRG	0-30 days: 2-146 U/L 31-182 days: 2-146 U/L 183-365 days: 18-138 U/L 1-3 years: 2-134 U/L 4-6 years: 8-147 U/L 7-9 years: 26-145 U/L 10-12 years: 6-137 U/L 13-15 years: 2-143 U/L 16-18 years: 13-144 U/L 19+ years: 30-184 U/L	0-30 days: 2-183 U/L 31-182 days: 2-129 U/L 183-365 days: 2-143 U/L 1-3 years: 2-163 U/L 4-6 years: 18-158 U/L 7-9 years: 2-177 U/L 10-12 years: 6-217 U/L 13-15 years: 2-251 U/L 16-18 years: 2-238 U/L 19+ years: 30-250 U/L		Photometric rate	Established by OSUWMC Reference Interval Study 2013, verified by OSUWMC Reference Interval study 2021.	10-2,000 U/L	10-200,000 U/L	
Clozapine Level	CK		350-1,000 ng/mL		Turbidimetric immunoassay	Established by OSUWMC Reference Interval Study 2023.		68-1,500 ng/mL	
	CLOZ								
		<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-6.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-6.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p>						

<p>CMPN without Glucose</p>	<p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p> <p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1-19 years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>	<p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-1.00 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR 0-4 days: 1.000-39.000 mL/min/1.73m²</p> <p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 2.8-4.8 g/dL 1-19 years: 3.2-4.7 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-283 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 42-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>	<p>NA/K/CL Indirect ion-selective electrode</p> <p>CO2 Photometric</p> <p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p> <p>CA Photometric, arsenazo</p> <p>ANION GAP (Na + K) - (Cl + CO2)</p> <p>eGFR eGFR = 142 x min(SerCr, 1) x max(SerCr, 1) - 1.200 x 0.9938Age x 1.012 [if female]</p> <p>Where x = 0.7 (females) or 0.9 (males) a = -0.241 (female) or -0.302 (male) SerCr = serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L</p> <p>TBIL Photometric with 3,5-dichlorophenylmiazonium tetrafluoroborate (DDP), and caffeine and a surfactant as accelerators.</p> <p>ALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5'-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5'-phosphate.</p> <p>TP Colorimetric with cupric ions in an alkaline solution.</p>	<p>See individual analytes</p>	<p>See individual analytes</p>	<p>See individual analytes</p>
<p>CO2 Total</p>	<p>CMPNG CO2</p>	<p>0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p>	<p>Photometric</p>	<p>Beckman Coulter BU verified by OSI/WMC Reference Interval Study 2021.</p>	<p>5-45 mmol/L</p>	<p>5-45 mmol/L</p>
<p></p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-6.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL</p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-6.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL</p>	<p>NA/K/CL Indirect ion-selective electrode</p>	<p></p>	<p></p>	<p></p>

Comprehensive Metabolic Panel	<p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p> <p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1-19 years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>JALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 108-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-332 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>	<p>183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR 0-4 days: 1,000-39,000 uU/mL</p> <p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 2.8-4.8 g/dL 1-19 years: 3.2-4.7 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>JALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 42-362 U/L 13-15 years: 74-300 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>		<p>CO2 Photometric</p> <p>GLUC Photometric rate with hexokinase</p> <p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p> <p>CA Photometric, arsenazo</p> <p>ANION GAP (Na + K) - (Cl + CO2)</p> <p>eGFR eGFR = 142 x min(Ser/Cr, 1) x max(Ser/Cr, 1) - 1.200 x 0.9938Age x 1.012 [if female]</p> <p>Where x = 0.7 (females) or 0.9 (males) a = -0.241 (female) or -0.302 (male) Ser = serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L Age (years) The "min(Ser/Cr, 1)" factor indicates the minimum of Ser/Cr or 1.0 and "max(Ser/Cr, 1)" indicates the maximum of Ser/Cr or 1.0.</p> <p>OSMO CALC (1.86(Na + K) + 1.15 (Glucose/18) + (Urea/2.8) + 14 where Na and K are in mmol/L; Glucose and Urea are in mg/dL.</p> <p>TBIL Photometric with 3,5-dichlorophenylidiazonium tetrafluoroborate (DDP), and caffeine and a surfactant as accelerators.</p> <p>JALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>TP Colorimetric with cupric ions in an alkaline solution.</p>	See individual analytes	See individual analytes	See individual analytes
Cortisol	CMP	3.09-22.40 mcg/dL		Competitive immunoassay using direct chemiluminescent technology	Atellica IM Cortisol Package Insert 11200393_EN Rev. 03-2020-03	0.50-75.00 mcg/dL	0.50-2,400.00 mcg/dL
Cortisol (DST) Overnight	COR1DX	<1.80 mcg/dL		Competitive immunoassay using direct chemiluminescent technology	Atellica IM Cortisol Package Insert 11200393_EN Rev. 03-2020-04	0.50-75.00 mcg/dL	0.50-2,400.00 mcg/dL
Cortisol, Baseline	COR1B	3.09-22.40 mcg/dL		Competitive immunoassay using direct chemiluminescent technology	Atellica IM Cortisol Package Insert 11200393_EN Rev. 03-2020-05	0.50-75.00 mcg/dL	0.50-2,400.00 mcg/dL
Cortisol, 30 Minute	COR1D	≥18.00 mcg/dL		Competitive immunoassay using direct chemiluminescent technology	Atellica IM Cortisol Package Insert 11200393_EN Rev. 03-2020-06	0.50-75.00 mcg/dL	0.50-2,400.00 mcg/dL
Cortisol, 60 Minute	COR1E	≥18.00 mcg/dL		Competitive immunoassay using direct chemiluminescent technology	Atellica IM Cortisol Package Insert 11200393_EN Rev. 03-2020-07	0.50-75.00 mcg/dL	0.50-2,400.00 mcg/dL
Creatinine Body Fluid	FLCREA			<p>CREA Kinetic Jaffe</p> <p>eGFR eGFR = 142 x min(Ser/Cr, 1) x max(Ser/Cr, 1) - 1.200 x 0.9938Age x 1.012 [if female]</p> <p>Where x = 0.7 (females) or 0.9 (males) a = -0.241 (female) or -0.302 (male) Ser = serum creatinine in mg/dL, divide by 88.4 for creatinine in μmol/L Age (years) The "min(Ser/Cr, 1)" factor indicates the minimum of Ser/Cr or 1.0 and "max(Ser/Cr, 1)" indicates the maximum of Ser/Cr or 1.0.</p> <p>Peritoneal and drainage: Fluid creatinine concentrations that are greater than serum/plasma creatinine concentrations may imply intraperitoneal leakage of urine outside of the urinary tract.</p> <p>Pleural: Pleural fluid creatinine to serum/plasma creatinine concentration ratio > 1 suggests urothorax.</p>	Manahm KJ, et al. Obstet Gynecol. 1999 May;93(5 Pt 1):780-2 Pleural: Toubes, et al. J Thorac Dis. 2017;9(5):1209-1218.	0.20-25.00 mg/dL	0.20-25.00 mg/dL
Creatinine	CREA	<p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>eGFR ≥60 mL/min/1.73m²</p>	<p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>eGFR ≥60 mL/min/1.73m²</p>	Kinetic Jaffe	OSLWMC Reference Range Study effective 12/11/2013, verified by OSLWMC Reference Interval Study 2021, Pediatric Reference Ranges, Soidin, 1999.	CREA: 0.20-25.00 mg/dL eGFR: ≥90 mL/min/1.73m ²	CREA: 0.20-25.00 mg/dL eGFR: ≥90 mL/min/1.73m ²
Creatinine, 8 HR Urine	RUCR	1.00-300.00 mg/dL	1.00-900.00 mg/dL	Kinetic Jaffe		1.00-300.00 mg/dL	1.00-900.00 mg/dL

Creatinine, 24 HR Urine	UCRF	0-2 years: 0.70-2.00 g/24 hrs 3-8 years: 0.11-0.68 g/24 hrs 9-12 years: 0.17-1.41 g/24 hrs 13-17 years: 0.29-1.87 g/24 hrs 18+ years: 0.60-1.80 g/24 hrs	0-2 years: 0.70-2.00 g/24 hrs 3-8 years: 0.11-0.68 g/24 hrs 9-12 years: 0.17-1.41 g/24 hrs 13-17 years: 0.29-1.87 g/24 hrs 18+ years: 0.60-2.00 g/24 hrs		Kinetic Jaffe	SKDEP traceable Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Soldin, 1999		
Creatinine, Random Urine	UCRFER			The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.	Kinetic Jaffe		1.00-300.00 mg/dL	1.00-900.00 mg/dL
Cystatin C and Creatinine with estimated GFR		CYSK 0.51-1.05 mg/L CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL eGFR ≥60 mL/min/1.73m ²	CYSK 0.51-1.05 mg/L CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL eGFR ≥60 mL/min/1.73m ²		CYSK Turbidometric CREA Kinetic Jaffe CKD-EPI 2021 Calculation: eGFR _{CYSK} = 135 x min(Ser/Cr, 1) x max(Ser/Cr, 1) - 0.544 x min(Ser/Cr, 1) + 0.323 x max(Ser/Cr, 1) + 0.778 x 0.9961 Age x 0.963 (if female) where x = 0.7 (female) or 0.9 (male) a = -0.219 (female) or -0.144 (male) Ser = serum creatinine in mg/dL Cysc = serum cystatin C in mg/L Age (years) The "min (Ser/Cr, 1)" factor indicates the minimum of Ser/Cr or 1.0, "max(Ser/Cr, 1)" indicates the maximum of Ser/Cr or 1.0, "min(Ser/Cr, 0.1)" indicates the minimum of Ser/Cr or 0.1, and "max(Ser/Cr, 0.1)" indicates the maximum of Ser/Cr or 0.1.	CREA: 0.20-25.00 mg/dL CYSK: 0.40-8.00 mg/L	CREA: 0.20-25.00 mg/dL CYSK: 0.40-32.00 mg/L	
Digoxin Level	DIG		Therapeutic Range: 0.3-1.0 ng/mL		Enzyme Immunoassay	Anselled Clinical Pharmacokinetics, Bloor, 2001	0.3-5.0 ng/mL	0.3-10.0 ng/mL
Estradiol, Enhanced	EE2B			Male: 19+ years: <11.8-39.8 Menstruating Females: Follicular phase: 19.5-144.2 Midcycle peak: 0.9-256.7 Luteal phase: 55.8-214.2 Post-menopausal: <11.8-32.2 This test is not recommended for patients receiving Fulvestrant (Faslodes) due to possible false elevations. EE2 levels vary widely throughout the menstrual cycle.	Competitive assay format. The endogenous estradiol contained in a sample is released from its binding proteins by a releasing agent. Then, a sheep monoclonal anti-estradiol antibody labeled with acridinium ester is added to bind available estradiol. Finally, an estradiol derivative capture solid phase is added to the reaction to compete with estradiol for the binding of the acridinium-labeled antibody. After washing, acid and base are dispensed to initiate the chemiluminescent reaction.	Advia Centaur, Enhanced Estradiol (eE2) Package Insert 10491467 Rev. C, 2016-09; Pediatric Reference Ranges, Soldin, 1999	11.8-3,000.0 pg/mL	11.8-150,000.0 pg/mL
EUS Pancreatic Lab (Fluid CEA, AMY)	FCEAR			CEA: Fluid CEA values >192 ng/mL may indicate mucinous cystic lesions of the pancreas. Results <192 ng/mL require clinical correlations with patient history and other imaging modalities. This chemiluminescent test was developed and its performance characteristics determined by the Critical Care Laboratory at The Ohio State University Wexner Medical Center. It has not been cleared or approved by the FDA. The laboratory is regulated under CLIA as qualified to perform high-complexity testing. This test is used for clinical purposes. It should not be regarded as investigational or for research. AMY: Very low pancreatic cyst fluid amylase concentrations (< 250 U/L) exclude a pseudocyst in the majority of cases.	AMY Photometric rate CEA Two-site Sandwich Immunoassay Chemiluminescent	See individual analytes	See individual analytes	See individual analytes
Ferritin	FERR	0-365 days: 5.7-421.0 ng/mL 1-12 years: 12.8-88.7 ng/mL 13-18 years: 6.8-75.6 ng/mL 19+ years: 7.3-270.7 ng/mL	0-365 days: 5.7-421.0 ng/mL 1-12 years: 12.8-88.7 ng/mL 13-18 years: 10.9-135.0 ng/mL 19+ years: 10.5-307.3 ng/mL		Two-site Sandwich Immunoassay Chemiluminescent.	Siemens Atellica IM Reference Interval Verification Study 2023 Siemens Atellica IM Reference Interval Verification Study Summary (Amendment)	0.9-1,650.0 ng/mL	0.9-1,650,000.0 ng/mL
Folate, Serum	FOLSB	0-1 years: 6.30-22.70 ng/mL 2-3 years: 1.70-15.70 ng/mL 4-6 years: 2.70-14.10 ng/mL 7-9 years: 2.40-13.40 ng/mL 10-12 years: 1.00-10.20 ng/mL 13-18 years: 1.20-7.20 ng/mL 19+ years: <5.38 ng/mL	0-1 years: 7.20-22.40 ng/mL 2-3 years: 2.50-15.00 ng/mL 4-6 years: 0.50-13.00 ng/mL 7-9 years: 2.30-11.90 ng/mL 10-12 years: 1.50-10.80 ng/mL 13-18 years: 1.20-8.80 ng/mL 19+ years: <5.38 ng/mL		Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Folate Package Insert 11200602_EN Rev. 04-2020-11; Pediatric Reference Ranges, Soldin, 1999	0.56-24.00 ng/mL	0.56-960.00 ng/mL
FSH	FSH	0-365 days: 1.9-22.9 mIU/mL 1-8 years: 1.1-9.5 mIU/mL 9-11 years: 1.2-9.2 mIU/mL 12-18 years: 1.9-11.2 mIU/mL	0-365 days: 0.8-5.2 mIU/mL 1-8 years: 0.8-2.5 mIU/mL 9-11 years: 1.0-4.2 mIU/mL 12-18 years: 2.3-10.2 mIU/mL	19+ Male: <18.1 mIU/mL 19+ years Female: Follicular: 2.5-10.2 mIU/mL Midcycle: 3.4-33.4 mIU/mL Luteal: 1.5-9.1 mIU/mL Pregnant: <0.3 mIU/mL Post Menopausal: 23.0-116.3 mIU/mL	Two-site sandwich immunoassay using direct chemiluminescent technology.	Atellica IM FSH Package Insert 11200384_EN Rev. 06-2020-09	0.3-200.0 mIU/mL	0.3-6,400.0 mIU/mL
GGT	GGT	0-182 days: 15-132 U/L 183-365 days: 8-39 U/L 1-12 years: 8-22 U/L 13-18 years: 8-24 U/L 19+ years: 8-64 U/L	0-182 days: 12-122 U/L 183-365 days: 8-39 U/L 1-12 years: 8-22 U/L 13-18 years: 8-42 U/L 19+ years: 8-64 U/L		Photometric rate	Beckman Coulter IFU verified by OSLW/MAC Reference Interval Study 2021 (lower end modified); Pediatric Reference Ranges, Soldin, 1999 (Lower end of reference range modified to agree with the linear limits)	3-1,200 U/L	3-6,000 U/L
Genstatinal 3HR Glucose Tolerance Test			GTFF 70-94 mg/dL GTIII 70-179 mg/dL GTIIH 70-154 mg/dL GTIII 70-139 mg/dL	The upper limits of the reference intervals for each fasting, 1 hour, 2 hour, and 3 hour samples are based on the Carpenter-Coustan criteria. According to the ADA practice recommendations, the diagnosis of GDM is made when at least 2 of the four plasma glucose levels meet or exceed the Carpenter-Coustan criteria, although The American College of Obstetricians and Gynecologists notes that some clinicians choose to use just one elevated value.	Photometric rate with hexokinase		10-800 mg/dL	10-2,400 mg/dL
Glucose	GLUC		0-30 days: 55-113 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL	The reference interval provided is for plasma glucose measurements from samples collected after an 8 hour fast (no caloric intake)	Photometric rate with hexokinase	ADA Standards October 2012, Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Soldin, 1999	10-800 mg/dL	10-2,400 mg/dL

Glucose Body Fluid	FGUC GLUCSF			<p>Amniotic: Amniotic fluid glucose concentrations < 10 mg/dL are consistent with intra-amniotic inflammation in patients with prelabor rupture of membranes.</p> <p>Peritoneal: Peritoneal glucose concentrations >50 mg/dL (2.8 mmol/L) are consistent with spontaneous bacterial peritonitis and concentrations below this are consistent with secondary bacterial peritonitis due to gut perforation.</p> <p>Pericardial: Pericardial fluid glucose to serum/plasma glucose ratios are 1.0 in presumed 'normal' patients.</p> <p>Pancreatic Body Cyst: Pancreatic Cyst Glucose measurements of \geq 50 mg/dL are suggestive of a mucinous lesion.</p> <p>Pleural: Pleural fluid glucose concentrations are equivalent to serum/plasma glucose concentrations in the absence of pleural pathology. Pleural fluid glucose concentrations < 60 mg/dL may indicate parapneumonic or malignant effusions. Other less common effusions associated with low glucose concentrations include hemothorax, tuberculosis, rheumatoid pleuritis, Churg-Strauss syndrome, parasitosis, and lupus pleuritis.</p> <p>Synovial: Synovial fluid glucose concentrations lower than glucose in serum/plasma are associated with infection.</p>	Photometric rate with hexokinase	Amniotic: Gonzalez-Bouquet, et al. J Matern Fetal Med. Jul-Aug 1999;8(4):155-8. Pancreatic cyst: Carr, et al. Surgery. 2018 Mar;163(3):600-605. Peritoneal: Runyon BA, Hoefs JC Hepatology. 1985;5(2):257. Pericardial: Ben Horta S, et al. Am J Med 2005;118:636-40. Pleural: Toebes, et al. J Thorac Dis. 2017;9(5):1299-1218. Light, RW. N Engl J Med. 2002 Jun 20;346(25):1971-1977 Synovial: Margaretem, et al. JAMA. 2007;297(13):1478-1488.	10-800 mg/dL	10-800 mg/dL
Glucose CSF	GLUCSF		40-70 mg/dL		Photometric rate with hexokinase	Clinical Guide to Laboratory Tests, Tietz, 1995	10-800 mg/dL	10-800 mg/dL
Glucose Fasting	GLUF		0-30 days: 55-115 mg/dL 31-182 days: 57.17 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL		Photometric rate with hexokinase	ADA Standards October 2012. Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Soldin, 1999.	10-800 mg/dL	10-2,400 mg/dL
Haptoglobin	HAPTO		44-215 mg/dL		Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021	30-400 mg/dL	30-1,200 mg/dL
hCG, Qualitative, Urine	hHCG			Negative	Lateral-flow test using a monoclonal antibody specific to the beta subunit of hCG.	Package insert		Negative, Positive
hCG, Quant (Tumor Marker)	hCGTM		<10.0 mIU/mL		Two-site sandwich immunoassay chemiluminescent.	Advia Centaur hCG Package Insert 10634917_EN Rev. F, 2011.04.	2.6-1,000.0 mIU/mL	2.6-128,000,000.0 mIU/mL
HDL Cholesterol	HDL		0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: \geq 40 mg/dL		Enzymatic colorimetric	National Cholesterol Education Project (NCEP) Adult Treatment Protocol (ATP-III) (Circulation, 2002;106:3143-3421)	3-200 mg/dL	3-200 mg/dL
Hepatic Function Panel	HFP	<p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1+ years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>DBIL -0.3 mg/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>ALP 0-30 days: 48-406 U/L 31-365 days: 124-341 U/L 1-3 years: 100-317 U/L 4-6 years: 96-297 U/L 7-9 years: 69-325 U/L 10-12 years: 51-333 U/L 13-15 years: 50-162 U/L 16-18 years: 47-119 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31 days-4 years: 8-30 U/L 4-6 years: 8-25 U/L 7-9 years: 8-25 U/L 10-17 years: 8-20 U/L 18+ years: 9-48 U/L</p> <p>AST 0-30 days: 0-49 U/L 31-365 days: 0-80 U/L 1-3 years: 0-70 U/L 4-6 years: 0-60 U/L 7-9 years: 0-42 U/L 10-12 years: 0-38 U/L 13-15 years: 0-33 U/L 16-18 years: 0-31 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1+ years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>	<p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 2.8-4.8 g/dL 1+ years: 2.4-7.7 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>DBIL -0.3 mg/dL</p> <p>TBIL 0-1 day: 1.4-8.7 mg/dL 1-2 days: 3.4-11.5 mg/dL 3-4 days: 1.5-12.0 mg/dL 5-365 days: 0.3-1.2 mg/dL 1+ years: <1.5 mg/dL</p> <p>ALP 0-30 days: 75-316 U/L 31-365 days: 82-383 U/L 1-3 years: 104-345 U/L 4-6 years: 93-309 U/L 7-9 years: 86-315 U/L 10-12 years: 42-362 U/L 13-15 years: 74-390 U/L 16-18 years: 52-171 U/L 19+ years: 32-126 U/L</p> <p>ALT 0-30 days: 8-25 U/L 31-365 days: 8-35 U/L 1-3 years: 8-30 U/L 4-6 years: 8-20 U/L 7-9 years: 8-25 U/L 10-17 years: 8-30 U/L 18+ years: 10-52 U/L</p> <p>AST 0-30 days: 0-52 U/L 31-365 days: 0-66 U/L 1-3 years: 0-57 U/L 4-6 years: 0-49 U/L 7-9 years: 0-43 U/L 10-12 years: 0-39 U/L 13-15 years: 0-40 U/L 16-18 years: 0-40 U/L 19+ years: 10-39 U/L</p> <p>TP 0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1+ years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p>	<p>ALB Colorimetric; Bromocresol green.</p> <p>DBIL Photometric with diazonium salt, 3,5-dichloroaniline (DPD).</p> <p>TBIL Photometric with 3,5-dichlorophenylmazonium tetrafluoroborate (DPD), and caffeine and a surfactant as accelerators.</p> <p>ALP Photometric rate with p-nitro-phenylphosphate (pNPP) in the presence of 2-amino-2-methyl-1-propanol (AMP) at pH 10.4.</p> <p>ALT Photometric rate with alanine to α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>AST Photometric rate with aspartate and α-oxoglutarate to form pyruvate and glutamate. This assay does not contain pyridoxyl-5-phosphate.</p> <p>TP Colorimetric with cupric ions in an alkaline solution.</p>	See individual analytes	See individual analytes	See individual analytes	
High Sensitivity Troponin I - Single Order	HSTLIX		<34 ng/L	<53 ng/L	Three-site sandwich immunoassay using direct chemiluminescent technology.	Atellica IM Troponin I Package Insert 11200498_EN Rev. 06, 2019-06.	3-25,000 ng/L	3-2,000,000 ng/L
High Sensitivity Troponin I x2	HSTLHX		<34 ng/L	<53 ng/L	Three-site sandwich immunoassay using direct chemiluminescent technology.	Atellica IM Troponin I Package Insert 11200498_EN Rev. 06, 2019-06.	3-25,000 ng/L	3-2,000,000 ng/L
Homocysteine	HOMCYS		3.7-13.9 μ mol/L		Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Homocysteine Package Insert 10995362_EN Rev. 04-2021.03	0.5-65.0 μ mol/L	0.5-130.0 μ mol/L

IgA	IgA	<p>0-30 days: 0-10 mg/dL 31-182 days: 5-42 mg/dL 183-365 days: 6-48 mg/dL 1-3 years: 15-111 mg/dL 4-6 years: 23-166 mg/dL 7-9 years: 28-180 mg/dL 10-12 years: 55-193 mg/dL 13-15 years: 62-241 mg/dL 16-18 years: 69-262 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p>	<p>0-30 days: ≤11 mg/dL 31-182 days: 5-40 mg/dL 183-365 days: 1-52 mg/dL 1-3 years: 9-137 mg/dL 4-6 years: 44-187 mg/dL 7-9 years: 58-204 mg/dL 10-12 years: 46-218 mg/dL 13-15 years: 29-251 mg/dL 16-18 years: 68-259 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p>	Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	10-700 mg/dL	10-14,000 mg/dL
IgG	IgG	<p>0-30 days: 162-872 mg/dL 31-182 days: 311-664 mg/dL 183-365 days: 325-647 mg/dL 1-3 years: 451-1,202 mg/dL 4-6 years: 560-1,319 mg/dL 7-9 years: 485-1,473 mg/dL 10-12 years: 586-1,609 mg/dL 13-15 years: 749-1,640 mg/dL 16-18 years: 804-1,817 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p>	<p>0-30 days: 197-833 mg/dL 31-182 days: 140-533 mg/dL 183-365 days: 130-821 mg/dL 1-3 years: 413-1,112 mg/dL 4-6 years: 468-1,328 mg/dL 7-9 years: 582-1,441 mg/dL 10-12 years: 685-1,620 mg/dL 13-15 years: 590-1,600 mg/dL 16-18 years: 522-1,703 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p>	Turbidimetry	OSUWMC Immunoglobulin Reference Range Study; Verified by OSUWMC Reference Interval Study 2021.	75-3,000 mg/dL	75-60,000 mg/dL
IgM	IgM	<p>0-29 days: 1-57 mg/dL 30-182 days: <128 mg/dL 183-365 days: <131 mg/dL 1-3 years: 35-184 mg/dL 4-6 years: 42-184 mg/dL 7-9 years: 30-165 mg/dL 10-12 years: 42-211 mg/dL 13-15 years: 34-225 mg/dL 16-18 years: 45-224 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p>	<p>0-29 days: 0-65 mg/dL 30-182 days: 6-80 mg/dL 183-365 days: 15-117 mg/dL 1-3 years: 30-146 mg/dL 4-6 years: 31-151 mg/dL 7-9 years: 21-140 mg/dL 10-12 years: 27-151 mg/dL 13-15 years: 26-184 mg/dL 16-18 years: 28-179 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p>	Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	20-500 mg/dL	20-50,000 mg/dL
Immunoglobulin IgG IgA IgM	IgA IgG IgM	<p>IgA 0-30 days: 0-10 mg/dL 31-182 days: 5-42 mg/dL 183-365 days: 6-48 mg/dL 1-3 years: 15-111 mg/dL 4-6 years: 23-166 mg/dL 7-9 years: 28-180 mg/dL 10-12 years: 55-193 mg/dL 13-15 years: 62-241 mg/dL 16-18 years: 69-262 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p> <p>IgG 0-30 days: 162-872 mg/dL 31-182 days: 311-664 mg/dL 183-365 days: 325-647 mg/dL 1-3 years: 451-1,202 mg/dL 4-6 years: 560-1,319 mg/dL 7-9 years: 485-1,473 mg/dL 10-12 years: 586-1,609 mg/dL 13-15 years: 749-1,640 mg/dL 16-18 years: 804-1,817 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p> <p>IgM 0-29 days: 1-57 mg/dL 30-182 days: <128 mg/dL 183-365 days: <131 mg/dL 1-3 years: 35-184 mg/dL 4-6 years: 42-184 mg/dL 7-9 years: 30-165 mg/dL 10-12 years: 42-211 mg/dL 13-15 years: 34-225 mg/dL 16-18 years: 45-224 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p>	<p>IgA 0-30 days: ≤11 mg/dL 31-182 days: 5-40 mg/dL 183-365 days: 1-52 mg/dL 1-3 years: 9-137 mg/dL 4-6 years: 44-187 mg/dL 7-9 years: 58-204 mg/dL 10-12 years: 46-218 mg/dL 13-15 years: 29-251 mg/dL 16-18 years: 68-259 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p> <p>IgG 0-30 days: 197-833 mg/dL 31-182 days: 140-533 mg/dL 183-365 days: 130-821 mg/dL 1-3 years: 413-1,112 mg/dL 4-6 years: 468-1,328 mg/dL 7-9 years: 582-1,441 mg/dL 10-12 years: 685-1,620 mg/dL 13-15 years: 590-1,600 mg/dL 16-18 years: 522-1,703 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p> <p>IgM 0-29 days: 0-65 mg/dL 30-182 days: 6-80 mg/dL 183-365 days: 15-117 mg/dL 1-3 years: 30-146 mg/dL 4-6 years: 31-151 mg/dL 7-9 years: 21-140 mg/dL 10-12 years: 27-151 mg/dL 13-15 years: 26-184 mg/dL 16-18 years: 28-179 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p>	Turbidimetry	See individual analytes	See individual analytes	See individual analytes
Iron	IRON	<p>0-30 days: 29-127 mcg/dL 31-365 days: 25-126 mcg/dL 1-3 years: 25-101 mcg/dL 4-6 years: 28-93 mcg/dL 7-9 years: 30-104 mcg/dL 10-12 years: 32-104 mcg/dL 13-15 years: 30-109 mcg/dL 16-18 years: 33-102 mcg/dL 19+ years: 40-174 mcg/dL</p>	<p>0-30 days: 32-112 mcg/dL 31-365 days: 27-109 mcg/dL 1-3 years: 29-91 mcg/dL 4-6 years: 25-115 mcg/dL 7-9 years: 27-96 mcg/dL 10-12 years: 28-112 mcg/dL 13-15 years: 26-110 mcg/dL 16-18 years: 27-138 mcg/dL 19+ years: 40-174 mcg/dL</p>	Colorimetric	Established by OSUWMC Reference Interval Study 2013, verified by OSUWMC Reference Interval study 2021.	10-1,000 mcg/dL	10-2,000 mcg/dL
Iron/Iron Binding/Transferrin	IRON TRAN TIBC	<p>IRON 0-30 days: 29-127 mcg/dL 31-365 days: 25-126 mcg/dL 1-3 years: 25-101 mcg/dL 4-6 years: 28-93 mcg/dL 7-9 years: 30-104 mcg/dL 10-12 years: 32-104 mcg/dL 13-15 years: 30-109 mcg/dL 16-18 years: 33-102 mcg/dL 19+ years: 40-174 mcg/dL</p> <p>TRAN 200-400 mg/dL</p> <p>TIBC 0-30 days: 94-236 mcg/dL 31-182 days: 89-311 mcg/dL 183-365 days: 138-365 mcg/dL 1-3 years: 184-377 mcg/dL 4-6 years: 162-352 mcg/dL 7-9 years: 167-336 mcg/dL 10-12 years: 198-383 mcg/dL 13-15 years: 169-358 mcg/dL 16-18 years: 194-372 mcg/dL 19+ years: 250-425 mcg/dL</p>	<p>IRON 0-30 days: 32-112 mcg/dL 31-365 days: 27-109 mcg/dL 1-3 years: 29-91 mcg/dL 4-6 years: 25-115 mcg/dL 7-9 years: 27-96 mcg/dL 10-12 years: 28-112 mcg/dL 13-15 years: 26-110 mcg/dL 16-18 years: 27-138 mcg/dL 19+ years: 40-174 mcg/dL</p> <p>TRAN 200-400 mg/dL</p> <p>TIBC 0-30 days: 94-232 mcg/dL 31-182 days: 116-322 mcg/dL 183-365 days: 176-384 mcg/dL 1-3 years: 204-382 mcg/dL 4-6 years: 180-390 mcg/dL 7-9 years: 183-369 mcg/dL 10-12 years: 173-356 mcg/dL 13-15 years: 193-377 mcg/dL 16-18 years: 174-351 mcg/dL 19+ years: 250-425 mcg/dL</p>	IRON Colorimetric TRAN In the procedure, the measurement of the decrease in light transmitted (increase in absorbance) through particles suspended in solution as a result of complexes formed during the antigen-antibody reaction, is the basis of this assay. IRON SATURATION (Iron/TIBC)*100	Transferrin: Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Soldin, 1999 Iron: Established by OSUWMC Reference Interval Study 2013, verified by OSUWMC Reference Interval study 2021.	TRANSFERRIN: 75-750 mg/dL IRON: 10-1,000 mcg/dL	TRANSFERRIN: 75-2,250 mg/dL IRON: 10-2,000 mcg/dL
Lactate Dehydrogenase	LDH	<p>IRON SATURATION 20-55%</p> <p>0-30 days: 145-765 U/L 31-365 days: 190-420 U/L 1-3 years: 165-395 U/L 4-6 years: 135-345 U/L 7-9 years: 140-280 U/L 10-12 years: 120-260 U/L 13-15 years: 100-235 U/L 16-18 years: 105-230 U/L 19+ years: 100-190 U/L</p>	<p>IRON SATURATION 20-55%</p> <p>0-30 days: 125-735 U/L 31-365 days: 170-450 U/L 1-3 years: 155-345 U/L 4-6 years: 155-345 U/L 7-9 years: 145-300 U/L 10-12 years: 120-325 U/L 13-15 years: 120-290 U/L 16-18 years: 105-235 U/L 19+ years: 100-190 U/L</p>	Photometric rate	Clinical Guide to Laboratory Tests, Tietz, 1995; verified by OSUWMC Reference Interval Study 2021. Pediatric Reference Ranges, Soldin, 1999.	25-1,200 U/L	25-60,000 U/L

Lactate Dehydrogenase Body Fluid	FLD			<p>Pleural: A Pleural fluid LDH to serum/plasma LDH ratio > 0.6 or a pleural fluid LDH concentration > two-thirds the upper limit of the serum/plasma LDH reference interval suggest an exudate.</p> <p>Pericardial: Pericardial fluid LDH to serum/plasma LDH ratio > 0.6 or > 300 U/L suggests an exudate.</p> <p>Peritoneal Ascites: Peritoneal fluid LDH to serum/plasma LDH ratio > 0.6 is consistent with an exudate.</p> <p>CSF: Elevated LDH in CSF specimens may indicate a non-specific immune process. CSF LDH measurements above 40 U/L may be associated with Jakob-Creutzfeldt Disease, Bacterial Meningitis, Neurosyphilis, or tumors of the central nervous system. Contamination of red blood cells can falsely increase LDH measurements.</p> <p>Amniotic Fluid: The reference range has not been established for this fluid type. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.</p>	Photometric rate	<p>CSF: Clinical Utility of Biochemical Analysis of Cerebrospinal Fluid Clinical Chemistry 1995 Watson MA</p> <p>Pericardial and peritoneal: Burgess. Clinical Chemical Acta 343 (2004) 61-84.</p> <p>Pleural Light, RW. N Engl J Med. 2002 Jun 20;346(25):1971-1977.</p>	25-1,200 U/L	25-30,000 U/L
Lactate, Blood	LACT		0.5-1.6 mmol/L		Enzymatic colorimetric		RADDOMETER: 0.0-30.0 mmol/L	RADDOMETER: 0.0-30.0 mmol/L
			0.5-2.2 mmol/L				DXC: 0.2-10.0 mmol/L	DXC: 0.2-30.0 mmol/L
Lactate, CSF	CSLACT		<2.8 mmol/L		Enzymatic colorimetric	Beckman Coulter literature which cites Clinical Guide to Laboratory Tests, Tetr, 1995	0.2-10.0 mmol/L	0.2-30.0 mmol/L
Lactate, Fluid	FLACT			The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.	Enzymatic colorimetric		0.2-10.0 mmol/L	0.2-30.0 mmol/L
LDL, Direct Measure	LDL		<100 mg/dL		Enzymatic colorimetric	National Cholesterol Education Project (NCEP) Adult Treatment Protocol (ATP, III) (Circulation, 2002;106:3143-3421)	7-400 mg/dL	7-400 mg/dL
LH	LH	0-365 days: 0.01-1.00 mIU/mL 1-10 years: 0.00-0.10 mIU/mL 11-13 years: 0.37-6.23 mIU/mL 14-18 years: 0.37-17.30 mIU/mL	0-365 days: <3.24 mIU/mL 1-10 years: 0.00-0.10 mIU/mL 11-13 years: 0.08-2.58 mIU/mL 14-18 years: 0.97-6.37 mIU/mL	<p>Male:</p> <p>19-70 years: 1.5-9.3 mIU/mL >70 years: 3.1-4.6 mIU/mL</p> <p>19+ years Female:</p> <p>Follicular phase: 1.9-12.5 mIU/mL Midcycle peak: 8.7-76.3 mIU/mL Luteal phase: 0.5-16.9 mIU/mL</p> <p>Pregnant: <0.1-1.5 mIU/mL Postmenopausal: 15.9-54.0 mIU/mL Caucasian: 0.5-2.6 mIU/mL</p>	Two-site sandwich immunoassay chemiluminescent	Atellica IM LH Package Insert 11200385_EN Rev. 04-2020-06; Pediatric Reference Ranges, Saldin, 1999	0.07-200.00 mIU/mL	0.07-6,400.00 mIU/mL
Lipase	LIPA		0-30 days: 6-55 U/L 31-182 days: 4-29 U/L 183-365 days: 4-23 U/L 1-3 years: 4-31 U/L 4-9 years: 3-32 U/L 10-18 years: 4-29 U/L 19+ years: 11-82 U/L		Photometric rate	Beckman Coulter Chemistry Information Sheet, 9/2020, verified by OSELWAC Reference Interval Study 2021. Pediatric Reference Ranges, 1999.	6-600 U/L	6-6,000 U/L
Lipid Panel w Calculated LDL	HDLT	<p>CHOL</p> <p>0-30 days: 62-155 mg/dL 31-182 days: 62-141 mg/dL 183-365 days: 76-216 mg/dL 1-3 years: 108-193 mg/dL 4-6 years: 106-193 mg/dL 7-9 years: 104-210 mg/dL 10-12 years: 105-218 mg/dL 13-15 years: 108-205 mg/dL 16-18 years: 92-234 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG</p> <p>0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 29-140 mg/dL 12-13 years: 37-130 mg/dL 14-15 years: 38-135 mg/dL 16-19 years: 37-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL</p> <p>0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL CALC</p> <p>0-59 days: <130 mg/dL 60-365 days: 32-117 mg/dL 1-2 years: 38-140 mg/dL 2+ years: 0-99 mg/dL</p> <p>TOT CHOL/HDL</p> <p><4.5</p> <p>NON HDL</p> <p><130 mg/dL</p>	<p>CHOL</p> <p>0-30 days: 54-151 mg/dL 31-182 days: 81-147mg/dL 183-365 days: 76-179 mg/dL 1-3 years: 85-182 mg/dL 4-6 years: 110-217 mg/dL 7-9 years: 110-211 mg/dL 10-12 years: 105-223 mg/dL 13-15 years: 91-204 mg/dL 16-18 years: 82-192 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG</p> <p>0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 28-137 mg/dL 12-13 years: 24-145 mg/dL 14-15 years: 34-165 mg/dL 16-19 years: 34-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL</p> <p>0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL CALC</p> <p>0-59 days: <130 mg/dL 60-365 days: 32-117 mg/dL 1-2 years: 38-140 mg/dL 2+ years: 0-99 mg/dL</p> <p>TOT CHOL/HDL</p> <p><4.5</p> <p>NON HDL</p> <p><130 mg/dL</p>	<p>CHOL</p> <p>Enzymatic colorimetric</p> <p>TRIG</p> <p>Enzymatic colorimetric</p> <p>HDL</p> <p>Enzymatic colorimetric</p> <p>LDL CALC</p> <p>T Cholesterol - HDL - (Trig/5)</p>	See individual analytes	See individual analytes	See individual analytes	

Lipid Panel with Reflex to Measured LDL	LIPD	<p>CHOL 0-30 days: 62-155 mg/dL 31-182 days: 62-141 mg/dL 183-365 days: 76-216 mg/dL 1-3 years: 108-193 mg/dL 4-6 years: 106-193 mg/dL 7-9 years: 104-210 mg/dL 10-12 years: 105-218 mg/dL 13-15 years: 108-205 mg/dL 16-18 years: 92-234 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG 0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 39-140 mg/dL 12-13 years: 37-130 mg/dL 14-15 years: 34-165 mg/dL 16-19 years: 37-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL 0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL CALC 0-9 years: <130 mg/dL 10-13 years: 32-117 mg/dL 14-19 years: 38-140 mg/dL 2+ years: 0-99 mg/dL</p> <p>TOT CHOL/HDL ≤4.5</p> <p>NON HDL ≤130 mg/dL</p>	<p>CHOL 0-30 days: 54-151 mg/dL 31-182 days: 81-147mg/dL 183-365 days: 76-179 mg/dL 1-3 years: 85-182 mg/dL 4-6 years: 110-217 mg/dL 7-9 years: 110-217 mg/dL 10-12 years: 105-223 mg/dL 13-15 years: 91-204 mg/dL 16-18 years: 82-192 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG 0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 24-137 mg/dL 12-13 years: 24-145 mg/dL 14-15 years: 34-165 mg/dL 16-19 years: 34-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL 0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL CALC 0-9 years: <130 mg/dL 10-13 years: 32-117 mg/dL 14-19 years: 38-140 mg/dL 2+ years: 0-99 mg/dL</p> <p>TOT CHOL/HDL ≤4.5</p> <p>NON HDL ≤130 mg/dL</p>		<p>CHOL Enzymatic colorimetric</p> <p>TRIG Enzymatic colorimetric</p> <p>HDL Enzymatic colorimetric</p> <p>LDL CALC T Cholesterol - HDL - (Trig/5)</p> <p>LDL DIRECT Enzymatic colorimetric</p>	See individual analytes	See individual analytes	See individual analytes
Lipids w Direct Measure LDL	HDLTB	<p>CHOL 0-30 days: 62-155 mg/dL 31-182 days: 62-141 mg/dL 183-365 days: 76-216 mg/dL 1-3 years: 108-193 mg/dL 4-6 years: 106-193 mg/dL 7-9 years: 104-210 mg/dL 10-12 years: 105-218 mg/dL 13-15 years: 108-205 mg/dL 16-18 years: 92-234 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG 0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 39-140 mg/dL 12-13 years: 37-130 mg/dL 14-15 years: 38-135 mg/dL 16-19 years: 37-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL 0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL DIRECT ≤100 mg/dL</p> <p>TOT CHOL/HDL ≤4.5</p> <p>NON HDL ≤130 mg/dL</p>	<p>CHOL 0-30 days: 54-151 mg/dL 31-182 days: 81-147mg/dL 183-365 days: 76-179 mg/dL 1-3 years: 85-182 mg/dL 4-6 years: 110-217 mg/dL 7-9 years: 110-211 mg/dL 10-12 years: 105-223 mg/dL 13-15 years: 91-204 mg/dL 16-18 years: 82-192 mg/dL 19+ years: <200 mg/dL</p> <p>TRIG 0-3 years: 27-125 mg/dL 4-6 years: 32-116 mg/dL 7-9 years: 28-129 mg/dL 10-11 years: 24-137 mg/dL 12-13 years: 24-145 mg/dL 14-15 years: 34-165 mg/dL 16-19 years: 34-140 mg/dL 20+ years: <150 mg/dL</p> <p>HDL 0-9 years: 35-82 mg/dL 10-13 years: 36-84 mg/dL 14-19 years: 35-65 mg/dL 20+ years: >40 mg/dL</p> <p>LDL DIRECT ≤100 mg/dL</p> <p>TOT CHOL/HDL ≤4.5</p> <p>NON HDL ≤130 mg/dL</p>		<p>CHOL Enzymatic colorimetric</p> <p>TRIG Enzymatic colorimetric</p> <p>HDL Enzymatic colorimetric</p> <p>LDL DIRECT Enzymatic colorimetric</p>	See individual analytes	See individual analytes	See individual analytes
Lithium Level	LI		Therapeutic Range: 0-59 years: 0.60-1.20 mmol/L 60+ years: 0.40-0.80 mmol/L		A spectrophotometric method which can be readily adapted to automated clinical chemistry analyzers. Lithium present in the sample reacts with a substituted porphyrin compound at an alkaline pH, resulting in a change in absorbance which is directly proportional to the concentration of Lithium in the sample.	Applied Clinical Pharmacokinetics, Baser, 2001; Clinical Pharmacokinetics, Eilers, 1995;29:442-50 Bipolar Disord. 2019 Mar;21(2):117-123. Bipolar Disord. 2019 May;21(3):90-191.	0.10-5.00 mmol/L	0.10-5.00 mmol/L
Lytes (Na, K, Cl) - Urine - Random	ULYTR			The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.	NAKCL Indirect ion-selective electrode	See individual analytes	See individual analytes	See individual analytes
Lytes (Na, K, Cl) - Urine 24HR	ULYT				NAKCL Indirect ion-selective electrode	See individual analytes		

Phenobarbital Level, Random			Therapeutic Range: 15.0-40.0 mcg/mL		The assay is based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Applied Clinical Pharmacokinetics, 2001	5.0-80.0 mcg/mL	5.0-240.0 mcg/mL	
	PHNOR								
Phenobarbital Level, Trough (Pre Drug Level)			Therapeutic Range: 15.0-40.0 mcg/mL		The assay is based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Applied Clinical Pharmacokinetics, 2001	5.0-80.0 mcg/mL	5.0-240.0 mcg/mL	
	PHNO								
Phenytoin Total Level			Therapeutic Range: 10.0-20.0 mcg/mL		The assay is based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Applied Clinical Pharmacokinetics, 2001	2.5-40.0 mcg/mL	2.5-200.0 mcg/mL	
	PTN								
Phosphate, Inorganic		0-30 days: 4.3-7.7 mg/dL 31-365 days: 3.7-6.5 mg/dL 1-3 years: 3.4-6.0 mg/dL 4-6 years: 3.2-5.5 mg/dL 7-9 years: 3.1-5.5 mg/dL 10-12 years: 3.3-5.3 mg/dL 13-15 years: 2.8-4.8 mg/dL 16-18 years: 2.5-4.8 mg/dL 19+ years: 2.7-4.6 mg/dL	0-30 days: 3.9-6.9 mg/dL 31-365 days: 3.5-6.6 mg/dL 1-3 years: 3.1-6.0 mg/dL 4-6 years: 3.3-5.6 mg/dL 7-9 years: 3.0-5.4 mg/dL 10-12 years: 3.2-5.7 mg/dL 13-15 years: 2.8-5.1 mg/dL 16-18 years: 2.7-4.9 mg/dL 19+ years: 2.7-4.6 mg/dL			Photometric	OSUWMC Reference Range Study effective 12/11/2013; verified by OSUWMC Reference Interval Study 2021. Pediatric Reference Ranges, Soldin, 1999.	1.0-20.0 mg/dL	1.0-60.0 mg/dL
	IP								
Phosphorus, Random, Urine				The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.		Photometric		10.0-200.0 mg/dL	10.0-1,000.0 mg/dL
	UPR								
Phosphorus, 24HR			0.4-1.3 g/24 hrs			Photometric	Clinical Guide to Laboratory Tests, Tietz, 1995		
	UIP								
Potassium		0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-7 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L				Indirect ion-selective electrode	OSUWMC Reference Range Study effective 12/11/2013; verified by OSUWMC Reference Interval Study 2021. Pediatric Reference Ranges, Soldin, 1999.	1.0-10.0 mmol/L	1.0-10.0 mmol/L
	K								
Potassium Body Fluid				Note: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.		Indirect ion-selective electrode		2.0-200.0 mmol/L	2.0-200.0 mmol/L
	FK								
Potassium 24 HR Urine			25-125 mmol/24 hrs			Indirect ion-selective electrode	Clinical Guide to Laboratory Tests, Tietz, 1995		
	UK								
Potassium, Random Urine				The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.		Indirect ion-selective electrode	Clinical Guide to Laboratory Tests, Tietz, 1995	2.0-200.0 mmol/L	2.0-200.0 mmol/L
	UKR								
Prealbumin			17-34 mg/dL			Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	3-80 mg/dL	3-1,600 mg/dL
	PALB								
Procalcitonin			<0.50 ng/mL			Two-site sandwich immunoassay chemiluminescent.	Atellica IM Procalcitonin 11200767_EN Rev. 03, 2019-06.	0.04-50.00 ng/mL	0.04-2,000.00 ng/mL
	PROCAL								
Progesterone					Male: 0.28-1.22 ng/mL Female: Follicular phase: 0.1-40 ng/mL Luteal phase: 3.34-25.56 ng/mL Midluteal phase: 4.44-29.03 ng/mL Postmenopausal: 0-0.73 ng/mL	Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Progesterone Package Insert 11200356_EN Rev. 04-2020-06	0.21-60.00 ng/mL	0.21-3,000.00 ng/mL
	PROG								
Prolactin					Male: 2.1-17.7 ng/mL Female: Nonpregnant: 2.8-29.2 ng/mL Pregnant: 9.7-208.5 ng/mL Postmenopausal: 1.8-26.3 ng/mL <2 years: 3.3-14.7 ng/mL 2-5 years: 1.0-12.8 ng/mL 6-10 years: 1.2-11.4 ng/mL 11-17 years: 1.4-14.3 ng/mL	Two-site sandwich immunoassay chemiluminescent.	Advia Centaur Prolactin Package Insert 111746 Rev. N, 2008-09; Pediatric Reference Intervals, 5th ed Soldin, 2005	0.3-200.0 ng/mL	0.3-800,000.0 ng/mL
	PROL								
Protein & Glucose, CSF			GLUCSF 40-70 mg/dL CFP 0-3 days: 40-120 mg/dL 4-30 days: 20-80 mg/dL 31+ days: 15-45 mg/dL			GLUCSF Photometric rate with hexokinase CFP Colorimetric with Pyrogallol red	See individual analytes	See individual analytes	See individual analytes
	CFPG CFP								
Protein Total		0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL	0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL			Colorimetric with cupric ions in an alkaline solution	Clinical Guide to Laboratory Tests, Tietz, 1995; verified by OSUWMC Reference Interval Study 2021. Pediatric Reference Ranges, Soldin, 1999	3.0-12.0 g/dL	3.0-24.0 g/dL
	TP								
Protein, CSF			0-3 days: 40-120 mg/dL 4-30 days: 20-80 mg/dL			Colorimetric with Pyrogallol red	Clinical Guide to Laboratory Tests, Tietz, 1995; Pediatric Reference Ranges, Soldin, 1999	4-200 mg/dL	4-5,000 mg/dL
	CFP								
Protein, Fluid				Plural: Pleural fluid protein to serum/plasma protein ratio > 0.5 are consistent with exudates. Pericardial: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation. Peritoneal/Ascites: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.		Colorimetric with cupric ions in an alkaline solution	4th Edition of Tietz	0.5-12.0 g/dL	0.5-24.0 g/dL
	FLP								
Protein, 24 HR Urine			40-225 mg/24 hrs			Colorimetric with Pyrogallol red	Clinical Guide to Laboratory Tests, Tietz, 1995		
	UPRO								
Protein, Random Urine				The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.		Colorimetric with Pyrogallol red		4-200 mg/dL	4-5,000 mg/dL
	UPROR								
PSA - Diagnostic/Tumor Marker			≤4.00 ng/mL			Two-site sandwich immunoassay chemiluminescent.	Atellica IM PSA Package Insert 10997799_EN Rev. 03-2019-09.	0.04-100.00 ng/mL	0.04-6,400,000.00 ng/mL
	PSATM								
PSA, Reflex to Free and Total PSA			≤4.00 ng/mL			Two-site sandwich immunoassay chemiluminescent.	Atellica IM PSA Package Insert 10997799_EN Rev. 03-2019-09.	0.04-100.00 ng/mL	0.04-6,400,000.00 ng/mL
	PSAR								

PSA, Screening	PSA	≤4.00 ng/mL		Two-site sandwich immunoassay chemiluminescent	Aetlica IM PSA Package Insert 10997799_EN Rev. 03-2019-09	0.04-100.00 ng/mL	0.04-6,400,000.00 ng/mL
Renal Panel	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-0.90 mg/dL 31-365 days: 0.40-0.60 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.50-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.70-1.10 mg/dL 16-18 years: 0.80-1.20 mg/dL 19+ years: 0.50-1.20 mg/dL</p> <p>CA 0-30 days: 8.4-10.6 mg/dL 31-365 days: 8.9-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p> <p>ALB 0-30 days: 2.7-4.3 g/dL 31-182 days: 2.9-4.2 g/dL 183-365 days: 3.3-4.8 g/dL 1-19 years: 2.9-4.2 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>IP 0-30 days: 4.3-7.7 mg/dL 31-365 days: 3.7-6.5 mg/dL 1-3 years: 3.4-6.0 mg/dL 4-6 years: 3.2-5.5 mg/dL 7-9 years: 3.1-5.5 mg/dL 10-12 years: 3.3-5.3 mg/dL 13-15 years: 2.8-4.8 mg/dL 16-18 years: 2.5-4.8 mg/dL 19+ years: 2.2-4.6 mg/dL</p>	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-30 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1-17 years: 3.3-4.6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+ years: 98-108 mmol/L</p> <p>CO2 0-2 years: 13-29 mmol/L 3+ years: 21-31 mmol/L</p> <p>GLUC 0-30 days: 55-115 mg/dL 31-182 days: 57-117 mg/dL 183-365 days: 70-126 mg/dL 1+ years: 70-99 mg/dL</p> <p>BUN 7-25 mg/dL</p> <p>CREA 0-30 days: 0.50-1.20 mg/dL 31-365 days: 0.40-0.70 mg/dL 1-3 years: 0.40-0.70 mg/dL 4-6 years: 0.50-0.80 mg/dL 7-9 years: 0.60-0.90 mg/dL 10-12 years: 0.60-1.00 mg/dL 13-15 years: 0.60-1.20 mg/dL 16-18 years: 0.80-1.40 mg/dL 19+ years: 0.70-1.30 mg/dL</p> <p>CA 0-30 days: 8.5-10.6 mg/dL 31-365 days: 8.7-10.5 mg/dL 1+ years: 8.6-10.5 mg/dL</p> <p>OSMO 278-305 mOsm/kg</p> <p>ANION GAP 7-17 mmol/L</p> <p>eGFR ≥60 mL/min/1.73m²</p> <p>ALB 0-30 days: 2.6-4.1 g/dL 31-182 days: 2.8-4.6 g/dL 183-365 days: 2.8-4.8 g/dL 1-19 years: 2.4-7.0 g/dL 19+ years: 3.5-5.0 g/dL</p> <p>IP 0-30 days: 3.9-6.9 mg/dL 31-365 days: 3.5-6.6 mg/dL 1-3 years: 3.1-6.0 mg/dL 4-6 years: 3.3-5.6 mg/dL 7-9 years: 3.0-5.4 mg/dL 10-12 years: 3.2-5.7 mg/dL 13-15 years: 2.9-5.1 mg/dL 16-18 years: 2.7-4.9 mg/dL 19+ years: 2.2-4.6 mg/dL</p>		<p>NA/K/CL Indirect ion-selective electrode</p> <p>CO2 Photometric</p> <p>GLUC Photometric rate with hexokinase</p> <p>BUN Photometric rate</p> <p>CREA Kinetic Jaffe</p> <p>CA Photometric, arsenazo</p> <p>BC RATIO BUN/Serum Creatinine</p> <p>ANION GAP (Na + K) - (Cl + CO2)</p> <p>eGFR eGFR = $142 \times \text{min}(\text{Scr}/x, 1) \times \text{max}(\text{Scr}/x, 1) - 1.200 \times 0.9938 \text{Age} \times 1.012$ [if female]</p> <p>Where: $x = -0.7$ (females) or 0.9 (males) $a = -0.241$ (female) or -0.302 (male) Scr = serum creatinine in mg/dL; divide by 88.4 for creatinine in μmol/L Age (years) The "min(Scr, 1)" factor indicates the minimum of Scr or 1.0 and "max(Scr, 1)" indicates the maximum of Scr or 1.0.</p> <p>OSMO CALC $1.86(\text{Na} + \text{K}) + 1.15(\text{Glucose}) + 1.43(\text{Urea}) + 14$ where Na and K are in mmol/L; Glucose and Urea are in mg/dL.</p> <p>ALB Colorimetric; Bromocresol green</p> <p>IP Colorimetric</p>	See individual analytes	See individual analytes	See individual analytes
Rheumatoid Factor	RENAL	≤14 IU/mL		Turbidimetry	Package Insert. Verified by OSUWMC Reference Interval Study 2021.	10-120 IU/mL	10-3,000 IU/mL
Salicylate Level	RF	Therapeutic Range: 20.0-30.0 mg/dL		Serum is mixed with Reagent 1, which contains antibodies to salicylic acid and the coenzyme nicotinamide adenine dinucleotide (NAD). Subsequently, Reagent 2, which contains salicylic acid labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH), is added. Salicylic acid in the sample and salicylic acid-labeled G6PDH compete for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the salicylic acid concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized NAD to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Applied Pharmacokinetics: Principles of Therapeutic Drug Monitoring, 2nd Edition 2002 Applied Therapeutics, Inc. and Micromedex. On OSU Intranet	5.0-80.0 mg/dL	5.0-240.0 mg/dL
Sodium	SALI	0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L		Indirect ion-selective electrode	Verified by OSUWMC Reference Interval Study 2021.	50-200 mmol/L	50-200 mmol/L
Sodium Body Fluid	NA		Stand: The reference range has not been established for this fluid specimen. The fluid results should be compared with the concentration in serum/plasma or integrated into the clinical context for interpretation.	Indirect ion-selective electrode		50-200 mmol/L	50-200 mmol/L
	PNA						

Sodium, Potassium, Chloride	NA, K, CL	<p>NA 0-6 days: 133-146 mmol/L 7-30 days: 134-144 mmol/L 31-182 days: 134-142 mmol/L 183-365 days: 133-142 mmol/L 1+ years: 135-145 mmol/L</p> <p>K 0-7 days: 3.2-5.5 mmol/L 8-36 days: 3.4-6.0 mmol/L 31-182 days: 3.5-5.6 mmol/L 183-365 days: 3.5-6.1 mmol/L 1+7 years: 3.4-6 mmol/L 18+ years: 3.5-5.0 mmol/L</p> <p>CL 0-365 days: 98-113 mmol/L 1-17 years: 102-112 mmol/L 18+: 98-109 mmol/L</p>	Indirect ion-selective electrode	See individual analytes	See individual analytes	See individual analytes		
Sodium, 24 HR Urine	UNA	40-220 mmol/24 hrs		The reference range has not been established for random urine specimens. The test result should be integrated into the clinical context for interpretation.	Indirect ion-selective electrode	Clinical Guide to Laboratory Tests, Tests, 1995		
Sodium, Random Urine	UNAR				Indirect ion-selective electrode		10-400 mmol/L	10-400 mmol/L
T3 Free	FT3	<p>0-3 days: 1.4-5.4 pg/mL 4-29 days: 1.5-5.0 pg/mL 30-365 days: 2.5-6.5 pg/mL 1-5 years: 3.0-6.0 pg/mL 6-10 years: 2.7-6.2 pg/mL 11-15 years: 2.6-5.7 pg/mL 16-18 years: 2.8-5.2 pg/mL 19+ years: 2.4-2.9 pg/mL</p>	<p>0-3 days: 1.4-4.8 pg/mL 4-29 days: 1.4-5.5 pg/mL 30-365 days: 2.0-6.9 pg/mL 1-5 years: 2.4-6.7 pg/mL 6-10 years: 2.9-6.0 pg/mL 11-15 years: 3.1-5.9 pg/mL 16-18 years: 3.5-5.7 pg/mL 19+ years: 2.3-4.2 pg/mL</p>	Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Free T3 Package Insert 10995347_EN Rev. 03-2020-06	0.2-20.0 pg/mL	0.2-20.0 pg/mL	
T3 Total (Triiodothyronine)	TT3	<p>0-29 days: 0.8-1.7 ng/mL 30-365 days: 0.73-2.21 ng/mL 1-5 years: 1.26-2.16 ng/mL 6-10 years: 1.10-1.05 ng/mL 11-15 years: 1.04-1.84 ng/mL 16-18 years: 1.01-1.51 ng/mL 19+ years: 0.60-1.41 ng/mL</p>	<p>0-29 days: 0.71-1.84 ng/mL 30-365 days: 1.03-2.29 ng/mL 1-5 years: 0.93-2.13 ng/mL 6-10 years: 1.04-1.09 ng/mL 11-15 years: 0.88-1.76 ng/mL 16-18 years: 0.86-1.78 ng/mL 19+ years: 0.60-1.81 ng/mL</p>	Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Total T3 Package Insert 10995424_EN Rev. 03-2020-06	0.10-8.00 ng/mL	0.10-80.00 ng/mL	
T4	TT4	<p>0-29 days: 3.4-11.8 mcg/dL 30-365 days: 4.7-11.8 mcg/dL 1-5 years: 6.0-11.3 mcg/dL 6-10 years: 5.2-9.7 mcg/dL 11-15 years: 4.9-9.0 mcg/dL 16-18 years: 5.1-9.0 mcg/dL 19+ years: 4.5-10.9 mcg/dL</p>	<p>0-29 days: 3.4-12.6 mcg/dL 30-365 days: 5.4-14.1 mcg/dL 1-5 years: 5.3-10.2 mcg/dL 6-10 years: 5.3-9.5 mcg/dL 11-15 years: 4.6-9.2 mcg/dL 16-18 years: 4.8-11.1 mcg/dL 19+ years: 4.5-10.9 mcg/dL</p>	Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Total T4 Package Insert 10995425_EN Rev. 03-2020-06; Pediatric Reference Ranges, Soldin, 1999	0.4-30.0 mcg/dL	0.4-300.0 mcg/dL	
T4 Free	FT4	<p>0-18 years: 1.04-1.64 ng/dL 19+ years: 0.89-1.76 ng/dL</p>		Competitive immunoassay using direct chemiluminescent technology.	Atellica IM Free T4 Package Insert 10995348_EN Rev. 06-2020-11; Pediatric Reference Ranges, Soldin, 1999	0.10-12.00 ng/dL	0.10-12.00 ng/dL	
Testosterone	TESTOS	<p>0-6 months: 20-80 ng/dL 6 months to 9 years: <20 ng/dL 10-11 years: 544 ng/dL 12-16 years: 575 ng/dL 17-18 years: 20-75 ng/dL 19+ years: 8-60 ng/dL</p>	<p>0-6 months: 35-400 ng/dL 6 months to 9 years: <20 ng/dL 10-11 years: <130 ng/dL 12-13 years: <800 ng/dL 14-15 years: <1,200 ng/dL 15-16 years: 100-1,200 ng/dL 17-18 years: 300-1,200 ng/dL 19+ years: 240-950 ng/dL</p>	Competitive immunoassay using direct chemiluminescent technology.	Siemens Atellica IM Reference Interval Verification Study 2023	7-1,500 ng/dL	7-3,000 ng/dL	
Testosterone Total and Free, Includes SHBG	TESTOS SHGB	<p>0-6 months: 20-80 ng/dL 6 months to 9 years: <20 ng/dL 10-11 years: 544 ng/dL 12-16 years: 575 ng/dL 17-18 years: 20-75 ng/dL 19+ years: 8-60 ng/dL</p> <p>SHGB 18.00-144.00 nmol/L</p> <p>FREE TESTOS 0-365 days: <0.25 ng/dL 1-6 years: <0.13 ng/dL 7-8 years: <0.23 ng/dL 8-9 years: <0.34 ng/dL 9-10 years: <0.46 ng/dL 10-11 years: <0.59 ng/dL 11-12 years: <0.72 ng/dL 12-13 years: <0.84 ng/dL 13-14 years: <0.96 ng/dL 14-15 years: <1.06 ng/dL 15-18 years: <1.09 ng/dL 19+ years: 0.00-1.05 ng/dL</p>	<p>0-6 months: 35-400 ng/dL 6 months to 9 years: <20 ng/dL 10-11 years: <130 ng/dL 12-13 years: <800 ng/dL 14-15 years: <1,200 ng/dL 15-16 years: 100-1,200 ng/dL 17-18 years: 300-1,200 ng/dL 19+ years: 240-950 ng/dL</p> <p>SHGB 10.00-57.00 nmol/L</p> <p>FREE TESTOS 0-365 days: 0.20-3.10 ng/dL 1-8 years: <0.13 ng/dL 9-10 years: <0.45 ng/dL 10-11 years: <1.26 ng/dL 11-12 years: <5.52 ng/dL 12-13 years: <9.28 ng/dL 13-14 years: <12.60 ng/dL 14-15 years: 0.48-15.30 ng/dL 15-16 years: 1.62-17.70 ng/dL 16-17 years: 2.93-19.50 ng/dL 17-18 years: 4.28-20.90 ng/dL 18-19 years: 5.40-21.80 ng/dL 19+ years: 2.20-20.70 ng/dL</p>		See individual analytes	See individual analytes	See individual analytes	
Theophylline Level	THEO	Therapeutic Range: 5.0-20.0 mcg/mL			Based on competition between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH) for antibody binding sites. Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Applied Clinical Pharmacokinetics, 2001	2.5-40.0 mcg/mL	2.5-200.0 mcg/mL
Tobramycin Level, Extended Interval	TOBREI	Peak: 10.0 - 15.0 mcg/mL, Trough: <1.0 mcg/mL			This assay is based on competition for antibody binding sites between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH). Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Antimicrobial Stewardship Program, 2013	0.6-10.0 mcg/mL	0.6-50.0 mcg/mL
Tobramycin Level, Peak (Post Drug Level)	TOBRPK	Therapeutic Range: 10.0-15.0 mcg/mL			This assay is based on competition for antibody binding sites between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH). Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Antimicrobial Stewardship Program, 2013	0.6-10.0 mcg/mL	0.6-50.0 mcg/mL
Tobramycin Level, Random	TOBR	Peak: 10.0 - 15.0 mcg/mL, Trough: <1.0 mcg/mL			This assay is based on competition for antibody binding sites between drug in the sample and drug labeled with the enzyme glucose-6-phosphate dehydrogenase (G6PDH). Enzyme activity decreases upon binding to the antibody, so the drug concentration in the sample can be measured in terms of enzyme activity. Active enzyme converts oxidized nicotinamide adenine dinucleotide (NAD) to NADH, resulting in an absorbance change that is measured spectrophotometrically.	Antimicrobial Stewardship Program, 2013	0.6-10.0 mcg/mL	0.6-50.0 mcg/mL

		<p>MCH 0-14 days: 31.1-35.9 pg 15-30 days: 30.4-35.3 pg 31-60 days: 28.0-32.5 pg 61-179 days: 24.4-29.5 pg 180 days-2 years: 23.2-27.5 pg 2-5 years: 23.7-28.6 pg 6-11 years: 24.8-29.7 pg 12-17 years: 24.8-30.2 pg 18+ years: 25.9-33.9 pg</p> <p>MCHC 0-14 days: 33.4-35.4 g/dL 15-30 days: 33.2-35.0 g/dL 31-60 days: 32.5-34.9 g/dL 61-179 days: 32.1-34.4 g/dL 180 days-2 years: 31.9-34.2 g/dL 2-5 years: 31.8-34.6 g/dL 6-11 years: 31.8-34.6 g/dL 12-17 years: 31.5-34.2 g/dL 18+ years: 31.4-35.9 g/dL</p> <p>RDW 0-14 days: 14.6-17.3 % 15-30 days: 14.4-16.2 % 31-60 days: 13.6-15.8 % 61-179 days: 12.2-14.3 % 180 days-2 years: 12.7-15.1 % 2-5 years: 12.4-14.9 % 6-11 years: 12.2-14.4 % 12-17 years: 12.3-14.6 % 18+ years: 10.8-14.9 %</p> <p>PLT 0-14 days: 144-449 K/uL 15-30 days: 279-571 K/uL 31-60 days: 331-597 K/uL 61-179 days: 247-580 K/uL 180 days-2 years: 214-459 K/uL 2-5 years: 189-394 K/uL 6-11 years: 199-367 K/uL 12-17 years: 194-345 K/uL 18+ years: 150-393 K/uL</p> <p>MPV 0-14 days: 10.4-12.0 fL 15-30 days: 10.0-12.2 fL 31-60 days: 9.4-11.1 fL 61-179 days: 9.0-10.9 fL 180 days-2 years: 8.8-10.6 fL 2-5 years: 8.9-11.0 fL 6-11 years: 9.1-11.3 fL 12-17 years: 9.6-11.7 fL 18+ years: 8.5-12.2 fL</p>	<p>MCH 0-14 days: 31.3-35.6 pg 15-30 days: 29.9-34.1 pg 31-60 days: 27.8-32.0 pg 61-179 days: 24.4-29.5 pg 180 days-2 years: 22.7-27.2 pg 2-5 years: 23.7-28.3 pg 6-11 years: 24.9-29.2 pg 12-17 years: 25.2-30.2 pg 18+ years: 26.1-33.3 pg</p> <p>MCHC 0-14 days: 33.0-35.7 g/dL 15-30 days: 32.7-35.1 g/dL 31-60 days: 32.3-34.8 g/dL 61-179 days: 31.9-34.4 g/dL 180 days-2 years: 31.6-34.4 g/dL 2-5 years: 32.0-34.7 g/dL 6-11 years: 32.2-34.9 g/dL 12-17 years: 31.8-34.8 g/dL 18+ years: 31.9-36.5 g/dL</p> <p>RDW 0-14 days: 14.8-17.0 % 15-30 days: 14.3-16.8 % 31-60 days: 13.8-16.1 % 61-179 days: 12.4-15.3 % 180 days-2 years: 12.9-15.6 % 2-5 years: 12.5-14.9 % 6-11 years: 12.3-14.1 % 12-17 years: 12.4-14.5 % 18+ years: 10.9-14.3 %</p> <p>PLT 0-14 days: 218-419 K/uL 15-30 days: 248-586 K/uL 31-60 days: 229-562 K/uL 61-179 days: 244-529 K/uL 180 days-2 years: 206-445 K/uL 2-5 years: 202-403 K/uL 6-11 years: 206-369 K/uL 12-17 years: 175-332 K/uL 18+ years: 146-337 K/uL</p> <p>MPV 0-14 days: 10.2-11.9 fL 15-30 days: 10.1-12.1 fL 31-60 days: 9.2-10.8 fL 61-179 days: 8.9-10.6 fL 180 days-2 years: 8.7-10.5 fL 2-5 years: 9.0-10.9 fL 6-11 years: 9.2-11.4 fL 12-17 years: 9.6-11.8 fL 18+ years: 8.7-12.3 fL</p>	(semiconductor laser), and sodium lauryl sulfate (SLS)-hemoglobin method.	Soldin, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	HCT: See individual analyte PLT: See individual analyte	HCT: See individual analyte PLT: See individual analyte
Clumped Platelet, Citrated Draw	CBP	<p>0-14 days: 144-449 K/uL 15-30 days: 279-571 K/uL 31-60 days: 331-597 K/uL 61-179 days: 247-580 K/uL 180 days-2 years: 214-459 K/uL 2-5 years: 189-394 K/uL 6-11 years: 199-367 K/uL 12-17 years: 194-345 K/uL 18+ years: 150-393 K/uL</p>	<p>0-14 days: 218-419 K/uL 15-30 days: 248-586 K/uL 31-60 days: 229-562 K/uL 61-179 days: 244-529 K/uL 180 days-2 years: 206-445 K/uL 2-5 years: 202-403 K/uL 6-11 years: 206-369 K/uL 12-17 years: 175-332 K/uL 18+ years: 146-337 K/uL</p>	Electronic resistance detection and flow cytometry.	OSU Internal Normal Range Study, October 2018	5-5,000 K/uL	≥5 K/uL
Electronic Diff	CDPLT	<p>NRBC 0-3 days: 0.1-8.3 /100 WBC 4 days - 17 years: <0.0 /100 WBC 18+ years: <0.2 /100 WBC</p> <p>SEG-BANDABS 0-14 days: 1.73-6.75 K/uL 15-30 days: 1.23-4.80 K/uL 31-60 days: 1.00-4.68 K/uL 61-179 days: 1.04-7.20 K/uL 180 days-2 years: 1.27-7.18 K/uL 2-5 years: 1.60-8.20 K/uL 6-11 years: 1.64-7.87 K/uL 12-17 years: 1.82-7.47 K/uL 18+ years: 1.64-7.29 K/uL</p> <p>LYMPHABS 0-14 days: 1.75-8.00 K/uL 15-30 days: 2.42-8.20 K/uL 31-60 days: 2.29-9.14 K/uL 61-179 days: 2.14-8.99 K/uL 180 days-2 years: 1.52-8.09 K/uL 2-5 years: 1.25-8.77 K/uL 6-11 years: 1.16-8.28 K/uL 12-17 years: 1.16-3.33 K/uL 18+ years: 1.16-3.51 K/uL</p> <p>MONOABS 0-14 days: 0.57-1.72 K/uL 15-30 days: 0.42-1.21 K/uL 31-60 days: 0.28-1.21 K/uL 61-179 days: 0.24-1.17 K/uL 180 days-2 years: 0.26-1.08 K/uL 2-5 years: 0.24-0.92 K/uL 6-11 years: 0.19-0.81 K/uL 12-17 years: 0.19-0.72 K/uL 18+ years: 0.22-0.87 K/uL</p> <p>EOSABS 0-14 days: 0.09-0.64 K/uL 15-30 days: 0.06-0.75 K/uL 31-60 days: 0.04-0.63 K/uL 61-179 days: 0.02-0.74 K/uL 180 days-2 years: 0.02-0.58 K/uL 2-5 years: 0.03-0.46 K/uL 6-11 years: 0.03-0.47 K/uL 12-17 years: 0.02-0.32 K/uL 18+ years: 0.00-0.42 K/uL</p>	<p>NRBC 0-3 days: 0.1-8.3 /100 WBC 4 days - 17 years: <0.0 /100 WBC 18+ years: <0.2 /100 WBC</p> <p>SEG-BANDABS 0-14 days: 1.60-6.06 K/uL 15-30 days: 1.18-4.45 K/uL 31-60 days: 0.83-4.23 K/uL 61-179 days: 0.97-5.45 K/uL 180 days-2 years: 1.19-7.21 K/uL 2-5 years: 1.54-7.92 K/uL 6-11 years: 1.63-7.55 K/uL 12-17 years: 1.54-7.04 K/uL 18+ years: 1.57-6.19 K/uL</p> <p>LYMPHABS 0-14 days: 2.07-7.53 K/uL 15-30 days: 2.11-8.38 K/uL 31-60 days: 2.47-7.95 K/uL 61-179 days: 2.45-8.89 K/uL 180 days-2 years: 1.56-7.83 K/uL 2-5 years: 1.13-8.52 K/uL 6-11 years: 0.97-3.96 K/uL 12-17 years: 0.97-3.26 K/uL 18+ years: 0.83-3.57 K/uL</p> <p>MONOABS 0-14 days: 0.52-1.77 K/uL 15-30 days: 0.28-1.38 K/uL 31-60 days: 0.28-1.05 K/uL 61-179 days: 0.28-1.07 K/uL 180 days-2 years: 0.25-1.15 K/uL 2-5 years: 0.19-0.94 K/uL 6-11 years: 0.19-0.85 K/uL 12-17 years: 0.18-0.79 K/uL 18+ years: 0.24-0.93 K/uL</p> <p>EOSABS 0-14 days: 0.12-0.66 K/uL 15-30 days: 0.08-0.80 K/uL 31-60 days: 0.05-0.57 K/uL 61-179 days: 0.03-0.61 K/uL 180 days-2 years: 0.02-0.82 K/uL 2-5 years: 0.03-0.53 K/uL 6-11 years: 0.03-0.52 K/uL 12-17 years: 0.04-0.38 K/uL 18+ years: 0.00-0.48 K/uL</p>	Flow cytometry	OSU Internal Normal Range Study, October 2018	SEG-BANDS%: 0.0-100.0 % LYMPHS%: 0.0-100.0 % EOS%: 0.0-100.0 % BAS%: 0.0-100.0 % IGRE%: 0.0-100.0 % NRBC: 0.0-600.0 /100 WBC SEG-BANDABS: 0.04-440.00 K/uL LYMPHABS: 0.04-440.00 K/uL MONOABS: 0.04-440.00 K/uL EOSABS: 0.04-440.00 K/uL IGREABS: 0.04-440.00 K/uL	SEG-BANDS%: 0.0-100.0 % LYMPHS%: 0.0-100.0 % EOS%: 0.0-100.0 % BAS%: 0.0-100.0 % IGRE%: 0.0-100.0 % NRBC: 0.0-600.0 /100 WBC SEG-BANDABS: >440.00 K/uL LYMPHABS: >440.00 K/uL MONOABS: >440.00 K/uL EOSABS: >440.00 K/uL IGREABS: >440.00 K/uL

		<p>BASOABS 0-14 days: 0.02-0.07 K/uL 15-30 days: 0.01-0.06 K/uL 31-60 days: 0.01-0.05 K/uL 61-179 days: 0.01-0.07 K/uL 180 days-2 years: 0.01-0.06 K/uL 2-5 years: 0.01-0.06 K/uL 6-11 years: 0.01-0.05 K/uL 12-17 years: 0.01-0.05 K/uL 18+ years: 0.00-0.15 K/uL</p> <p>IGREABS 0-1 days: <0.28 K/uL 2-13 days: <0.27 K/uL 14-30 days: <0.22 K/uL 31-90 days: <0.09 K/uL 91-180 days: <0.06 K/uL 181 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.08 K/uL</p>	<p>BASOABS 0-14 days: 0.02-0.11 K/uL 15-30 days: 0.01-0.07 K/uL 31-60 days: 0.01-0.07 K/uL 61-179 days: 0.01-0.08 K/uL 180 days-2 years: 0.01-0.06 K/uL 2-5 years: 0.01-0.06 K/uL 6-11 years: 0.01-0.06 K/uL 12-17 years: 0.01-0.05 K/uL 18+ years: 0.00-0.09 K/uL</p> <p>IGREABS 0-1 days: <0.28 K/uL 2-13 days: <0.27 K/uL 14-30 days: <0.22 K/uL 31-90 days: <0.09 K/uL 91-180 days: <0.06 K/uL 181 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.07 K/uL</p>					
	CBC,EDDF EDDF							
Extended Reticulocyte Panel		<p>RETIC 0-3 days: 3.47-5.40 % 4-30 days: 1.06-2.27 % 31-60 days: 2.12-3.47 % 61-179 days: 1.55-2.70 % 180 days-2 years: 0.99-1.82 % 2-5 years: 0.82-1.45 % 6-11 years: 0.98-1.94 % 12-17 years: 0.90-1.49 % 18+ years: 0.74-2.94 %</p> <p>RETICABS 0-3 days: 0.1475-0.2164 M/uL 4-30 days: 0.0513-0.1104 M/uL 31-60 days: 0.0518-0.0779 M/uL 61-179 days: 0.0482-0.0882 M/uL 180 days-2 years: 0.0435-0.1111 M/uL 2-5 years: 0.0364-0.0680 M/uL 6-11 years: 0.0424-0.0702 M/uL 12-17 years: 0.0416-0.0651 M/uL 18+ years: 0.0324-0.1142 M/uL</p> <p>IRETF 0-3 days: 30.5-35.1 % 4-30 days: 14.5-24.6 % 31-60 days: 19.1-28.9 % 61-179 days: 13.4-23.3 % 180 days-2 years: 11.4-25.8 % 2-5 years: 8.4-21.7 % 6-11 years: 8.9-24.1 % 12-17 years: 9.0-18.7 % 18+ years: 1.1-16.2 %</p> <p>RETHE 0-179 days: 29.2-37.5 pg 180 days-2 years: 30.1-35.7 pg 2-5 years: 29.3-37.3 pg 6-11 years: 30.4-39.7 pg 12-17 years: 29.9-38.4 pg 18+ years: 28.8-39.9 pg</p>	<p>RETIC 0-3 days: 3.47-5.40 % 4-30 days: 1.062-2.37 % 31-60 days: 2.12-3.47 % 61-179 days: 1.55-2.70 % 180 days-2 years: 0.99-1.82 % 2-5 years: 0.82-1.45 % 6-11 years: 0.98-1.94 % 12-17 years: 0.90-1.49 % 18+ years: 0.68-2.64 %</p> <p>RETICABS 0-3 days: 0.1475-0.2164 M/uL 4-30 days: 0.0513-0.1104 M/uL 31-60 days: 0.0518-0.0779 M/uL 61-179 days: 0.0482-0.0882 M/uL 180 days-2 years: 0.0435-0.1111 M/uL 2-5 years: 0.0364-0.0680 M/uL 6-11 years: 0.0424-0.0702 M/uL 12-17 years: 0.0416-0.0651 M/uL 18+ years: 0.0317-0.1377 M/uL</p> <p>IRETF 0-3 days: 30.5-35.1 % 4-30 days: 14.5-24.6 % 31-60 days: 19.1-28.9 % 61-179 days: 13.4-23.3 % 180 days-2 years: 11.4-25.8 % 2-5 years: 8.4-21.7 % 6-11 years: 8.9-24.1 % 12-17 years: 9.0-18.7 % 18+ years: 0.2-16.3 %</p> <p>RETHE 0-179 days: 27.6-38.7 pg 180 days-2 years: 28.7-35.7 pg 2-5 years: 27.7-37.8 pg 6-11 years: 32.4-37.6 pg 12-17 years: 30.3-40.4 pg 18+ years: 29.9-38.7 pg</p>	Flow cytometry / calculation	OSU Internal Normal Range Study, October 2018 Siddins, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	<p>RETIC: 0.25-30.00 % RETICABS: 0.0100-2.5800 XN: 0.0100-0.7200 XN-L: 0.0100-0.4576 Manual: 0.0100-2.5800 IRETF: 0.0-100.0 %</p>	<p>RETIC: 0.25-30.00 % RETICABS: 0.0100-2.5800 Manual: 0.0100-2.5800 IRETF: 0.0-100.0 %</p> <p>*Instrument may prompt dilution but final result should not exceed 2.5800</p>	
	EXRETIC RF RET-HE	<p>See HEM3 and LAB26</p> <p>BLAST% <0.0 %</p> <p>LYMPHOMA% <0.0 %</p> <p>HAIRY% <0.0 %</p> <p>PLASMA% <0.0 %</p> <p>PROLYMPH% <0.0 %</p> <p>OTHER% <0.0 %</p> <p>NRBC 0.0-0.2 /100 WBC</p> <p>SEG-BANDABS 0-14 days: 1.76-6.75 K/uL 15-30 days: 1.23-4.80 K/uL 31-60 days: 1.00-4.68 K/uL 61-179 days: 1.06-7.20 K/uL 180 days-2 years: 1.27-7.18 K/uL 2-5 years: 1.60-8.29 K/uL 6-11 years: 1.64-7.87 K/uL 12-17 years: 1.82-7.47 K/uL 18+ years: 1.64-7.28 K/uL</p> <p>LYMPHABS 0-14 days: 1.75-8.00 K/uL 15-30 days: 2.42-8.20 K/uL 31-60 days: 2.29-9.14 K/uL 61-179 days: 2.14-8.99 K/uL 180 days-2 years: 1.52-8.09 K/uL 2-5 years: 1.25-5.79 K/uL 6-11 years: 1.16-4.28 K/uL 12-17 years: 1.16-3.33 K/uL 18+ years: 1.16-3.51 K/uL</p> <p>MONOABS 0-14 days: 0.57-1.72 K/uL 15-30 days: 0.42-1.21 K/uL 31-60 days: 0.28-1.21 K/uL 61-179 days: 0.24-1.17 K/uL</p>	<p>See HEM3 and LAB26</p> <p>BLAST% <0.0 %</p> <p>LYMPHOMA% <0.0 %</p> <p>HAIRY% <0.0 %</p> <p>PLASMA% <0.0 %</p> <p>PROLYMPH% <0.0 %</p> <p>OTHER% <0.0 %</p> <p>NRBC 0.0-0.2 /100 WBC</p> <p>SEG-BANDABS 0-14 days: 1.60-6.06 K/uL 15-30 days: 1.18-5.45 K/uL 31-60 days: 0.83-4.23 K/uL 61-179 days: 0.97-5.45 K/uL 180 days-2 years: 1.19-7.21 K/uL 2-5 years: 1.54-7.92 K/uL 6-11 years: 1.63-7.55 K/uL 12-17 years: 1.54-7.04 K/uL 18+ years: 1.57-6.19 K/uL</p> <p>LYMPHABS 0-14 days: 2.07-7.53 K/uL 15-30 days: 2.11-8.38 K/uL 31-60 days: 2.47-7.95 K/uL 61-179 days: 2.45-8.89 K/uL 180 days-2 years: 1.56-7.83 K/uL 2-5 years: 1.18-5.82 K/uL 6-11 years: 0.97-3.96 K/uL 12-17 years: 0.97-3.26 K/uL 18+ years: 0.83-3.57 K/uL</p> <p>MONOABS 0-14 days: 0.51-1.77 K/uL 15-30 days: 0.28-1.38 K/uL 31-60 days: 0.28-1.05 K/uL 61-179 days: 0.28-1.07 K/uL</p>					
								<p>See HEM3 and LAB26</p> <p>BANDS: 0.0-100.0 % SEGs: 0.0-100.0 % LYMPHs: 0.0-100.0 % MONOs: 0.0-100.0 % Eos: 0.0-100.0 % BASOs: 0.0-100.0 % METAs: 0.0-100.0 % MYELOs: 0.0-100.0 % Pp: 0.0-100.0 %</p>

<p>Fetal (Pre) CBC, DIFF, Retc</p>	<p>180 days-2 years: 0.26-1.08 K/uL 2-5 years: 0.24-0.93 K/uL 6-11 years: 0.19-0.81 K/uL 12-17 years: 0.19-0.72 K/uL 18+ years: 0.22-0.87 K/uL</p> <p>EOSABS 0-14 days: 0.09-0.64 K/uL 15-30 days: 0.06-0.75 K/uL 31-60 days: 0.04-0.63 K/uL 61-179 days: 0.02-0.74 K/uL 180 days-2 years: 0.02-0.58 K/uL 2-5 years: 0.03-0.46 K/uL 6-11 years: 0.03-0.47 K/uL 12-17 years: 0.02-0.32 K/uL 18+ years: 0.00-0.42 K/uL</p> <p>BASOABS 0-14 days: 0.02-0.07 K/uL 15-30 days: 0.01-0.06 K/uL 31-60 days: 0.01-0.05 K/uL 61-179 days: 0.01-0.07 K/uL 180 days-2 years: 0.01-0.06 K/uL 2-5 years: 0.01-0.06 K/uL 6-11 years: 0.01-0.05 K/uL 12-17 years: 0.01-0.05 K/uL 18+ years: 0.00-0.15 K/uL</p> <p>METAABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.08 K/uL</p> <p>MYELOABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.08 K/uL</p> <p>PROMYELOABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.08 K/uL</p> <p>BLASTABS <0.00 K/uL</p> <p>LYMPHOMAABS <0.00 K/uL</p> <p>HARYABS <0.00 K/uL</p> <p>PLASMAABS <0.00 K/uL</p>	<p>180 days-2 years: 0.25-1.15 K/uL 2-5 years: 0.19-0.94 K/uL 6-11 years: 0.19-0.85 K/uL 12-17 years: 0.18-0.78 K/uL 18+ years: 0.24-0.93 K/uL</p> <p>EOSABS 0-14 days: 0.12-0.66 K/uL 15-30 days: 0.08-0.80 K/uL 31-60 days: 0.05-0.57 K/uL 61-179 days: 0.03-0.61 K/uL 180 days-2 years: 0.02-0.82 K/uL 2-5 years: 0.03-0.53 K/uL 6-11 years: 0.03-0.52 K/uL 12-17 years: 0.04-0.38 K/uL 18+ years: 0.00-0.48 K/uL</p> <p>BASOABS 0-14 days: 0.02-0.11 K/uL 15-30 days: 0.01-0.07 K/uL 31-60 days: 0.01-0.07 K/uL 61-179 days: 0.01-0.06 K/uL 180 days-2 years: 0.01-0.06 K/uL 2-5 years: 0.01-0.06 K/uL 6-11 years: 0.01-0.06 K/uL 12-17 years: 0.01-0.05 K/uL 18+ years: 0.00-0.09 K/uL</p> <p>METAABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.07 K/uL</p> <p>MYELOABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.07 K/uL</p> <p>PROMYELOABS 0-1 days: <0.28 K/uL 2-12 days: <0.27 K/uL 13-29 days: <0.22 K/uL 30-89 days: <0.09 K/uL 90-179 days: <0.06 K/uL 180 days-2 years: <0.14 K/uL 2-5 years: <0.06 K/uL 6-11 years: <0.04 K/uL 12-17 years: <0.04 K/uL 18+ years: <0.07 K/uL</p> <p>BLASTABS <0.00 K/uL</p> <p>LYMPHOMAABS <0.00 K/uL</p> <p>HARYABS <0.00 K/uL</p> <p>PLASMAABS <0.00 K/uL</p>	<p>The Sysmex XN performs hematology analysis according to the hydrodynamic focusing (DF) detection, flow cytometry method (semiconducting laser), and sodium lauryl sulfate (SLS) hemoglobin method.</p>	<p>OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i>. 7th ed., AACCPress, 2011.</p>	<p>LYMPH%: 0.0-100.0 % MONO%: 0.0-100.0 % EOS%: 0.0-100.0 % BASO%: 0.0-100.0 % META%: 0.0-100.0 % MYELO%: 0.0-100.0 % PROMYELO%: 0.0-100.0 % BLAST%: 0.0-100.0 % LYMPHOMA%: 0.0-100.0 % HARY%: 0.0-100.0 % PLASMA%: 0.0-100.0 % SEG-BANDABS: 0.00-440.00 K/uL LYMPHABS: 0.00-440.00 K/uL MONOABS: 0.00-440.00 K/uL EOSABS: 0.00-440.00 K/uL BASOABS: 0.00-440.00 K/uL METAABS: 0.00-440.00 K/uL MYELOABS: 0.00-440.00 K/uL PROMYELOABS: 0.00-440.00 K/uL BLASTABS: 0.00-440.00 K/uL LYMPHOMAABS: 0.00-440.00 K/uL HARYABS: 0.00-440.00 K/uL PLASMAABS: 0.00-440.00 K/uL POLYMPHABS: 0.00-440.00 K/uL OTHERABS: 0.00-440.00 K/uL</p> <p>BLAST%: 0.0-100.0 % LYMPHOMA%: 0.0-100.0 % HARY%: 0.0-100.0 % PLASMA%: 0.0-100.0 % SEG-BANDABS: 0.00-440.00 L/uL EOSABS: 0.00-440.00 L/uL BASOABS: 0.00-440.00 L/uL METAABS: 0.00-440.00 L/uL MYELOABS: 0.00-440.00 L/uL PROMYELOABS: 0.00-440.00 L/uL OTHERABS: 0.00-440.00 L/uL SCHISTO: 1+, 2+, 3+ POLY: 1+, 2+, 3+ SKNLE: 1+, 2+, 3+ SPHEROCYTES: 1+, 2+, 3+ ACANTHO: Present ECHIN: Present OVALO: Present TARGET: Present TEAR: Present HBODDS: Present ROULLEUX: Present AGGLUT: Present PLTCLUMP: Present GANT: Present SATLITELISM: Present</p>
<p>Hematocrit</p>	<p>0-14 days: 39.6-57.2 % 15-30 days: 32.0-44.5 % 31-60 days: 27.7-35.1 % 61-179 days: 29.5-37.1 % 180 days-2 years: 30.9-37.9 % 2-5 years: 31.2-37.8 % 6-11 years: 32.4-39.5 % 12-17 years: 33.4-40.4 % 18+ years: 34.9-44.3 %</p>	<p>0-14 days: 39.8-53.6 % 15-30 days: 30.5-45.0 % 31-60 days: 26.8-37.5 % 61-179 days: 28.6-37.2 % 180 days-2 years: 30.8-37.8 % 2-5 years: 31.0-37.7 % 6-11 years: 32.2-39.8 % 12-17 years: 33.9-43.5 % 18+ years: 39.6-48.8 %</p>	<p>Cumulative pulse height detection.</p>	<p>OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i>. 7th ed., AACCPress, 2011.</p>	<p>0.1-75.0 % 0.1 %-dilate to endpoint</p>
<p>Hemoglobin</p>	<p>0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180 days-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL</p>	<p>0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180 days-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL</p>	<p>Photometrically measured.</p>	<p>OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i>. 7th ed., AACCPress, 2011.</p>	<p>0.2-26.0 g/dL 0.2 g/dL-dilate to endpoint</p>
<p>Hemoglobin & Hematocrit</p>	<p>HGB 0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180 days-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL</p> <p>HCT 0-14 days: 39.6-57.2 % 15-30 days: 32.0-44.5 % 31-60 days: 27.7-35.1 % 61-179 days: 29.5-37.1 % 180 days-2 years: 30.9-37.9 % 2-5 years: 31.2-37.8 % 6-11 years: 32.4-39.5 % 12-17 years: 33.4-40.4 % 18+ years: 34.9-44.3 %</p>	<p>HGB 0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180 days-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL</p> <p>HCT 0-14 days: 39.8-53.6 % 15-30 days: 30.5-45.0 % 31-60 days: 26.8-37.5 % 61-179 days: 28.6-37.2 % 180 days-2 years: 30.8-37.8 % 2-5 years: 31.0-37.7 % 6-11 years: 32.2-39.8 % 12-17 years: 33.9-43.5 % 18+ years: 39.6-48.8 %</p>	<p>HGB: Photometrically measured. HCT: Cumulative pulse height detection.</p>	<p>OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i>. 7th ed., AACCPress, 2011.</p>	<p>See individual analytes See individual analytes</p>

Immature Platelet Fraction	IPF 0-179 days: 1.3-6.8 % 180 days-2 years: 1.4-4.5 % 2-5 years: 1.0-3.6 % 6-11 years: 1.0-4.7 % 12-17 years: 1.4-6.4 % 18+ years: 0.0-8.6 % PLICIPP 0-14 days: 144-449 K/uL 15-30 days: 279-571 K/uL 31-60 days: 331-597 K/uL 61-179 days: 247-580 K/uL 180 days-2 years: 214-459 K/uL 2-5 years: 189-394 K/uL 6-11 years: 199-367 K/uL 12-17 years: 194-345 K/uL 18+ years: 150-393 K/uL	IPF 0-179 days: 2.0-6.8 % 180 days-2 years: 1.4-3.8 % 2-5 years: 1.1-3.6 % 6-11 years: 1.0-4.9 % 12-17 years: 1.6-6.1 % 18+ years: 0.0-9.0 % PLICIPP 0-14 days: 218-419 K/uL 15-30 days: 248-586 K/uL 31-60 days: 229-562 K/uL 61-179 days: 244-529 K/uL 180 days-2 years: 206-445 K/uL 2-5 years: 202-403 K/uL 6-11 years: 206-369 K/uL 12-17 years: 175-332 K/uL 18+ years: 146-337 K/uL	Calculation	OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	IPF: 0.0-100.0 % PLICIPP: 5-5,000 K/uL	IPF: 0.0-100.0 % PLICIPP: 5 K/uL-dilute to endpoint
Malaria Prep	Parasite Screen ID Blood MPH		Giemsa stain			No parasitic organisms seen, including plasmodium organisms Positive for Plasmodium species
Platelet Count	PLT 0-14 days: 144-449 K/uL 15-30 days: 279-571 K/uL 31-60 days: 331-597 K/uL 61-179 days: 247-580 K/uL 180 days-2 years: 214-459 K/uL 2-5 years: 189-394 K/uL 6-11 years: 199-367 K/uL 12-17 years: 194-345 K/uL 18+ years: 150-393 K/uL MPV 0-14 days: 10.4-12.0 fL 15-30 days: 10.0-12.2 fL 31-60 days: 9.4-11.1 fL 61-179 days: 9.0-10.9 fL 180 days-2 years: 8.8-10.6 fL 2-5 years: 8.9-11.0 fL 6-11 years: 9.3-11.3 fL 12-17 years: 9.6-11.7 fL 18+ years: 8.5-12.2 fL	PLT 0-14 days: 218-419 K/uL 15-30 days: 248-586 K/uL 31-60 days: 229-562 K/uL 61-179 days: 244-529 K/uL 180 days-2 years: 206-445 K/uL 2-5 years: 202-403 K/uL 6-11 years: 206-369 K/uL 12-17 years: 175-332 K/uL 18+ years: 146-337 K/uL MPV 0-14 days: 10.2-11.9 fL 15-30 days: 10.1-12.1 fL 31-60 days: 9.2-10.8 fL 61-179 days: 8.9-10.6 fL 180 days-2 years: 8.7-10.5 fL 2-5 years: 9.0-10.9 fL 6-11 years: 9.2-11.4 fL 12-17 years: 9.6-11.8 fL 18+ years: 8.7-12.3 fL	Electronic resistance detection and flow cytometry.	OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	5-5,000 K/uL	5 K/uL-dilute to endpoint
Reticulocytes	PLAT RETIC 0-3 days: 3.47-5.40 % 4-30 days: 1.06-2.37 % 31-60 days: 2.12-3.47 % 61-179 days: 1.55-2.70 % 180 days-2 years: 0.99-1.82 % 2-5 years: 0.82-1.45 % 6-11 years: 0.98-1.94 % 12-17 years: 0.90-1.49 % 18+ years: 0.74-2.54 % RETICABS 0-3 days: 0.1475-0.2164 MuL 4-30 days: 0.0513-0.1104 MuL 31-60 days: 0.0518-0.0779 MuL 61-179 days: 0.0482-0.0882 MuL 180 days-2 years: 0.0435-0.1111 MuL 2-5 years: 0.0364-0.0680 MuL 6-11 years: 0.0424-0.0702 MuL 12-17 years: 0.0416-0.0651 MuL 18+ years: 0.0324-0.1142 MuL	RETIC 0-3 days: 3.47-5.40 % 4-30 days: 1.06-2.37 % 31-60 days: 2.12-3.47 % 61-179 days: 1.55-2.70 % 180 days-2 years: 0.99-1.82 % 2-5 years: 0.82-1.45 % 6-11 years: 0.98-1.94 % 12-17 years: 0.90-1.49 % 18+ years: 0.68-2.64 % RETICABS 0-3 days: 0.1475-0.2164 MuL 4-30 days: 0.0513-0.1104 MuL 31-60 days: 0.0518-0.0779 MuL 61-179 days: 0.0482-0.0882 MuL 180 days-2 years: 0.0435-0.1111 MuL 2-5 years: 0.0364-0.0680 MuL 6-11 years: 0.0424-0.0702 MuL 12-17 years: 0.0416-0.0651 MuL 18+ years: 0.0317-0.1377 MuL	Flow cytometry	OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	RETIC: 0.25-30.00% RETICABS: XX: 0.0100-0.7200 XX-L: 0.0100-0.4576 Manual: 0.0100-2.5800	RETIC: 0.25-30.00 % RETICABS: Manual: 0.0100-2.5800*
Schistocytes	PSCHISTO		Manual			
Sedimentation Rate, Automated	ESR	0-49 years: <20 mm/hr 50-84 years: <30 mm/hr 85+ years: <42 mm/hr	Westergren	McPherson, R.A., & Pincus, M.R. (2017). <i>Henry's Clinical Diagnosis and Management by Laboratory Methods</i> (23rd ed.). St. Louis, MO: Elsevier Inc., pp. 533.	1-140 mm/hr	1-140 mm/hr
White Blood Count	WBC	0-14 days: 8.16-14.56 K/uL 15-30 days: 8.36-14.42 K/uL 31-60 days: 7.05-14.68 K/uL 61-179 days: 6.00-13.25 K/uL 180 days-2 years: 6.48-13.02 K/uL 2-5 years: 4.86-13.18 K/uL 6-11 years: 4.27-11.40 K/uL 12-17 years: 4.19-9.43 K/uL 18+ years: 3.99-11.19 K/uL	Flow cytometry	OSU Internal Normal Range Study, October 2018 Soldin, Steven J. <i>Pediatric Reference Intervals</i> . 7th ed., AACCPress, 2011.	0.30-440.00 K/uL	0.30 K/uL-dilute to endpoint
Body Fluid Cell Count	FCNT		Hemocytometer counts and Iris instrument.		TNC: 2-3,500 /uL RBC: 3-50,000 /uL	TNC: 23 /uL RBC: 23 /uL
Body Fluid Cell Count with Diff	FCNT FPDIFF Body fluid differential	SFBLAST 0% SFMALG 0% SFLYPHOM 0%	Hemocytometer counts, Iris instrument and manual differential.		TNC: 3-2,500 /uL RBC: 3-50,000 /uL FSEG: 0-100 % FSEGL: 0-100 % FLYM: 0-100 % FMOMAC: 0-100 % FMESO: 0-100 % FEOS: 0-100 % FRASO: 0-100 % FPLAS: 0-100 % FPLASL: 0-100 % FMALIG: 0-100 % FRLAS: 0-100 % FLPHOM: 0-100 % FSYNLC: 0-100 % FNBRIC: 0-100 % FOTHR: 0-100 %	TNC: 23 /uL RBC: 23 /uL FSEG: 0-100 % FSEGL: 0-100 % FLYM: 0-100 % FMOMAC: 0-100 % FMESO: 0-100 % FEOS: 0-100 % FRASO: 0-100 % FPLAS: 0-100 % FPLASL: 0-100 % FMALIG: 0-100 % FRLAS: 0-100 % FLPHOM: 0-100 % FSYNLC: 0-100 % FNBRIC: 0-100 % FOTHR: 0-100 %

	URIN			<p>SQUAMOUS: 0-2/hpf, 3-5/hpf = 1+</p> <p>BACTERIA: Absent</p> <p>RENAL TUBULAR CELLS: 0-2/hpf</p> <p>WBC CASTS: Absent</p> <p>WBC/RENAL TUBULAR CASTS: Absent</p> <p>RBC CASTS: Absent</p> <p>YEAST: Absent</p> <p>TRICHOMONAS: Absent</p> <p>LEUCINE: Absent</p> <p>TYROSINE: Absent</p> <p>CYSTINE: Absent</p> <p>TRIPLE PHOSPHATE: Absent</p>	<p>Dry pad urine chemistry or optic refractive index method</p> <p>URINE MICROSCOPIC Automatic particle counter and/or manual microscopic</p>			<p>substance:</p> <p>UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL, 2.0 E.U./dL, 4.0 E.U./dL, ≥8.0 E.U./dL. Unable to analyze due to interfering substance.</p> <p>NITRITE: Negative. Positive. Unable to analyze due to interfering substance.</p> <p>LEUKOCYTE ESTERASE: Negative, Trace, Small, Moderate, Large. Unable to analyze due to interfering substance.</p> <p>RBC: 0-2, 3-5, 6-10, 11-25, ≥25 /HPF</p> <p>WBC: 0-5, 6-10, 11-20, ≥20 /HPF</p> <p>SQUAMOUS: 0-2/hpf, 3-5/hpf = 1+, 6-10/hpf = 2+, 11-20/hpf = 3+, ≥20/hpf = 4+</p> <p>BACTERIA: Absent, Trace, Present</p> <p>RENAL TUBULAR CELLS: 0-2/hpf, 3-5/hpf = 1+, 6-10/hpf = 2+, 11-20/hpf = 3+, ≥20/hpf = 4+</p> <p>AMORPHOUS: <25% = Slight, 26-50% = Moderate, >51% = Heavy</p> <p>HYALINE CASTS: 0-2, 3-5, 6-10, 11-</p>
Urinalysis Reflex to Culture	URIN1			<p>COLOR: Yellow</p> <p>APPEARANCE: Clear</p> <p>GLUCOSE: Negative</p> <p>KETONE: Negative</p> <p>SPECIFIC GRAVITY: 1.001-1.035</p> <p>BLOOD: Negative</p> <p>PH: 5.0, 5.5, 6.0, 6.5, 7.0</p> <p>PROTEIN: Negative</p> <p>UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL</p> <p>NITRITE: Negative</p> <p>LEUKOCYTE ESTERASE: Negative</p> <p>RBC: 0-2 /HPF</p> <p>WBC: 0-5 /HPF</p> <p>SQUAMOUS: 0-2/hpf, 3-5/hpf = 1+</p> <p>BACTERIA: Absent</p> <p>RENAL TUBULAR CELLS: 0-2/hpf</p> <p>WBC CASTS: Absent</p> <p>WBC/RENAL TUBULAR CASTS: Absent</p> <p>RBC CASTS: Absent</p> <p>YEAST: Absent</p> <p>TRICHOMONAS: Absent</p> <p>LEUCINE: Absent</p> <p>TYROSINE: Absent</p> <p>CYSTINE: Absent</p> <p>TRIPLE PHOSPHATE: Absent</p>	<p>DIPSTICK Dry pad urine chemistry</p> <p>SG Dry pad urine chemistry or optic refractive index method</p> <p>URINE MICROSCOPIC Automatic particle counter and/or manual microscopic</p>	Varies	See LAB347	
Urine Dipstick with Reflex to Microscopy	UASR			<p>COLOR: Yellow</p> <p>APPEARANCE: Clear</p> <p>GLUCOSE: Negative</p> <p>KETONE: Negative</p> <p>SPECIFIC GRAVITY: 1.001-1.035</p> <p>BLOOD: Negative</p> <p>PH: 5.0, 5.5, 6.0, 6.5, 7.0</p> <p>PROTEIN: Negative</p> <p>UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL</p> <p>NITRITE: Negative</p> <p>LEUKOCYTE ESTERASE: Negative</p>	<p>DIPSTICK Dry pad urine chemistry</p> <p>SG Dry pad urine chemistry or optic refractive index method</p>	Varies	See URN1	
Urine Dipstick; Reflex Microscopy; Reflex Culture	UASR1			<p>COLOR: Yellow</p> <p>APPEARANCE: Clear</p> <p>GLUCOSE: Negative</p> <p>KETONE: Negative</p> <p>SPECIFIC GRAVITY: 1.001-1.035</p> <p>BLOOD: Negative</p> <p>PH: 5.0, 5.5, 6.0, 6.5, 7.0</p> <p>PROTEIN: Negative</p> <p>UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL</p> <p>NITRITE: Negative</p> <p>LEUKOCYTE ESTERASE: Negative</p>	<p>DIPSTICK Dry pad urine chemistry</p> <p>SG Dry pad urine chemistry or optic refractive index method</p>	National guidelines and recommendations	See URN1	

Urine Screen	UAS Urino dipstick			<p>COLOR: Yellow APPEARANCE: Clear GLUCOSE: Negative KETONE: Negative SPECIFIC GRAVITY: 1.001-1.035 BLOOD: Negative PH: 5.0, 5.5, 6.0, 6.5, 7.0 PROTEIN: Negative UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL NITRITE: Negative LEUKOCYTE ESTERASE: Negative</p>	<p>DIPSTICK Dry pad urine chemistry</p> <p>SG Dry pad urine chemistry or optic refractive index method</p>	Varies	<p>COLOR: Yellow, Orange, Red, See Comment APPEAR: Clear, Cloudy, Turbid, Unable to analyze due to interfering substance. GLUCOSE: Negative, 100 mg/dL, 250 mg/dL, 500 mg/dL, ≥1000 mg/dL, Unable to analyze due to interfering substance. KETONE: Negative, Trace, 15 mg/dL - Small, 40 mg/dL - Moderate, ≥80 mg/dL - Large, Unable to analyze due to interfering substance. SPECIFIC GRAVITY: See individual analyte BLOOD: Negative, Trace, Small, Moderate, Large, Unable to analyze due to interfering substance. PH: See individual analyte PROTEIN: Negative, Trace, 30 mg/dL, 100 mg/dL, 300 mg/dL, ≥300 mg/dL, Unable to analyze due to interfering substance. UROBILINOGEN: 0.2 E.U./dL, 1.0 E.U./dL, 2.0 E.U./dL, 4.0 E.U./dL, ≥8.0 E.U./dL, Unable to analyze due to interfering substance. NITRITE: Negative, Positive, Unable to analyze due to interfering substance. LEUKOCYTE ESTERASE: Negative, Trace, Small, Moderate, Large, Unable to analyze due to interfering substance.</p>																									
Anti Phospholipid Antibody	PLAB Anti-cardiolipin Antibodies IgG, IgM Anti-phospholipid Antibodies IgG, IgM Cardiolipin Antibodies IgG, IgM Phospholipid Antibodies IgG, IgM	PLABG 0.0-20.0 CU PLABM 0.0-20.0 CU			Chemiluminescent two-step immunoassay.	Inova Quanta-Flash Package Insert	PLABG: 10.0-500.0 CU PLABM: 10.0-200.0 CU	PLABG: 10.0-10,000.0 CU PLABM: 10.0-4,000.0 CU																								
Anti Xa DOAC (Apixaban)	AXAPIX Eliquis			<p>Therapeutic reference ranges have not been established. At steady state - median (5th-95th percentile) peak and trough levels have been observed in clinical trials:</p> <table border="1"> <thead> <tr> <th>Dose</th> <th>Indication</th> <th>Peak Conc., ng/mL (a)</th> <th>Trough Conc., ng/mL (b)</th> </tr> </thead> <tbody> <tr> <td>2.5 mg twice daily</td> <td>VTE prevention after total hip replacement</td> <td>123 (69-221)</td> <td>79 (34-162)</td> </tr> <tr> <td>5 mg twice daily</td> <td>Stroke and systemic embolism prevention in patients with AF</td> <td>171 (91-321)</td> <td>103 (41-230)</td> </tr> <tr> <td>2.5 mg twice daily</td> <td>Prevention and treatment of DVT and PE</td> <td>67 (30-153)</td> <td>32 (11-90)</td> </tr> <tr> <td>5 mg twice daily</td> <td></td> <td>132 (59-302)</td> <td>63 (22-77)</td> </tr> <tr> <td>10 mg twice daily</td> <td></td> <td>251 (111-572)</td> <td>120 (41-335)</td> </tr> </tbody> </table> <p>(a) Defined as samples collected 2-4 hours after dosing (b) Defined as samples collected 10-12 hours after dosing</p> <p>AF-atrial fibrillation, CR-CL-creatinine clearance, DVT-deep vein thrombosis, VTE-venous thromboembolism</p>	Dose	Indication	Peak Conc., ng/mL (a)	Trough Conc., ng/mL (b)	2.5 mg twice daily	VTE prevention after total hip replacement	123 (69-221)	79 (34-162)	5 mg twice daily	Stroke and systemic embolism prevention in patients with AF	171 (91-321)	103 (41-230)	2.5 mg twice daily	Prevention and treatment of DVT and PE	67 (30-153)	32 (11-90)	5 mg twice daily		132 (59-302)	63 (22-77)	10 mg twice daily		251 (111-572)	120 (41-335)	Chromogenic measurement at 405nm.	<p>1. Package insert: Rivaroxaban: Diagnostics Stage. Revised December 2014</p> <p>2. Mueck W, Stampfuss J, Kubiza D, Becka M: Clinical pharmacokinetic and pharmacodynamic profile of rivaroxaban. Clinical Pharmacokinetics 2014; 53(1):1-16 doi: 10.1007/s40262-013-0100-7</p> <p>3. Bayer Pharma AG, Xarelto (rivaroxaban) Summary of Product Characteristics, 2013. Available at: www.ema.europa.eu/docs/en_GB/document_library/yEPAR_... Product_Information/human/000944/WC5000571/08.pdf</p> <p>4. EINSTEIN Investigators, Basenachs R, Beckowitz SD, et al: Oral rivaroxaban for symptomatic venous thromboembolism. N Engl J Med 2010; 363:2499-510</p> <p>5. EINSTEIN-PE Investigators, Buller HR, Prins MH, et al: Oral rivaroxaban for the treatment of symptomatic pulmonary embolism. N Engl J Med 2012; 366:1287-1297</p> <p>6. Patel MR, Mahaffey KW, Garg J, et al: Rivaroxaban versus warfarin in nonvalvular atrial fibrillation. N Engl J Med 2011; 365:883891</p> <p>7. Sison DM, Curran JT, Connolly SJ, et al: 1. Package insert Rivaroxaban: Diagnostics Stage. Revised December 2014</p> <p>2. Mueck W, Stampfuss J, Kubiza D, Becka M: Clinical pharmacokinetic and pharmacodynamic profile of rivaroxaban. Clinical Pharmacokinetics 2014; 53(1):1-16 doi: 10.1007/s40262-013-0100-7</p> <p>3. Bayer Pharma AG, Xarelto (rivaroxaban) Summary of Product Characteristics, 2013. Available at: www.ema.europa.eu/docs/en_GB/document_library/yEPAR_... Product_Information/human/000944/WC5000571/08.pdf</p> <p>4. EINSTEIN Investigators, Basenachs R, Beckowitz SD, et al: Oral rivaroxaban for symptomatic venous thromboembolism. N Engl J Med 2010; 363:2499-510</p> <p>5. EINSTEIN-PE Investigators, Buller HR, Prins MH, et al: Oral rivaroxaban for the treatment of symptomatic pulmonary embolism. N Engl J Med 2012; 366:1287-1297</p> <p>6. Patel MR, Mahaffey KW, Garg J, et al: Rivaroxaban versus warfarin in nonvalvular atrial fibrillation. N Engl J Med 2011; 365:883891</p> <p>7. Sison DM, Curran JT, Connolly SJ, et al:</p>	23-500 ng/mL	23-500 ng/mL
Dose	Indication	Peak Conc., ng/mL (a)	Trough Conc., ng/mL (b)																													
2.5 mg twice daily	VTE prevention after total hip replacement	123 (69-221)	79 (34-162)																													
5 mg twice daily	Stroke and systemic embolism prevention in patients with AF	171 (91-321)	103 (41-230)																													
2.5 mg twice daily	Prevention and treatment of DVT and PE	67 (30-153)	32 (11-90)																													
5 mg twice daily		132 (59-302)	63 (22-77)																													
10 mg twice daily		251 (111-572)	120 (41-335)																													
Anti Xa DOAC (Rivaroxaban)	AXRIVA Xarelto			<p>Therapeutic reference ranges have not been established. At steady state - median (5th-95th percentile) peak and trough levels have been observed in clinical trials:</p> <table border="1"> <thead> <tr> <th>Dose</th> <th>Indication</th> <th>Peak Conc., ng/mL (a)</th> <th>Trough Conc., ng/mL (b)</th> </tr> </thead> <tbody> <tr> <td>2.5 mg</td> <td>Acute coronary twice daily syndrome</td> <td>46 (28-70)</td> <td>17 (6-37)</td> </tr> <tr> <td>10 mg</td> <td>VTE prevention once daily after total hip replacement</td> <td>125 (91-196)</td> <td>9 (1-38)</td> </tr> <tr> <td>15 mg</td> <td>Stroke prevention once daily in patients with AF (CR-CL 30-49 mL/min)</td> <td>229 (178-313)</td> <td>57 (18-136)</td> </tr> <tr> <td>20 mg</td> <td>DVT treatment once daily</td> <td>270 (189-419)</td> <td>26 (6-87)</td> </tr> <tr> <td>20 mg</td> <td>Stroke prevention once daily in patients with AF (CR-CL ≥50 mL/min)</td> <td>249 (184-343)</td> <td>44 (12-137)</td> </tr> </tbody> </table> <p>(a) Defined as samples collected 2-4 hours after dosing (b) Defined as samples collected 20-28 hours after dosing</p> <p>AF-atrial fibrillation, CR-CL-creatinine clearance, DVT-deep vein thrombosis, VTE-venous thromboembolism</p>	Dose	Indication	Peak Conc., ng/mL (a)	Trough Conc., ng/mL (b)	2.5 mg	Acute coronary twice daily syndrome	46 (28-70)	17 (6-37)	10 mg	VTE prevention once daily after total hip replacement	125 (91-196)	9 (1-38)	15 mg	Stroke prevention once daily in patients with AF (CR-CL 30-49 mL/min)	229 (178-313)	57 (18-136)	20 mg	DVT treatment once daily	270 (189-419)	26 (6-87)	20 mg	Stroke prevention once daily in patients with AF (CR-CL ≥50 mL/min)	249 (184-343)	44 (12-137)	Chromogenic measurement at 405nm.	<p>1. Package insert Rivaroxaban: Diagnostics Stage. Revised December 2014</p> <p>2. Mueck W, Stampfuss J, Kubiza D, Becka M: Clinical pharmacokinetic and pharmacodynamic profile of rivaroxaban. Clinical Pharmacokinetics 2014; 53(1):1-16 doi: 10.1007/s40262-013-0100-7</p> <p>3. Bayer Pharma AG, Xarelto (rivaroxaban) Summary of Product Characteristics, 2013. Available at: www.ema.europa.eu/docs/en_GB/document_library/yEPAR_... Product_Information/human/000944/WC5000571/08.pdf</p> <p>4. EINSTEIN Investigators, Basenachs R, Beckowitz SD, et al: Oral rivaroxaban for symptomatic venous thromboembolism. N Engl J Med 2010; 363:2499-510</p> <p>5. EINSTEIN-PE Investigators, Buller HR, Prins MH, et al: Oral rivaroxaban for the treatment of symptomatic pulmonary embolism. N Engl J Med 2012; 366:1287-1297</p> <p>6. Patel MR, Mahaffey KW, Garg J, et al: Rivaroxaban versus warfarin in nonvalvular atrial fibrillation. N Engl J Med 2011; 365:883891</p> <p>7. Sison DM, Curran JT, Connolly SJ, et al:</p>	25-500 ng/mL	25-500 ng/mL
Dose	Indication	Peak Conc., ng/mL (a)	Trough Conc., ng/mL (b)																													
2.5 mg	Acute coronary twice daily syndrome	46 (28-70)	17 (6-37)																													
10 mg	VTE prevention once daily after total hip replacement	125 (91-196)	9 (1-38)																													
15 mg	Stroke prevention once daily in patients with AF (CR-CL 30-49 mL/min)	229 (178-313)	57 (18-136)																													
20 mg	DVT treatment once daily	270 (189-419)	26 (6-87)																													
20 mg	Stroke prevention once daily in patients with AF (CR-CL ≥50 mL/min)	249 (184-343)	44 (12-137)																													
Anti Xa Heparin (Unfractionated)	AXHEP LPH Heparin assay Unfractionated heparin	0.30-0.70 IU/mL			Chromogenic measurement at 405nm.	Chest, vol. 119, issue 1, January 2001, pp. 64S-75S.	0.10-0.80 IU/mL	0.10-1.60 IU/mL																								
Anti Xa LMWH (Enoxaparin) Random	AXLMWR Low molecular weight heparin LMWH Random	0.60-1.00 Anti-Xa IU/mL			Chromogenic measurement at 405nm.	Chest, vol. 119, issue 1, January 2001, pp. 64S-75S.	0.10-1.60 Anti-Xa IU/mL	0.10-1.60 Anti-Xa IU/mL																								
Anti Xa LMWH (Enoxaparin), *Exact Time Required* 4 Hr Post	AXLMWRP Low molecular weight heparin LMWH Peak	0.60-1.00 Anti-Xa IU/mL			Chromogenic measurement at 405nm.	Chest, vol. 119, issue 1, January 2001, pp. 64S-75S.	0.10-1.60 Anti-Xa IU/mL	0.10-1.60 Anti-Xa IU/mL																								
Antithrombin III	AT Antithrombin Activity Antithrombin Functional ATH	AT 0-4 days: 39-87 % 5-29 days: 41-93 % 30-89 days: 48-108 % 90-179 days: 73-121 % 180-365 days: 84-124 %			Chromogenic measurement at 405nm.	OSUMMC Normal Range Study 02/2004, Blood, Vol 80, 1998-2005, Andrews, 1992; Amer. Jour. Ped. Hematol. Oncol., Vol 12, 95-104, Andrews, 1990		9-200 %																								
Beta-2 Glycoprotein 1 Ab, IgG & IgM	B2GP1 B2GP1 IgG B2GP1 IgM B2GP1 IgM	B2GP1G 0.0-20.0 CU B2GP1M 0.0-20.0 CU			Chemiluminescent two-step immunoassay.	Inova Quanta-Flash Package Insert Verified in House	B2GP1G: 10.0-500.0 CU B2GP1M: 10.0-200.0 CU	B2GP1G: 10.0-10,000.0 CU B2GP1M: 10.0-4,000.0 CU																								

D-Dimer, Quantitative			-0.50 mcg/ml FEU		A suspension of latex microparticles, coated by covalent bonding with monoclonal antibodies specific for D-dimer, is mixed with the test plasma whose D-dimer level is to be assayed. Agglutination of the microparticles results via an antigen-antibody reaction, which induces an increase in turbidity of the reaction medium. This increase in turbidity is reflected by an increase in absorbance, the latter being measured photometrically at 540 nm. The increase in absorbance is a function of the D-dimer level present in the test sample.	OSUWMC Normal Range Study (08/2007)	0.27-4.00 mcg/ml FEU	0.27-20.00 mcg/ml FEU
	HSDDI High sensitivity D-Dimer		PT 11.9-14.2 sec INR 0.9-1.1 PTT 24.0-34.3 sec TT 13.0-20.0 sec FIB 220-410 mg/dL DDIMER <0.50 mcg/ml FEU		PT, PTT, Mixing Studies, FIB, TT: Mechanical Clot Detection D-Dimer: Optical measurement at 540 nm	See individual analytes	See individual analytes	See individual analytes
Factor II Activity	DC.DICMIXES Disseminated Intravascular Coagulation Workup Consumptive Coagulopathy Workup		0-4 days: 26-70 % Activity 5-29 days: 33-93 % Activity 30-89 days: 34-102 % Activity 90-179 days: 45-105 % Activity 180-365 days: 66-116 % Activity 1-5 years: 71-116 % Activity 6-10 years: 67-107 % Activity 11-16 years: 61-104 % Activity 17+ years: 60-150 % Activity		1 stage clotting assay.	Clinical Guide to Laboratory Tests, Tietz, 1995; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	3-500 % Activity	3-500 % Activity
	FA2 Factor 2 Activity Prothrombin Activity							
Factor II Inhibitor			0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
	FAZIN Factor 2 Inhibitor Factor II Bethesda Titer Bethesda Titer							
Factor IX Activity			0-4 days: 15-91 % Activity 5-29 days: 15-91 % Activity 30-89 days: 21-81 % Activity 90-179 days: 21-113 % Activity 180-365 days: 36-136 % Activity 1-5 years: 47-104 % Activity 6-10 years: 63-89 % Activity 11-16 years: 59-122 % Activity 17+ years: 77-147 % Activity	<1% severe hemophilia 1-~5% moderate hemophilia 5-~40% mild hemophilia	1 stage clotting assay.	OSUWMC Normal Range Study 02/2004; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	1-500 % Activity	1-500 % Activity
	FA9 Factor 9 Activity Hemophilia B Christmas Factor							
Factor IX Inhibitor			0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
	FA9IN Factor 9 Inhibitor Factor IX Bethesda Titer Bethesda Titer							
Factor V Activity			0-4 days: 34-108 % Activity 5-29 days: 45-145 % Activity 30-89 days: 62-124 % Activity 90-179 days: 48-132 % Activity 180-365 days: 55-127 % Activity 1-5 years: 79-127 % Activity 6-10 years: 63-116 % Activity 11-16 years: 55-99 % Activity 17+ years: 50-150 % Activity		1 stage clotting assay.	Clinical Guide to Laboratory Tests, Tietz, 1995; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	3-500 % Activity	3-500 % Activity
	FA5 Factor 5 Activity Labile Factor Proaccelerin							
Factor V Inhibitor			0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
	FA5IN Factor 5 Inhibitor Factor V Bethesda Titer Bethesda Titer							
Factor VII Activity			0-4 days: 28-104 % Activity 5-29 days: 35-143 % Activity 30-89 days: 42-138 % Activity 90-179 days: 39-143 % Activity 180-365 days: 47-127 % Activity 1-5 years: 55-116 % Activity 6-10 years: 52-120 % Activity 11-16 years: 58-115 % Activity 17+ years: 65-135 % Activity		1 stage clotting assay.	Clinical Guide to Laboratory Tests, Tietz, 1995; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	3-1,000 % Activity	3-1,000 % Activity
	FA7 Factor 7 Activity Stable Factor Proconvertin							
Factor VII Inhibitor			0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
	FA7IN Factor 7 Inhibitor Factor VII Bethesda Titer Bethesda Titer							

Factor VIII Activity	F8 Factor 8 Activity Factor VIII Clotting Anti-Hemophilic Factor	0-4 days: 50-178 % Activity 5-29 days: 50-154 % Activity 30-89 days: 50-157 % Activity 90-179 days: 50-125 % Activity 180-365 days: 50-100 % Activity 1-5 years: 59-142 % Activity 6-10 years: 58-132 % Activity 11-16 years: 53-131 % Activity 17+ years: 50-200 % Activity	<1% severe hemophilia 1- <5% moderate hemophilia 5- <40% mild hemophilia	1 stage clotting assay.	Clinical Laboratory Reference Values. In: Laposata M. eds. <i>Laboratory Medicine: The Diagnosis of Disease in the Clinical Laboratory</i> . McGraw-Hill Education; 2014; Blood, Vol 80, 165-172, Andrew, 1987	1-500 % Activity	1-500 % Activity
Factor VIII Activity, Chromogenic	F8C Factor 8 Activity Factor VIII Clotting Factor VIII Chromogenic	2-5 years: 59-142 % Activity 6-10 years: 58-132 % Activity 11-16 years: 53-131 % Activity 17+ years: 50-200 % % Activity	<1% severe hemophilia 1- <5% moderate hemophilia 5- <40% mild hemophilia	Two stage chromogenic assay, measurement at 405nm.	Clinical Laboratory Reference Values. In: Laposata M. eds. <i>Laboratory Medicine: The Diagnosis of Disease in the Clinical Laboratory</i> . McGraw-Hill Education; 2014; Blood, Vol 80, 165-172, Andrew, 1987	1-200 % Activity	1-200 % Activity
Factor VIII Inhibitor	F8IN Factor 8 Inhibitor Factor VIII Bethesda Titer Bethesda Titer	0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
Factor X Activity	FA10 Factor 10 Activity Stuenkel-Power Factor	0-4 days: 12-68 % Activity 5-29 days: 19-79 % Activity 30-89 days: 31-87 % Activity 90-179 days: 35-107 % Activity 180-365 days: 38-118 % Activity 1-5 years: 58-116 % Activity 6-10 years: 55-101 % Activity 11-16 years: 50-117 % Activity 17+ years: 60-130 % Activity		1 stage clotting assay.	Clinical Guide to Laboratory Tests, Tietz, 1995; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	3-500 % Activity	3-500 % Activity
Factor X Inhibitor	FA10IN Factor 10 Inhibitor Factor X Bethesda Titer Bethesda Titer	0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
Factor XI Activity	FA11 Factor 11 Activity Hemophilia C Plasma Thromboplastin Antecedent (PTA)	0-4 days: 10-66 % Activity 5-29 days: 23-87 % Activity 30-89 days: 27-79 % Activity 90-179 days: 41-97 % Activity 180-365 days: 49-134 % Activity 1-5 years: 56-150 % Activity 6-10 years: 52-120 % Activity 11-16 years: 50-97 % Activity 17+ years: 65-135 % Activity		1 stage clotting assay.	Clinical Guide to Laboratory Tests, Tietz, 1995; Blood, Vol 80, 1998-2005, Andrew, 1992; Blood, Vol 70, 165-172, Andrew, 1987	1-500 % Activity	1-500 % Activity
Factor XI Inhibitor	FA11IN Factor 11 Inhibitor Factor XI Bethesda Titer Bethesda Titer	0.0 Bethesda units	Negative	Bethesda method using a 1 stage clotting assay.			≥0.0 Bethesda units Negative, Positive
Factor XIII Activity	FA13 Factor 13 Screen Factor 13 Qualitative Factor XIII Qualitative Fibrin Stabilization Factor		Present	Clot solubility.			Absent, Present
Fibrinogen, Clottable	FIB Factor I Activity Fibrinogen Activity Claus Fibrinogen	220-410 mg/dL		Mechanical clot detection of dilute plasma in the presence of excess thrombin.	OSU Lab Normal Range Study (05/2003)	60-900 mg/dL	60-900 mg/dL
Fibrinogen, Obstetrical	FIB Fibrinogen, OB	OB 1st Trimester: 244-510 mg/dL OB 2nd Trimester: 291-538 mg/dL OB 3rd Trimester: 373-619 mg/dL	Fibrinogen levels may be altered by the normal physiologic changes of pregnancy and should be interpreted considering reference ranges specific to gestational age.	Mechanical clot detection of dilute plasma in the presence of excess thrombin.	Reference: Abbas-Ghanavati M, Greer LG, Cunningham FG. Pregnancy and laboratory studies: a reference table for clinicians. <i>Obstet Gynecol</i> 2009; 114:1326.	60-900 mg/dL	60-900 mg/dL
Heparin Platelet Factor 4 (HIT Screen) with Reflex to (SRK)	PF4ICP	PF4 IgG OD: <0.400 % Heparin Inhibition: <50		Enzyme-linked immunosorbent assay (ELISA), IgG.	Immucor LIFE CODES® PF4 IgG Assay Package Insert	PF4 IgG OD: 0.000-3.000	PF4 IgG OD: 0.000-3.000
Lupus Anticoagulant	PTLT PTT-LA,DRVVT LA Screen APS Screen Antiphospholipid Antibody Screen Siclet LA DRVVT	PT 11.9-14.2 sec INR 0.9-1.1 TT 13.0-20.0 sec PTT-LA ≥43.6 sec DRVVT SCREEN RATIO ≤1.20 DRVVT NORMALIZED RATIO ≤1.29 HEXAGONAL PL NEUTRALIZATION ≤10.3 sec		Hexagonal PL Neutralization: STACLOT-LA: Mechanical clot detection in the presence/absence of hexagonal ring phospholipid DRVVT: Mechanical clot detection in the presence of Diluted Russell's Viper Venom PTT-LA: Mechanical clot detection in the presence of cephalin and activator PF, TT, Mixing Studies: Mechanical clot detection	OSUWMC Normal Range Study	PT: See individual analyte TT: See individual analyte PTT-LA: 20.0-180.0 sec	PT: See individual analyte TT: See individual analyte PTT-LA: 20.0-180.0 sec

Platelet Aggregation	PLATGG Plt. Agg. Plt. Aggregation	Aggregation w/ADP 5 umol/L: 67-82 % Aggregation Aggregation w/ADP 10.0 umol/L: 73-91 % Aggregation Aggregation w/Arach Acid 0.5 mmol/L: 72-91 % Aggregation Aggregation w/Epinephrine 5 umol/L: 64-105 % Aggregation Aggregation w/Ristocetin 0.5 mg/mL: 0-6 % Aggregation Aggregation w/Ristocetin 1.25 mg/mL: 70-105 % Aggregation Aggregation w/Collagen 2 ug/mL: 80-96 % Aggregation Aggregation w/Collagen 5 ug/mL: 71-92 % Aggregation Aggregation w/TXA 2.0 umol/L: 73-112 % Aggregation ATP Release w/ Collagen 2 ug/mL: 0.26-1.07 nmol ATP Release w/ Collagen 5 ug/mL: 0.49-1.32 nmol ATP Release w/Arach Acid 0.5 mmol/L: 0.37-0.90 nmol ATP Release w/ADP 5 umol/L: 0.28-0.93 nmol ATP Release w/ADP 10 umol/L: 0.44-1.19 nmol ATP Release w/Thrombin IU: 0.36-0.99 nmol ATP Release w/Epinephrine 5 umol/L: 0.35-0.96 nmol		Born method of turbidimetric aggregation with simultaneous measurement of ATP release by platelet lumi-aggregometry.	OSUWMC Normal Range Study (02-2023)		
Platelet Function Test	PFST FEA Screen	Collagen / Epi Closure Time: 73-172 sec Collagen / ADP Closure Time: 53-111 sec		Instrument PFA-100 closure time: the time measured from the start of the test until a platelet to close aperture after exposure to agonist.	OSU Normal Range Study (07/2004)	Collagen / Epi Closure Time: 31-300 sec Collagen / ADP Closure Time: 31-300 sec	Collagen / Epi Closure Time: 31-300 sec Collagen / ADP Closure Time: 31-300 sec
Platelet P2Y12 Inhibition Test	P2Y12 Clopidogrel Inhibition Ticagrelor Inhibition	194-418 PRU		Test results are reported in P2Y12 Reaction Units (PRU). The pre-drug Normal Reference Range is 194 - 418 PRU. PRU measures the extent of platelet aggregation in the presence of P2Y12 inhibitor drugs such as clopidogrel (Plavix), prasugrel (Effient), ticlopidine (Ticlid), and ticagrelor (Brilinta).	Whole blood platelet aggregometry based on the ability of activated platelets to bind fibrinogen. Fibrinogen-coated microparticles aggregate in whole blood in proportion to the number of expressed platelet GP IIb/IIIa receptors.	Accriva Diagnostics Verify Now P2Y12 Package Insert VNI009WEU	0-999 PRU 0-999 PRU
Protein C Activity	PROTC Protein C Functional PC Functional PC Activity	0-4 days: 17-33 % Activity 5-29 days: 20-64 % Activity 30-89 days: 21-65 % Activity 90-179 days: 24-80 % Activity 180-365 days: 37-81 % Activity 1-5 years: 56-150 % Activity 6-10 years: 40-92 % Activity 11-16 years: 55-111 % Activity		This clotting assay measures Protein C which is activated in the presence of the specific activator extracted from Aghastrodion c. constrictor venom. The resulting activated protein C inhibits the factors V and VIII, and thus prolongs the APTT of a system in which all the factors are present, constant and in excess (provided by the test system), except for the protein C which is derived from the sample being tested.	OSUWMC Normal Range Study; Blood, Vol 80, 1998-2005, Andrew, 1992; Amer. Jour. Fed. Hematol. Oncol. Vol 12, 95-104, Andrew, 1990		10-300 % Activity
Protein S Activity	PROTS Protein S Functional PS Functional PS Activity	0-4 days: 12-60 % Activity 5-29 days: 22-78 % Activity 30-89 days: 33-93 % Activity 90-179 days: 54-118 % Activity 180-365 days: 55-119 % Activity 1-5 years: 53-173 % Activity		Clotting assay based on the cofactor activity of protein S which enhances the anticoagulant action of activated protein C. This enhancement is reflected by the prolongation of the clotting time of a system enriched with factor Va which is a physiological substrate for activated protein C.	OSUWMC Normal Range Study; Blood, Vol 80, 1998-2005, Andrew, 1992; Amer. Jour. Fed. Hematol. Oncol. Vol 12, 95-104, Andrew, 1990		10-300 % Activity
Protime - INR	PTI PT/INR Protime/INR	PT 11.9-14.2 sec INR		PT: Mechanical clot detection initiated by Calcium Thromboplastin INR: Calculation	OSUWMC Normal Range Study	PT: 7.0-109.0 sec INR: 0.5-15.2	PT: 7.0-109.0 sec INR: 0.5-15.2
PT and PT Mixing Study	PT PT Mixing PT Inhibitor Screen	PT 11.9-14.2 sec INR 0.9-1.1		PT: Mechanical clot detection initiated by Calcium Thromboplastin INR: Calculation PT Mixing Study: PT performed immediately subsequent to 1:1 mix with normal plasma pool		PT: 7.0-109.0 sec INR: 0.5-15.2	PT: 7.0-109.0 sec INR: 0.5-15.2
PT, INR, PTT	PT/PT PT/INR PTT Protime/INR Activated Partial Thromboplastin Time	PT 11.9-14.2 sec INR PT		PT, PTT: Mechanical clot detection INR: Calculation	See individual analytes	See individual analytes	See individual analytes
PT INR - Stroke	PT/INR	PT 11.9-14.2 sec		PT, PTT: Mechanical clot detection INR: Calculation	OSUWMC Normal Range Study	PT: 7.0-109.0 sec INR: 0.5-15.2	PT: 7.0-109.0 sec INR: 0.5-15.2
PTT	PTT Partial Thromboplastin Time aPTT Activated Partial Thromboplastin Time	24.0-34.3 sec		Mechanical clot detection initiated by Calcium in the presence of cephalin (platelet substitute) and a particulate activator (silica).	OSUWMC Normal Range Study	20.0-180.0 sec	20.0-180.0 sec
PTT with Mixing Study	PTTMS PTT Mixing aPTT Mixing PTT Inhibitor Screen	24.0-34.3 sec		PTT: Mechanical clot detection initiated by Calcium in the presence of cephalin (platelet substitute) and a particulate activator (silica). PTT Mixing Study: PTT performed immediately subsequent to 1:1 mix with normal plasma pool		20.0-180.0 sec	20.0-180.0 sec
ROTEM Main Lab: Heparin Panel	RTMHEP	INTEM CT: 122-208 sec INTEM CFT: 45-110 sec INTEM ALPHA: 70-81 degree INTEM A20: 51-72 mm INTEM MCF: 51-72 mm EXTEM CT: 43-82 sec EXTEM CFT: 48-127 sec EXTEM ALPHA: 65-80 degree EXTEM A20: 50-70 mm EXTEM MCF: 52-70 mm FIBTEM A20: 7-21 mm FIBTEM MCF: 7-21 mm Reference ranges are not available for all ROTEM components unless otherwise noted: Interpretation of ROTEM results must include all available parameters, patient clinical context and current therapy. Interpretation of the HEPTM results must include clinical correlation between the INTEM and the HEPTM data. All results should be interpreted carefully based on patient clinical context and current therapy. Interpretation of APTM results (fibrinolytic activity) must include correlation between the EXTEM and APTM data. All results should be interpreted carefully based on patient clinical context and current therapy.	Rotem Physician Advice	Rotational Viscoelastic Testing: Rotem Delta: The patented ROTEM technology is based on a fixed cylindrical cup and a permanently oscillating vertical axis. The axis is supported by a high precision ball bearing and oscillates to the left and to the right through an angle of 4.75°.	Normal Range Study 2013		

ROTEM Main Lab: Routine Panel	RTMRN	<p>INTEM CT: 122-208 sec INTEM CFT: 45-110 sec INTEM ALPHA: 70-81 degree INTEM A20: 51-72 mm INTEM MCF: 51-72 mm INTEM ML: No reference range available; see comment INTEM LLD: No reference range available; see comment</p> <p>EXTM CT: 43-82 sec EXTM CFT: 48-127 sec EXTM ALPHA: 65-80 degree EXTM A20: 50-70 mm EXTM MCF: 52-70 mm</p> <p>FIBTEM A20: 7-21 mm FIBTEM MCF: 7-21 mm</p> <p>Reference ranges are not available for all ROTEM components unless otherwise noted: Interpretation of ROTEM results must include all available parameters, patient clinical context and current therapy.</p> <p>Interpretation of APTEM results (fibrinolytic activity) must include correlation between the EXTEM and APTEM data. All results should be interpreted carefully based on patient clinical context and current therapy.</p>	<p><u>Rotem Physician Advice</u></p>	<p>Rotational Viscoelastic Testing: Rotem Delta: The patented ROTEM technology is based on a fixed cylindrical cup and a permanently oscillating vertical axis.</p> <p>The axis is supported by a high precision ball bearing and oscillates to the left and to the right through an angle of 4.75°.</p>	Normal Range Study 2013		
ROTEM Main Lab: Trauma Panel	RTMTRA	<p>EXTM CT: 43-82sec EXTM CFT: 48-127 sec EXTM ALPHA: 65-80 degree EXTM A20: 50-70 mm EXTM MCF: 52-70 mm</p> <p>FIBTEM A20: 7-21 mm FIBTEM MCF: 7-21 mm</p> <p>Reference ranges are not available for all ROTEM components unless otherwise noted: Interpretation of ROTEM results must include all available parameters, patient clinical context and current therapy.</p> <p>Interpretation of APTEM results (fibrinolytic activity) must include correlation between the EXTEM and APTEM data. All results should be interpreted carefully based on patient clinical context and current therapy.</p>	<p><u>Rotem Physician Advice</u></p>	<p>Rotational Viscoelastic Testing: Rotem Delta: The patented ROTEM technology is based on a fixed cylindrical cup and a permanently oscillating vertical axis.</p> <p>The axis is supported by a high precision ball bearing and oscillates to the left and to the right through an angle of 4.75°.</p>	Normal Range Study 2013		
Thrombin Time	TT	13.0-20.0 sec		Mechanical clot detection of undiluted plasma in the presence of a predetermined quantity of thrombin, resulting in a fibrin clot.	OSUWMC Normal Range Study	10.0-120.0 sec	10.0-120.0 sec
Thrombin Time w/ Mixing Studies	TTMS TT Mixing TT w/ah Mixing	13.0-20.0 sec		<p>TT: Mechanical clot detection of undiluted plasma in the presence of a predetermined quantity of thrombin, resulting in a fibrin clot.</p> <p>TT Mixing Study: Thrombin Time is performed subsequent to a 1:1 mix with normal plasma pool and also in the presence of heparinase (inflow of medication)</p>	OSUWMC Normal Range Study	10.0-120.0 sec	10.0-120.0 sec
Von Willebrand Battery Ag + Factor VIII	<p>PTT (vW) FR, VWFAC, VWFAG FR, GPIBM, VWFAG VWF Antigen and Factor VIII VWF Immunologic and Factor VIII VWF Antigen and Factor 8 VWF Immunologic and Factor 8 VWF Activity and Factor VIII VWF Activity and Factor 8</p>	<p>PTT 24.0-34.3 sec</p> <p>FAS 0-4 days: 50-178 % Activity 5-29 days: 50-154 % Activity 30-89 days: 50-157 % Activity 90-179 days: 50-125 % Activity 180-365 days: 50-109 % Activity 1-5 years: 59-142 % Activity 6-10 years: 58-132 % Activity 11-16 years: 53-131 % Activity 17+ years: 50-200 % Activity</p> <p>VWFAC 50-203% Activity</p> <p>VWFAG 0-4 days: 50-287 % 5-29 days: 50-254 % 30-89 days: 50-246 % 90-179 days: 50-206 % 180-365 days: 50-197 % 1-5 years: 60-120 % 6-10 years: 44-144 % 11-16 years: 46-153 % 17+ years: 50-180 %</p>		<p>PTT: Mechanical clot detection initiated by Calcium in the presence of cephalin (platelet substitute) and a particulate activator (silica).</p> <p>VWF Ag: Immuno-turbidimetric method (STA - Latex VWF:Ag)</p> <p>Factor VIII: One Stage Clotting Assay</p> <p>VWF Activity: Latex turbidimetric immunoassay (Siemens Innoance® VWF:Ac)</p>	See individual analytes	See individual analytes	See individual analytes
Von Willebrand Factor Ag	VWFAG VWF Antigen VWF Immunologic	<p>0-4 days: 50-287 % 5-29 days: 50-254 % 30-89 days: 50-246 % 90-179 days: 50-206 % 180-365 days: 50-197 % 1-5 years: 60-120 % 6-10 years: 44-144 % 11-16 years: 46-153 % 17+ years: 50-180 %</p>		Immuno-turbidimetric method (STA - Latex VWF:Ag)	OSUWMC Normal Range Study: Blood, Vol 80, 1998-2005; Andrew, 1992; Blood, Vol 70, 165-172; Andrew, 1987		1-400 %
Von Willebrand Factor GPIBM Activity	GPIBM GPIBM von Willebrand VWF VWF Activity	50-203 % Activity		Latex turbidimetric immunoassay (Siemens Innoance® vWF Ac)	Siemens Healthineers Innoance® VWF Ac Package Insert		4-300 % Activity
Amikacin Level, Extended Interval	AMIKEL	Peak: 30-60 mcg/mL, Trough: <6 mcg/mL		Homogeneous particle-enhanced turbidimetric immunoassay.	OSU Pharmacy	3.0-50.0 mcg/mL	3.0-150.0 mcg/mL
Amikacin Level, Peak (Post Drug Level)	AMIKPK	Therapeutic Range: 30.0-60.0 mcg/mL		Homogeneous particle-enhanced turbidimetric immunoassay.	OSU Pharmacy	3.0-50.0 mcg/mL	3.0-150.0 mcg/mL
Amikacin Level, Random	AMIKR	Peak: 30-60 mcg/mL, Trough: <6 mcg/mL		Homogeneous particle-enhanced turbidimetric immunoassay.	OSU Pharmacy	3.0-50.0 mcg/mL	3.0-150.0 mcg/mL
Amikacin Level, Trough (Pre Drug Level)	AMIKTR	Therapeutic Range: <6.0 mcg/mL		Homogeneous particle-enhanced turbidimetric immunoassay.	OSU Pharmacy	3.0-50.0 mcg/mL	3.0-150.0 mcg/mL

Amphetamines, Urine, Confirmation	AMPC Adderall Meth Speed	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Amphetamine: 25 ng/mL Methamphetamine: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	CONFIRMATION (Amphetamine/Methamphetamine): 25-5,000 ng/mL	CONFIRMATION: (Amphetamine/Methamphetamine): 25-5,000 ng/mL
Benzodiazepines, Urine, Confirmation	BENZMS	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: 2-Hydroxyethylflurazepam: 50 ng/mL 7-Aminoclonazepam: 10 ng/mL 7-Amino-flunitrazepam: 10 ng/mL Alpha-hydroxyvalproic acid: 10 ng/mL Alpha-hydroxyvalproic acid: 10 ng/mL Alprazolam: 10 ng/mL Diazepam: 10 ng/mL Flurazepam: 10 ng/mL Midazolam: 10 ng/mL Nefazepam: 10 ng/mL Temazepam: 10 ng/mL Clonazepam: 25 ng/mL Lorazepam: 25 ng/mL Oxazepam: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	2-Hydroxyethylflurazepam: 50-1,750 ng/mL 7-Aminoclonazepam: 10-1,000 ng/mL 7-Amino-flunitrazepam: 10-1,300 ng/mL Alpha-hydroxyvalproic acid: 10-2,000 ng/mL Alpha-hydroxyvalproic acid: 10-1,750 ng/mL Alprazolam: 10-2,000 ng/mL Diazepam: 10-2,000 ng/mL Flurazepam: 10-2,000 ng/mL Midazolam: 10-2,000 ng/mL Nefazepam: 10-2,000 ng/mL Temazepam: 10-2,000 ng/mL Clonazepam: 25-2,000 ng/mL Lorazepam: 25-2,000 ng/mL Oxazepam: 25-2,000 ng/mL	2-Hydroxyethylflurazepam: 50-1,750 ng/mL 7-Aminoclonazepam: 10-1,000 ng/mL 7-Amino-flunitrazepam: 10-1,300 ng/mL Alpha-hydroxyvalproic acid: 10-2,000 ng/mL Alpha-hydroxyvalproic acid: 10-1,750 ng/mL Alprazolam: 10-2,000 ng/mL Diazepam: 10-2,000 ng/mL Flurazepam: 10-2,000 ng/mL Midazolam: 10-2,000 ng/mL Nefazepam: 10-2,000 ng/mL Temazepam: 10-2,000 ng/mL Clonazepam: 25-2,000 ng/mL Lorazepam: 25-2,000 ng/mL Oxazepam: 25-2,000 ng/mL
Buprenorphine and Norbuprenorphine, Urine, Confirmation	QBUPR Subutex Suboxone Buprenex	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Buprenorphine: 5.0 ng/mL Norbuprenorphine: 5.0 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	Buprenorphine: 5.0 - 5,000.0 ng/mL Norbuprenorphine: 5.0 - 5,000.0 ng/mL	Buprenorphine: 5.0 - 25,000.0 ng/mL Norbuprenorphine: 5.0 - 25,000.0 ng/mL
Carboxy THC, Urine Confirmation	THCCON Cannabinoids THC Marijuana Mary Jane	None Detected Cutoff concentrations by gas chromatography-tandem mass spectrometry: 9-Carboxy-11-Nor-Delta-THC: 5.0 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	5.0-500.0 ng/mL	5.0-500.0 ng/mL
Cocaine, Urine, Confirmation	COCCON Coke Crack	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Benzoylecgonine: 25 ng/mL Cocaine: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	Benzoylecgonine: 25-5,000 ng/mL Cocaine: 25-5,000 ng/mL	Benzoylecgonine: 25-25,000 ng/mL Cocaine: 25-25,000 ng/mL
Cyclosporine Level, ZHR	CSAN2 Sandimmune Neoral Gengraf	Therapeutic Range: 320-960 ng/mL			OSU Pharmacy	30-1,500 ng/mL 30-3,000 ng/mL
Cyclosporine Level, Random	CSASR Sandimmune Neoral Gengraf	Kidney/Pancreas Transplant: 0 to 3 months: 800-1000 ng/mL 3 to 12 months: 600-800 ng/mL +12 months: 400-600 ng/mL Kidney/Pancreas Transplant: 0 to 3 months: 800-1000 ng/mL 3 to 12 months: 600-800 ng/mL +12 months: 400-600 ng/mL			OSU Pharmacy	30-1,500 ng/mL 30-3,000 ng/mL
Cyclosporine Level, Trough (Pre Drug Level)	CSANT Sandimmune Neoral Gengraf	Aplastic anemia and stem cell transplant: 200-400			OSU Pharmacy	30-1,500 ng/mL 30-3,000 ng/mL
Ethanol (Alcohol), Urine	ALCOU	<10 mg/dL Negative	Detectable ethanol in urine indicates exposure to ethanol within the past 8-12 hours.	Beckman Coulter DxC700AU; Emi* B Plus Ethyl Alcohol Assay	10-600 mg/dL	10-600 mg/dL
Ethylene Glycol, Blood, Screen with Reflex to Confirmation	EGU	10 mg/dL		Seren; Beckman Coulter DxC700AU; Catachem DiacrefPak Ethylene Glycol Reagent Kit	CONFIRMATION: 10-250 mg/dL	SCREEN: Negative, Presumptive Positive, Confirmation to follow, CONFIRMATION: 10-250 mg/dL
Everolimus, Trough (Pre Drug Level)	EVERTR	Kidney/Pancreas Transplant:		Particle enhanced turbidimetry, immunoassay	OSU Pharmacy	2.0-20.0 ng/mL 2.0-40.0 ng/mL
Fentanyl, Urine, Confirmation	FENTQ GENTEH Garamycin	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Fentanyl: 2.5 ng/mL Norfentanyl: 2.5 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	Fentanyl: 2.5-5,000.0 ng/mL Norfentanyl: 2.5-5,000.0 ng/mL	Fentanyl: 2.5-25,000.0 ng/mL Norfentanyl: 2.5-25,000.0 ng/mL
Gentamicin Level, Extended Infusion	GENTEH Garamycin	Peak: 3-15.0 mcg/mL, Trough: <1.0 mcg/mL		Enzyme immunoassay	OSU Pharmacy	0.3-10.0 mcg/mL 0.3-20.0 mcg/mL
Gentamicin Level, Peak (Post Drug Level)	GENTPK Garamycin	Therapeutic Range: 3.0-15.0 mcg/mL		Enzyme immunoassay	OSU Pharmacy	0.3-10.0 mcg/mL 0.3-20.0 mcg/mL
Gentamicin Level, Random	GENTE Garamycin	Peak: 3-15.0 mcg/mL, Trough: <1.0 mcg/mL		Enzyme immunoassay	OSU Pharmacy	0.3-10.0 mcg/mL 0.3-20.0 mcg/mL
Gentamicin, Trough (Pre Drug Level)	GENTR Garamycin	0-365 days: Therapeutic Range: <1.0 mcg/mL, +1 years: Therapeutic Range: <1.0 mcg/mL, Therapeutic Range: 1.5-5.0 mcg/mL		Enzyme immunoassay	OSU Pharmacy	0.3-10.0 mcg/mL 0.3-20.0 mcg/mL
Lidocaine Level	LIDOU	Therapeutic Range: 1.5-5.0 mcg/mL		Enzyme immunoassay	OSU Pharmacy	0.5-12.0 mcg/mL 0.5-36.0 mcg/mL
Methadone, Urine, Confirmation	METHMS	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Methadone: 25 ng/mL EDDP Methadone: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	Methadone: 25-5,000 ng/mL EDDP Methadone: 25-5,000 ng/mL	Methadone: 25-25,000 ng/mL EDDP Methadone: 25-25,000 ng/mL
Nicotine Screen Urine	NICOTU	None Detected Cutoff concentrations by immunoassay detection:	Positive results indicate recent exposure to cigarette smoke.	Beckman Coulter DxC700AU Thermo Scientific DRP Cotinine Assay		None Detected, Positive, Presumptive Positive, Confirmation to Follow.

Opioids, Urine, Confirmation	OPICON	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: 6-Monoacetylmorphine: 5.0 ng/mL Codeine: 25 ng/mL Hydrocodone: 25 ng/mL Hydroxycodone: 25 ng/mL Morphine: 25 ng/mL Tramadol: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	6-Monoacetylmorphine: 5.0-5,000.0 ng/mL Codeine: 25-5,000 ng/mL Hydrocodone: 25-5,000 ng/mL Hydroxycodone: 25-5,000 ng/mL Morphine: 25-5,000 ng/mL	6-Monoacetylmorphine: 5.0-25,000.0 ng/mL Codeine: 25-25,000 ng/mL Hydrocodone: 25-25,000 ng/mL Hydroxycodone: 25-25,000 ng/mL Morphine: 25-25,000 ng/mL	
Oxycodone, Urine, Confirmation	OXYCON	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: Oxycodone: 25 ng/mL Oxycodone: 25 ng/mL Noroxycodone: 25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	Oxycodone: 25-5,000 ng/mL Oxycodone: 25-5,000 ng/mL Noroxycodone: 25-5,000 ng/mL	Oxycodone: 25-25,000 ng/mL Oxycodone: 25-25,000 ng/mL Noroxycodone: 25-25,000 ng/mL	
Pentobarbital Level	PENTU	Intracranial Pressure Therapy: 30-40 ug/mL		Gas Chromatography (GC)	OSU Pharmacy	3-50 ug/mL	3-50 ug/mL
Phenytoin Free Level	PFPN	Therapeutic Range: 0.6-2.4 mcg/mL		Chemiluminescent microparticle immunoassay	OSU Pharmacy	0.5-40.0 mcg/mL	0.5-40.0 mcg/mL
Phenytoin Free	Dilutin, Free						
Sirolimus (Rapamycin) Level, Random	SIROR	Bone Marrow Transplant: 3-12		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Sirolimus (Rapamycin) Level, Trough (Pre Drug Level)	SIROTR	Bone Marrow Transplant: 3-12		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Tacrolimus Level, Trough (Pre Drug Level)	TACRO	Bone Marrow Transplant: 5-15		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Tacrolimus Level, Trough (Pre Drug Level)	Psoarf	Kidney/Pancreas Transplant: 0 to 3 months: 8-16; 3 to 12 months: 6-8; >12 months: 4-6		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Tacrolimus, Random	TACROR	Bone Marrow Transplant: 5-15		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Tacrolimus, Random	Psoarf	Kidney/Pancreas Transplant: 0 to 3 months: 8-16; 3 to 12 months: 6-8; >12 months: 4-6		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2.0-30.0 ng/mL	2.0-60.0 ng/mL
Tramadol, Urine	TramU Ultram ConZin	None Detected Cutoff concentrations by liquid chromatography-tandem mass spectrometry: <25 ng/mL	Numeric results (positive) indicate the measured concentration of drug in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection. Drug concentrations in urine cannot determine the dose, route of administration, or exact time of exposure.	Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)	25-5,000 ng/mL	25-25,000 ng/mL	
Urine Drug Screen 10		None Detected Cutoff concentrations by immunoassay detection: Amphetamine/Methamphetamine: 500 ng/mL Barbiturates: 200 ng/mL Benzodiazepines: 200 ng/mL Buprenorphine: 5 ng/mL Cannabinoids: 50 ng/mL Cocaine: 150 ng/mL Fentanyl: 1 ng/mL Methadone: 300 ng/mL Opiates: 300 ng/mL Oxycodone: 100 ng/mL	Presumptive positive results indicate the presence of a compound within the specified class of drugs or a structurally related compound in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of detection. The detection of drugs in urine is dependent on several factors such as the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection.	Beckman Coulter Dc700AU; Emit® II Plus Amphetamines Assay, Barbiturates Assay, Benzodiazepine Assay, Buprenorphine Assay, Cannabinoids Assay, Cocaine Metabolite Assay, Methadone Assay, and Opiates Assay; ARK™ Fentanyl II Assay; DRB Oxycodone Assay.		Negative, Positive, Presumptive Positive, Confirmation to Follow.	
Urine Drug Screen 10 with Confirmation	UDRUG	None Detected Cutoff concentrations by immunoassay detection: Amphetamine/Methamphetamine: 500 ng/mL Barbiturates: 200 ng/mL Benzodiazepines: 200 ng/mL Buprenorphine: 5 ng/mL Cannabinoids: 50 ng/mL Cocaine: 150 ng/mL Fentanyl: 1 ng/mL Methadone: 300 ng/mL Opiates: 300 ng/mL Oxycodone: 100 ng/mL	Presumptive positive results indicate the presence of a compound within the specified class of drugs or a structurally related compound in the urine sample. Negative results may reflect the absence of drug or drug present below the assay's limit of quantitation. The detection of drugs in urine is dependent on several factors including the amount of drug ingested, time since exposure, individual metabolism, kidney and liver function, urine concentration, and the assay's limit of detection.	Screen: Beckman Coulter Dc700AU; Emit® II Plus Amphetamines Assay, Barbiturates Assay, Benzodiazepine Assay, Buprenorphine Assay, Cannabinoids Assay, Cocaine Metabolite Assay, Methadone Assay, and Opiates Assay; ARK™ Fentanyl II Assay; DRB Oxycodone Assay. Confirmation: Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS)		See individual reflex tests	
Valproic Acid, Free	VVPA Depakote Divalproex Valproate Depakene	Therapeutic Range: 5-35 mcg/mL		Chemiluminescent microparticle immunoassay	OSU Pharmacy	2-150 mcg/mL	
Valproic Acid, Total	VPA Depakote Divalproex Valproate Depakene	Therapeutic Range: 50-120 mcg/mL		Chemiluminescent microparticle immunoassay		4-150 mcg/mL	4-750 mcg/mL
Volatile Alcohol Panel, Blood	VOLARB Alcohol Drug Methanol Ethanol Isopropanol Acetone Ethyl Alcohol Isopropyl Alcohol	METHANOL <10 mg/dL ETHYL ALCOHOL <10 mg/dL ACETONE <10 mg/dL		Gas Chromatography (GC) Flame Ionization Detection (FID)		METHANOL: 10-400 mg/dL ETHYL ALCOHOL: 10-400 mg/dL ACETONE: 10-400 mg/dL ISOPROPANOL: 10-400 mg/dL	METHANOL: 10-400 mg/dL ETHYL ALCOHOL: 10-400 mg/dL ACETONE: 10-400 mg/dL ISOPROPANOL: 10-400 mg/dL
Adrenal Vein Salivine Pre Stimulation		CORTISOL LEFT AV1 3.09-22.40 mcg/dL CORTISOL RIGHT AV1 3.09-22.40 mcg/dL CORTISOL IVC1 3.09-22.40 mcg/dL CORTISOL LEFT AV2 3.09-22.40 mcg/dL CORTISOL RIGHT AV2 3.09-22.40 mcg/dL CORTISOL IVC2 3.09-22.40 mcg/dL CORTISOL LEFT AV3 3.09-22.40 mcg/dL CORTISOL RIGHT AV3 3.09-22.40 mcg/dL CORTISOL IVC3 3.09-22.40 mcg/dL	Albosterone Reference Range Plasma Upright: <35.30 ng/dL	Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes

			<p>ALDOSTERONE LEFT AV1 <23.20 ng/dL</p> <p>ALDOSTERONE RIGHT AV1 <23.20 ng/dL</p> <p>ALDOSTERONE IVC1 <23.20 ng/dL</p> <p>ALDOSTERONE LEFT AV2 <23.20 ng/dL</p> <p>ALDOSTERONE RIGHT AV2 <23.20 ng/dL</p> <p>ALDOSTERONE IVC2 <23.20 ng/dL</p> <p>ALDOSTERONE LEFT AV3 <23.20 ng/dL</p> <p>ALDOSTERONE RIGHT AV3 <23.20 ng/dL</p> <p>ALDOSTERONE IVC3 <23.20 ng/dL</p>	<p>Plasma Supine: <23.60 ng/dL Serum Upright: <39.20 ng/dL Serum Supine: <23.20 ng/dL</p>				
Adrenal Vein Sampling Additional Pre Stimulation	<p>CORT LAV1 CORT RAV1 CORT IVC1 CORT LAV2 CORT RAV2 CORT IVC2 CORT LAV3 CORT RAV3 CORT IVC3 ALDO LAV1 ALDO RAV1 ALDO IVC1 ALDO LAV2 ALDO RAV2 ALDO IVC2 ALDO LAV3 ALDO RAV3 ALDO IVC3</p> <p>CORT LAV4 CORT RAV4 CORT IVC4 CORT LAV5 CORT RAV5 CORT IVC5 CORT LAV6 CORT RAV6 CORT IVC6 ALDO LAV4 ALDO RAV4 ALDO IVC4 ALDO LAV5 ALDO RAV5 ALDO IVC5 ALDO LAV6 ALDO RAV6 ALDO IVC6</p>		<p>CORTISOL LEFT AV4 3.09-22.40 mcg/dL</p> <p>CORTISOL RIGHT AV4 3.09-22.40 mcg/dL</p> <p>CORTISOL IVC4 3.09-22.40 mcg/dL</p> <p>CORTISOL LEFT AV5 3.09-22.40 mcg/dL</p> <p>CORTISOL RIGHT AV5 3.09-22.40 mcg/dL</p> <p>CORTISOL IVC5 3.09-22.40 mcg/dL</p> <p>CORTISOL LEFT AV6 3.09-22.40 mcg/dL</p> <p>CORTISOL RIGHT AV6 3.09-22.40 mcg/dL</p> <p>CORTISOL IVC6 3.09-22.40 mcg/dL</p> <p>ALDOSTERONE LEFT AV4 <23.20 ng/dL</p> <p>ALDOSTERONE RIGHT AV4</p>	<p>Aldosterone Reference Range</p> <p>Plasma Upright: <35.30 ng/dL Plasma Supine: <23.60 ng/dL Serum Upright: <39.20 ng/dL Serum Supine: <23.20 ng/dL</p>	Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes
Adrenal Vein Sampling Post Stimulation	<p>ALDO LAV 5M ALDO RAV 5M ALDO IVC 5M ALDO LAV 10M ALDO RAV 10M ALDO IVC 10M ALDO LAV 15M ALDO RAV 15M ALDO IVC 15M CORT LAV 5M CORT RAV 5M CORT IVC 5M CORT LAV 10M CORT RAV 10M CORT IVC 10M CORT LAV 15M CORT RAV 15M CORT IVC 15M</p> <p>AFFTM AFFTMR AFP</p>		<p>ALDOSTERONE 5 MIN LEFT AV No established reference range</p> <p>ALDOSTERONE 5 MIN RIGHT AV <23.20 ng/dL</p> <p>ALDOSTERONE 5 MIN IVC AV <23.20 ng/dL</p> <p>ALDOSTERONE 10 MIN LEFT AV <23.20 ng/dL</p> <p>ALDOSTERONE 10 MIN RIGHT AV <23.20 ng/dL</p> <p>ALDOSTERONE 10 MIN IVC AV <23.20 ng/dL</p> <p>ALDOSTERONE 15 MIN LEFT AV <23.20 ng/dL</p> <p>ALDOSTERONE 15 MIN RIGHT AV <23.20 ng/dL</p> <p>ALDOSTERONE 15 MIN IVC AV <23.20 ng/dL</p> <p>CORTISOL 5 MIN LEFT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 5 MIN RIGHT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 5 MIN IVC AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 10 MIN LEFT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 10 MIN RIGHT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 10 MIN IVC AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 15 MIN LEFT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 15 MIN RIGHT AV 3.09-22.40 mcg/dL</p> <p>CORTISOL 15 MIN IVC AV 3.09-22.40 mcg/dL</p>		Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes
AFP Tumor Marker			<8.1 ng/mL		Two-site sandwich immunoassay chemiluminescent	Siemens AFP Package Insert 10995310_E3N Rev. 03, 2020-02	2.2-1,000.0 ng/mL	2.2-1,000,000.0 ng/mL

Aldosterone	ALDOS ALDNIH ALDPRL ALDPRI ALDOPBR ALDPOL ALDPOI LDOPOR			Reference Range: Plasma Upright: <35.20 ng/dL Plasma Supine: <23.60 ng/dL Serum Upright: <39.20 ng/dL Serum Supine: <23.20 ng/dL	Chemluminescent immunoassay	Package Insert	4.00-100.00 ng/dL	4.00-100,000.00 ng/dL
ANA Multiplex Screen	ANAS			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
ANA Multiplex Scm with Reflex	ANASR			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
ANA Screen IFA	ANAB			Negative	Indirect immunofluorescence antibody	INOVA IFU 2018 Revision 7		SCREEN: Negative, Positive QUANTITATIVE: 1:160, 1:160, 1:320, 1:640, 1:1280, >1:2560
ANCA Init Scm (ANCA, PR3AB, MPO)	ANCA Anti-PR3 MPO			Negative	ANCA: Indirect fluorescent antibody MPO & PR3: Multiplex flow immunoassay	INOVA IFU 2019 Revision 3 (ANCA). Instructions for use manual (MPO & PR3)		See individual analytes
Anti Microsomal Antibody	MIAN ATPO	<60.0 IU/mL			Chemluminescent immunoassay	Siemens Anti-Thyroid Peroxidase Package Insert 10995280_EN Rev. 02-2019-07	28.0-1,300.0 IU/mL	28.0-130,000.0 IU/mL
Anti Mitochondrial Antibody	AMA			Negative	Indirect immunofluorescence antibody	INOVA IFU 2019 Revision 21		SCREEN: Negative, Positive, Indeterminate TITER: 1:20, 1:40, 1:80, 1:160, >1:320
Anti Mullerian Hormone		18-19 years: 0.530-10.580 ng/mL 20-24 years: 1.350-16.460 ng/mL 25-29 years: 0.910-13.520 ng/mL 30-34 years: 0.480-11.000 ng/mL 35-39 years: <10.160 ng/mL 40-44 years: <7.360 ng/mL 45-49 years: <3.300 ng/mL 50+ years: <1.000 ng/mL	18+ years: <18.000 ng/mL		One-stop sandwich immunoassay	Siemens Insulin Package Insert 10998438_EN Rev. 01-2023-06	0.043-24.000 ng/mL	0.043-24.000 ng/mL
Anti Neutrophil Cytoplasmic Antibody	AMH			Negative	Indirect immunofluorescence antibody	INOVA IFU 2019 Revision 3		SCREEN: Negative, Positive, Atypical fluorescence present TITER: 1:20, 1:40, 1:80, 1:160, 1:320, 1:640, >1:1280
Anti Parietal Antibody	ANCA			Negative	Indirect immunofluorescence antibody	INOVA IFU 2019 Revision 21		SCREEN: Negative, Positive, Indeterminate TITER: 1:20, 1:40, 1:80, 1:160, >1:320
Anti Smooth Muscle Antibody	PCA			Negative	Indirect immunofluorescence antibody	INOVA IFU 2019 Revision 21		SCREEN: Negative, Positive, Indeterminate TITER: 1:20, 1:40, 1:80, 1:160, >1:320
Anti-Proteinase 3 Ab	PR3AB Anti-PR3			Negative	Multiplex flow immunoassay	IFU		Negative, Positive
Anti-Scleroderma Ab (SCL70)	SCL70T			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
Beta 2 Microglobulin Serum	B2M	0.80-2.34 mg/L			Turbidimetry	Package insert (Insert Code: INS043.OPT.A, Version: 09th August 2018)	0.30-20.00 mg/L	0.30-40.00 mg/L
Calcitonin	CALCIT CALCT	>9.53 pg/mL	<=13.38 pg/mL		Two-site sandwich immunoassay chemiluminescent	Siemens Calcitonin Package Insert RPH11991801_EN Rev. 01-2019-04	5.00-1,800.00 pg/mL	5.00-180,000.00 pg/mL
Centromere B Antibody	CEN1B			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
Chromatin Antibody	CHROMT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
Chronic Hepatitis Panel (Dialysis, ESRD, AKI)	HBSAGS HEP1B5 HEP1B5			HBSAG: Negative HBSAB: Negative HBcGM: Negative HBcAB: Negative	Chemluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes
CMV IgG Ab	CMVG			Negative	Chemluminescent immunoassay	Package Insert		Negative, Positive, Indeterminate
CMV IgM Ab	CMVM			Negative	Chemluminescent immunoassay	Package Insert		Negative, Positive, Indeterminate
CMVG, CMVM, EBV, HSVG12, HSYM	CMVG CMVM EBV HSVG12 HSVM LPEP Cp8			CMVG: Negative CMVM: Negative EBV: Negative HSVG12: Negative HSVM: Negative	CMVM & CMVG: Chemluminescent immunoassay EBV & HSVG12: Multiplex flow immunoassay HSVM: ELISA manual method	See individual analytes	See individual analytes	See individual analytes
C-Peptide	CPEP Cp8	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) - 5 Minutes	CPMSM	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) 0 Minute	CPZBO	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) 2 Minutes	CPEP2M	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) 5 Minutes	CPEP5M	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) 10 Minutes	CP10M	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
C-Peptide Tolerance (Part of the Insulin Glucose Tolerance Battery) 30 Minutes	CP30M	0.81-3.85 ng/mL			Two-site sandwich immunoassay chemiluminescent	Siemens C-Peptide Package Insert 10997742_EN Rev. 03-2021-06	0.05-25.00 ng/mL	0.05-5,000.00 ng/mL
Cryptococcal Antigen	CRAG			Negative	Lateral flow assay	Package Insert		SCREEN: Negative, Positive TITER: 1:2, 1:5, 1:10, 1:20, 1:40, 1:80, 1:160, 1:320, 1:640, 1:1280, >1:2560
Cryptococcus, Antigen CSF	FCRAG			Negative	Lateral flow assay	Package Insert		SCREEN: Negative, Positive TITER: 1:2, 1:5, 1:10, 1:20, 1:40, 1:80, 1:160, 1:320, 1:640, 1:1280, >1:2560
Cyclic Citrullinated Peptide Ab	ACCP Anti-CCP	<5.00 U/mL			Chemluminescent microparticle immunoassay	Siemens ACCP Package Insert Rev. 06-2022-05	0.54-200.00 U/mL	0.54-20,000.00 U/mL
DHEA-Sulfate	DHEAS DHEAS	25.90-460.20 ug/dL	34.50-568.90 ug/dL		Chemluminescent immunoassay	Siemens DHEAS Package Insert 11200382_EN	3.00-1,500.00 ug/dL	3.00-3,000.00 ug/dL
dDNA Antibody	DSDNAB			Negative	Multiplex flow immunoassay	Package Insert Textbook	QUANTITATIVE: 1-300 IU/mL	QUALITATIVE: Negative, Positive, Indeterminate QUANTITATIVE: 1-30,000 IU/mL
EBV VCA IgG Ab	EBVG			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate
EBV VCA IgG and IgM	EBV			EBVG: Negative EBVM: Negative	Multiplex flow immunoassay	See individual analytes	See individual analytes	See individual analytes
EBV VCA IgM Ab	EBVM			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate
ENA Battery (SSA, SSB, SM, RNP)	ENAB PLASHB			RF: Negative SM: Negative SSA: Negative SSB: Negative	Multiplex flow immunoassay	See individual analytes	See individual analytes	See individual analytes
Free Hemoglobin, Plasma	ENAB PLASHB	<5.0 mg/dL			Photometric	HemoCue Oserating Manual	30.0-2,100.0 mg/dL	230.0 mg/dL
G6PD, Qualitative	GFPO			Enzyme Activity Present	Visual Fluorescence	Package Insert		Enzyme Activity Absent, Enzyme Activity Present, Enzyme Activity Indeterminate
Anti-Glomerular Basement Mem Ab	GBMAB			Negative	Multiplex flow immunoassay	Bio-Rad IFU Vasculitis Revision 665-0516E		Negative, Positive
Granul Hormone	GRHR	<=8.88 ng/mL	<1.23 ng/mL		Chemluminescent immunoassay	Package Insert	0.05-80.00 ng/mL	0.05-1,000.00 ng/mL
Hemoglobin A1C	A1CB	4.7-5.6 %			High-Performance Liquid Chromatography (HPLC)	Package Insert	3.5-15.0 %	3.5-15.0 %
Hemoglobin, Fetal	HFB	<1.0 %			High-Performance Liquid Chromatography (HPLC)	Package Insert Textbook	1.0-40.0 %	1.0-40.0 %

Hemoglobinopathy Eval		<p>HEPB: HGBA ≥95.0%</p> <p>HGBA2 2.1-3.3 %</p> <p>HGBF ~1.0 %</p> <p>HGBC ≤0.0%</p> <p>HGBS ≤0.0%</p> <p>HGB OTHER ≤0.0%</p> <p>GCBCHE RBC 0-14 days: 4.12-5.74 M/uL 15-30 days: 3.32-4.80 M/uL 31-60 days: 2.93-3.87 M/uL 61-179 days: 3.45-4.75 M/uL 180 days-2 years: 3.97-5.01 M/uL 2-5 years: 3.84-4.92 M/uL 6-11 years: 3.60-4.96 M/uL 12-17 years: 3.93-4.90 M/uL 18+ years: 3.91-5.04 M/uL</p> <p>HGB 0-7 days: 13.4-20.0 g/dL 8-14 days: 13.4-20.0 g/dL 15-30 days: 10.8-14.6 g/dL 31-60 days: 9.2-11.4 g/dL 61-179 days: 9.9-12.4 g/dL 180 days-2 years: 10.2-12.7 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.6-13.2 g/dL 12-17 years: 10.8-13.3 g/dL 18+ years: 11.4-15.2 g/dL</p> <p>MCH 0-14 days: 31.1-35.9 pg 15-30 days: 30.4-35.3 pg 31-60 days: 28.0-32.5 pg 61-179 days: 24.4-29.5 pg 180 days-2 years: 23.2-27.5 pg 2-5 years: 23.7-28.6 pg 6-11 years: 24.8-29.5 pg 12-17 years: 24.8-30.2 pg 18+ years: 25.9-33.9 pg</p> <p>MCV 0-14 days: 92.7-106.4 fL 15-30 days: 90.1-103.0 fL 31-60 days: 83.4-96.4 fL 61-179 days: 74.8-88.3 fL 180 days-2 years: 71.3-82.6 fL 2-5 years: 72.3-85.0 fL 6-11 years: 75.9-87.6 fL 12-17 years: 76.9-90.6 fL 18+ years: 79.6-97.7 fL</p> <p>RDW 0-14 days: 14.6-17.3 % 15-30 days: 14.4-16.2 % 31-60 days: 13.6-15.8 % 61-179 days: 12.2-14.3 % 180 days-2 years: 12.7-15.1 % 2-5 years: 12.4-14.9 % 6-11 years: 12.2-14.4 % 12-17 years: 12.3-14.6 % 18+ years: 10.8-14.9 %</p>	<p>HEPB: HGBA ≥95.0%</p> <p>HGBA2 2.1-3.3 %</p> <p>HGBF ~1.0 %</p> <p>HGBC ≤0.0%</p> <p>HGBS ≤0.0%</p> <p>HGB OTHER ≤0.0%</p> <p>GCBCHE: RBC 0-14 days: 4.10-5.55 M/uL 15-30 days: 3.16-4.63 M/uL 31-60 days: 3.02-4.22 M/uL 61-179 days: 3.43-4.80 M/uL 180 days-2 years: 4.03-5.07 M/uL 2-5 years: 3.89-4.97 M/uL 6-11 years: 3.96-5.03 M/uL 12-17 years: 4.03-5.29 M/uL 18+ years: 4.38-5.83 M/uL</p> <p>HGB 0-7 days: 13.9-19.1 g/dL 8-14 days: 13.9-19.1 g/dL 15-30 days: 10.0-15.3 g/dL 31-60 days: 8.9-12.7 g/dL 61-179 days: 9.6-12.4 g/dL 180 days-2 years: 10.1-12.5 g/dL 2-5 years: 10.2-12.7 g/dL 6-11 years: 10.7-13.4 g/dL 12-17 years: 11.0-14.5 g/dL 18+ years: 13.4-16.8 g/dL</p> <p>MCH 0-14 days: 31.3-35.6 pg 15-30 days: 29.9-34.1 pg 31-60 days: 27.8-32.0 pg 61-179 days: 24.4-28.9 pg 180 days-2 years: 22.7-27.2 pg 2-5 years: 23.7-28.3 pg 6-11 years: 24.9-29.2 pg 12-17 years: 25.2-30.2 pg 18+ years: 26.1-33.3 pg</p> <p>MCV 0-14 days: 91.3-103.1 fL 15-30 days: 89.4-99.7 fL 31-60 days: 84.3-94.2 fL 61-179 days: 74.1-87.5 fL 180 days-2 years: 69.5-81.7 fL 2-5 years: 71.3-84.0 fL 6-11 years: 74.4-86.1 fL 12-17 years: 76.7-89.2 fL 18+ years: 79.6-94.5 fL</p> <p>RDW 0-14 days: 14.8-17.0 % 15-30 days: 14.3-16.8 % 31-60 days: 13.8-16.1 % 61-179 days: 12.4-15.1 % 180 days-2 years: 12.9-15.6 % 2-5 years: 12.5-14.9 % 6-11 years: 12.3-14.1 % 12-17 years: 12.4-14.5 % 18+ years: 10.9-14.3 %</p>		<p>HEPB High-Performance Liquid Chromatography (HPLC)</p> <p>GCBCHE The Sysmex XN performs hematology analysis according to the hydrodynamic focusing (DC detection), flow cytometry method (semiconductor laser), and sodium lauryl sulfate (SLS) hemoglobin method</p>	<p>Package Insert Textbook</p>	<p>HGBA2: 1.0-7.0 % HGBF: 1.0-40.0 % RBC: See individual analyte HGB: See individual analyte MCH: See individual analyte MCV: See individual analyte RDW: See individual analyte</p>	<p>HGBA2: 1.0-7.0 % HGBF: 1.0-40.0 % RBC: See individual analyte HGB: See individual analyte MCH: See individual analyte MCV: See individual analyte RDW: See individual analyte</p>
	HEPB GCBCHE Abnormal HGB Detection HEPB							
Hepatitis A Ab, Total (IgG + IgM)	HAACM		Negative	Competitive direct chemiluminescent immunoassay	Package Insert		Negative, Positive, Indeterminate	
Hepatitis A IgM Ab	HAABM		Negative	Two-step IgM capture immunoassay	Package Insert		Negative, Positive, Indeterminate	
Hepatitis B Core Ab, Total (IgG + IgM)	HBCBM		Negative	Two-well antigen sandwich immunoassay	Package Insert		Negative, Positive, Indeterminate	
Hepatitis B Core IgM Ab	HBCBM		Negative	Two-step IgM capture immunoassay	Package Insert		Negative, Positive, Indeterminate	
Hepatitis B Surface Antibody	HBSAB		Negative	Sandwich direct chemiluminescent immunoassay	Package Insert		Negative, Positive	
Hepatitis B Surface Antigen	HBSAG		Negative	Sandwich direct chemiluminescent immunoassay	Package Insert		Negative, Positive, initially reactive, to be confirmed by neutralization	
Hepatitis Battery, Acute	HEPB HBSAG HBCBM		HBSAG: Negative HBCGM: Negative HBCBM: Negative HAABM: Negative HBAB: Negative	Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes	
Hepatitis Battery, Chronic	HEPB HBSAG HBCBM		HBSAG: Negative HBSAB: Negative HBCGM: Negative HCBAB: Negative	Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes	
Hepatitis Be Antibody	HEPB							
Hepatitis Be Antigen	HBEAG		Non Reactive	Chemiluminescent immunoassay	Package Insert		Nonreactive, Reactive	
Hepatitis C Antibody	HCAB		Negative	Indirect sandwich immunoassay	Package Insert		Negative, Positive, Indeterminate	
Hepatitis Infections	HEPB HBSAG HBCGM HBSAB		HBSAG: Negative HBCGM: Negative HBSAB: Negative	Chemiluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes	
Histoplasma Antigen, Urine	HIS1G		Negative	Enzyme Immunoassay	Package Insert	0.20-25.00	Not Detected, Detected	
HIV 1 and 2 Antibody/p24 Antigen	HIV12		Negative	Two-step sandwich immunoassay	Package Insert		Non Reactive, Reactive	
HIV-1 HIV-2 Differentiation	HIV12C		HIV-1: Non Reactive HIV-2: Non Reactive	Immunochromatographic assay	Package Insert		HIV-1: Non Reactive, Reactive, Indeterminate HIV-2: Non Reactive, Reactive, Indeterminate	
HPV Human Papilloma Virus Genotyping w/o Pap Smear	HPVG		HPVG16: Negative HPVG18: Negative HPVG19: Negative	Polymerase chain reaction (PCR)	Package Insert		HPVG16: Negative, Positive HPVG18: Negative, Positive HPVG19: Negative, Positive	
			The assay detects HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 59, 66, 68.					

HPV Testing, Reflex to Cytology for All Positive Results	HPVGSF HPVCYT HPV			HPVGI6: Negative HPVGI8: Negative HPVGI9: Negative The assay detects HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66, 68	Polymerase chain reaction (PCR)	Package Insert			HPVGI6: Negative, Positive HPVGI8: Negative, Positive HPVGI9: Negative, Positive
HPV Testing, Reflex to Cytology for HPV Positive Other Category	HPVGOIH HPVCYT HPV			HPVGI6: Negative HPVGI8: Negative HPVGI9: Negative The assay detects HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66, 68	Polymerase chain reaction (PCR)	Package Insert			HPVGI6: Negative, Positive HPVGI8: Negative, Positive HPVGI9: Negative, Positive
HSV 1 and 2 IgG Antibody	HSVGI2			HSVGI1: Negative HSVGI2: Negative	Multiplex flow immunoassay	Package Insert Textbook			HSVGI1: Negative, Positive, Indeterminate HSVGI2: Negative, Positive, Indeterminate
HSV 2 IgG Antibody	HSVGI2			Negative	Multiplex flow immunoassay	Package Insert Textbook			Negative, Positive, Indeterminate
HSV IgM Antibody	HSVIM			Negative	ELISA manual method	Textbooks, HSV-1 & 2 IgG procedure, Package Insert			Negative, Positive, Equivocal
Immune Status Hepatitis Battery	HEP2			HBCGM: Negative HBSAB: Negative HCBAB: Negative	Chemluminescent immunoassay	See individual analytes	See individual analytes	See individual analytes	See individual analytes
Immuno-fixation, Serum	SIMFX SIMFXB Serum Protein Electrophoresis with Immuno-fixation and Serum Total Protein		MPRO1 ≤0.0 mg/dL MPRO2 ≤0.0 mg/dL MPRO3 ≤0.0 mg/dL MPRO4 ≤0.0 mg/dL		Capillary electrophoresis	Package Insert			MPRO1: ≤0.0 mg/dL MPRO2: ≤0.0 mg/dL MPRO3: ≤0.0 mg/dL MPRO4: ≤0.0 mg/dL
Immunoglobulin Free Chains			KFLC 3.9-26.0 mg/L LFLC 6.4-22.1 mg/L RATIO 0.51-1.72		Turbidimetry	2017 OSU Study	KFLC: 2.9-127.0 mg/L LFLC: 5.2-139.0 mg/L	KFLC: 0.6-63,500.0 mg/L LFLC: 1.3-139,000.0 mg/L	
Immunoglobulin IgE	IgE IgE, Total		≤165.3 IU/mL		Two-site sandwich direct chemiluminescent immunoassay	Reference Range Study 11.3.2016	2.5-3,000.0 IU/mL	2.5-3,000.0 IU/mL	
Insulin	INSL		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance Minus 5 Min	INSL1		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance 0 Minute	INSL2		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance 2 Minute	INSL3		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance 5 Minute	INSL4		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance 10 Minute	INSL5		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin Glucose Tolerance 30 Minute	INSL6		3.0-25.0 uIU/mL		Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05_2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Insulin-Like Growth Factor 1	IGF1 Somatomedin C		All Reference Ranges		Chemiluminescent immunoassay	Package Insert	10.0-1,000.0 ng/mL	10.0-1,000.0 ng/mL	
Intact PTH (Intraoperative)	PTH Intact PTH Rapid Parathyroid Hormone		14.0-72.0 pg/mL		Two-site sandwich immunoassay	Atellica product insert (PTH) REV 4, 2020/11	6.3-2,000.0 pg/mL	6.3-160,000.0 pg/mL	
JO-1 Antibody	JO1 JO1			Negative	Multiplex flow immunoassay	Package Insert Textbook			Negative, Positive
Lyme Antibody	LYME			Negative	Chemiluminescent immunoassay	Package Insert			Negative, Positive
Legionella Urinary Ag	LYME LEGUMI			Negative	Ezyme Immunoassay (EIA)	Package Insert			Negative, Positive
M Tuberculosis By Quantiferon	QFTB QFT M. Tuberculosis Antigen			Negative	Chemiluminescent immunoassay	Package Insert			MTUBERQUANT: Negative, Positive, Indeterminate MTB TB1-NIL: 0.00-10.00 IU/mL MTB TB2-NIL: 0.00-10.00 IU/mL MTB MITOGEN-NIL: 0.00-10.00 IU/mL MTB-NIL: 0.00-10.00 IU/mL
Maternal Hepatitis B Surface Ag	MHBSAG			Negative	Sandwich direct chemiluminescent immunoassay	Package Insert			Negative, Positive, Initially reactive, to be confirmed by neutralization
Maternal Rubella, IgG/Hep B Ag	MHBSAG RUBAB			MHBSAG: Negative RUBAB: Positive	Multiplex flow immunoassay	See individual analytes	See individual analytes	See individual analytes	See individual analytes
Monoclonal Prot Inf, Urine 24 Hr			MPROU1 ≤0 mg/24hrs MPROU2 ≤0 mg/24hrs MPROU3 ≤0 mg/24hrs PROTA 40-225 mg/24 hrs		Electrophoresis	Package Insert			
			PROTEIN 0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL	PROTEIN 0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL					
			ALBUMIN 3.5-5.0 g/dL	ALBUMIN 3.5-5.0 g/dL					
			ALPHA 1 0.2-0.4 g/dL	ALPHA 1 0.2-0.4 g/dL					
			ALPHA 2 0.5-1.0 g/dL	ALPHA 2 0.5-1.0 g/dL					
			BETA ...	BETA ...					

Monoclonal Pro Immuno Serum	<p>0.5-1.1 g/dL</p> <p>GAMMA 0.6-1.5 g/dL</p> <p>IGA 0-30 days: 0-10 mg/dL 31-182 days: <42 mg/dL 183-365 days: 6-69 mg/dL 1-3 years: 15-111 mg/dL 4-6 years: 23-166 mg/dL 7-9 years: 28-180 mg/dL 10-12 years: 55-193 mg/dL 13-15 years: 62-241 mg/dL 16-18 years: 69-262 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p> <p>IGG 0-30 days: 163-872 mg/dL 31-182 days: 311-664 mg/dL 183-365 days: 325-647 mg/dL 1-3 years: 451-1,202 mg/dL 4-6 years: 560-1,219 mg/dL 7-9 years: 485-1,473 mg/dL 10-12 years: 586-1,609 mg/dL 13-15 years: 649-1,640 mg/dL 16-18 years: 804-1,817 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p> <p>IGM 0-29 days: 1-57 mg/dL 30-182 days: <128 mg/dL 183-365 days: <131 mg/dL 1-3 years: 35-184 mg/dL 4-6 years: 42-184 mg/dL 7-9 years: 30-165 mg/dL 10-12 years: 42-211 mg/dL 13-15 years: 34-225 mg/dL 16-18 years: 45-224 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p> <p>MPRO1 ≤0.0 mg/dL</p> <p>MPRO2 ≤0.0 mg/dL</p> <p>MPRO3 ≤0.0 mg/dL</p> <p>MPRO4 ≤0.0 mg/dL</p> <p>SERUM TP PSE QMM SIMFX</p>	<p>0.5-1.1 g/dL</p> <p>GAMMA 0.6-1.5 g/dL</p> <p>IGA 0-30 days: ≤11 mg/dL 31-182 days: <40 mg/dL 183-365 days: 1-82 mg/dL 1-3 years: 9-137 mg/dL 4-6 years: 44-187 mg/dL 7-9 years: 58-204 mg/dL 10-12 years: 46-218 mg/dL 13-15 years: 29-251 mg/dL 16-18 years: 68-259 mg/dL 19-59 years: 66-433 mg/dL 60+ years: 90-410 mg/dL</p> <p>IGG 0-30 days: 197-833 mg/dL 31-182 days: 140-523 mg/dL 183-365 days: 130-823 mg/dL 1-3 years: 413-1,112 mg/dL 4-6 years: 468-1,328 mg/dL 7-9 years: 582-1,441 mg/dL 10-12 years: 685-1,620 mg/dL 13-15 years: 590-1,600 mg/dL 16-18 years: 522-1,703 mg/dL 19-59 years: 600-1,714 mg/dL 60+ years: 600-1,560 mg/dL</p> <p>IGM 0-29 days: 0-65 mg/dL 30-182 days: 6-84 mg/dL 183-365 days: 15-117 mg/dL 1-3 years: 30-146 mg/dL 4-6 years: 31-151 mg/dL 7-9 years: 21-140 mg/dL 10-12 years: 27-151 mg/dL 13-15 years: 26-184 mg/dL 16-18 years: 28-179 mg/dL 19-59 years: 45-281 mg/dL 60+ years: 30-360 mg/dL</p> <p>MPRO1 ≤0.0 mg/dL</p> <p>MPRO2 ≤0.0 mg/dL</p> <p>MPRO3 ≤0.0 mg/dL</p> <p>MPRO4 ≤0.0 mg/dL</p>		Capillary electrophoresis	See individual analytes	See individual analytes	See individual analytes		
Mumps, IgG Ab, Immune Status	MUMPSB			Positive	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate	
Myeloperoxidase Antibodies	MPO			Negative	Multiplex flow immunoassay	Textbook (PI)		Negative, Positive	
Post-Distal Splenic Artery	DISTSPART30 DISTSPART60 DISTSPART90 DISTPART120 DISTPART180	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Post Gastrooduodenal Artery	GASTART30 GASTART60 GASTART90 GASTART120 GASTART180	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Post Miscellaneous Artery	MISC30 MISC60 MISC90 MISC120 MISC180	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Post Proximal Splenic Artery	PROXSPART30 PROXSPART60 PROXSPART90 PROXSPART120 PROXSPART180	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Post Superior Mesenteric Artery	SUPMAR30 SUPMAR60 SUPMAR90 SUPMAR120 SUPMAR180	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Pre-Distal Splenic Artery	DISTSPARTB1 DISTSPARTB2	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Pre Gastrooduodenal Artery	GASTARTB1 GASTARTB2	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Pre Miscellaneous Artery	MISCB1 MISCB2	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Pre-Proximal Splenic Artery	PROXSPARTB1 PROXSPARTB2	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Pre Superior Mesenteric Artery	SUPMARTB1 SUPMARTB2	3.0-25.0 uIU/mL			Two-site sandwich chemiluminescent immunoassay	Siemens Insulin Package Insert 10997752_EN Rev. 05, 2021-06	0.5-300.0 uIU/mL	0.5-1,500.0 uIU/mL	
Protein Electrophoresis Serum, with Reflex	<p>PROTEIN 0-30 days: 4.2-6.2 g/dL 31-182 days: 4.4-6.6 g/dL 183-365 days: 5.6-7.9 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p> <p>ALBUMIN 3.5-5.0 g/dL</p> <p>ALPHA 1 0.2-0.4 g/dL</p> <p>ALPHA 2 0.5-1.0 g/dL</p> <p>SERUM TP SPE Serum Electrophoresis PSE Serum Protein Electrophoresis with Reflex to immunofixation and serum total protein</p> <p>GAMMA (0.4, 0.5, 0.6)</p>	<p>PROTEIN 0-30 days: 4.1-6.3 g/dL 31-182 days: 4.7-6.7 g/dL 183-365 days: 5.5-7.0 g/dL 1-18 years: 5.7-8.0 g/dL 19+ years: 6.4-8.3 g/dL</p> <p>ALBUMIN 3.5-5.0 g/dL</p> <p>ALPHA 1 0.2-0.4 g/dL</p> <p>ALPHA 2 0.5-1.0 g/dL</p> <p>BETA 0.5-1.1 g/dL</p> <p>GAMMA (0.4, 0.5, 0.6)</p>				<p>PSE Capillary electrophoresis</p> <p>PROTEIN Colorimetric with cupric ions in an alkaline solution</p>	Package Insert	<p>PROTEIN: 3.0-12.0 g/dL ALBUMIN: <0.0 g/dL ALPHA 1: >0.0 g/dL ALPHA 2: >0.0 g/dL BETA: >0.0 g/dL GAMMA: >0.0 g/dL</p>	<p>PROTEIN: 3.0-24.0 g/dL ALBUMIN: >0.0 g/dL ALPHA 1: >0.0 g/dL ALPHA 2: >0.0 g/dL BETA: >0.0 g/dL GAMMA: >0.0 g/dL</p>
PTH Intact	PTH Parathyroid Hormone	14.0-72.0 pg/mL			Two-site sandwich immunoassay	Alicella product insert (PTH) REV 4, 2020/11	6.3-2,000.0 pg/mL	6.3-160,000.0 pg/mL	
Renin	RENIN Direct	0.39 years: 4.2-32.2 pg/mL 40+ years: 3.6-81.6 pg/mL			Chemiluminescent immunoassay	Package Insert	2.1-300.0 pg/mL	2.1-3,000.0 pg/mL	
Ribosomal P Antibody	RIBOPT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive	
RNP Antibody	RNP			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive	

RPR	Rapid Plasma Reagin RPRC			Non Reactive	Macroscopic nonreproducible flocculation	Package Insert Textbook		SCREEN: Non Reactive, Reactive TITER: 1:1, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, 1:256, 1:512, 1:1024, ≥1:2048
RPR with Titer	RPR T			Non Reactive	Macroscopic nonreproducible flocculation	Package Insert Textbook		SCREEN: Non Reactive, Reactive TITER: 1:1, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, 1:256, 1:512, 1:1024, ≥1:2048
Rubella IgG With Reflex To IgM	RUBAGR			Positive	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate
Rubella Immune Status IgG Antibody	RUBAG			Positive	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate
Rubella IgG Ab (Immune Status)	RUBOH			Positive	Multiplex flow immunoassay	Package Insert Textbook		SCREEN: Non Reactive, Reactive TITER: 1:1, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, 1:256, 1:512, 1:1024, ≥1:2048
Sex Hormone Binding Globulin	SHBG	18.00-144.00 nmol/L	10.00-57.00 nmol/L		Sandwich immunoassay	BioPlex 2200 MMRV IgG Procedure March 2010 Siemens IMMULITE 2000 SHBG (PIL2KSH-20, 20) (4.01-13)	1.60-180.00 nmol/L	1.60-360.00 nmol/L
Sm Antibody	SMA SMT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
SmRNP Antibody	SMRNPT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
SS-A/Ro Antibody	SSAT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
SS-B/LA Antibody	SSBT			Negative	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive
Syphilis Ab w/ Reflex RPR	SYPIT Syphilis IgG/ObM Antibody			Non Reactive	Direct sandwich assay	Package Insert		SYPHG: Non Reactive, Reactive, Equivocal RPR SCREEN: Non Reactive, Reactive RPR TITER: 1:1, 1:2, 1:4, 1:8, 1:16, 1:32, 1:64, 1:128, 1:256, 1:512, 1:1024, ≥1:2048
Toxoplasma IgG Antibody	TOXG			Negative	Sandwich direct chemiluminescent immunoassay	Package Insert		Negative, Positive, Equivocal
Urine Immunofixation, Random	UMFXR Monoclonal Prot Immunf, Urine - Random UMFXR				Electrophoresis	Package Insert		
Varicella IgG Ab (Immune Status)	VZVIB			Positive	Multiplex flow immunoassay	Package Insert Textbook		Negative, Positive, Indeterminate
Vitamin D (25-Hydroxy, Total)	D25OH Vitamin D, Total	10.0-100.0 ng/mL			Chemiluminescent immunoassay	Package Insert	4.0-150.0 ng/mL	4.0-150.0 ng/mL
Vitamin D, (1,25 Dihydroxy)	D125 ABORH	20.0-79.0 pg/mL			In vitro chemiluminescent immunoassay	Package Insert	3.0-180.0 pg/mL	3.0-540.0 pg/mL
ABORH(D) Typing	Blood Type			A, B, O, AB Rh Negative, Rh Positive	Agglutination			
ABORH - Not Valid for Transfusion	ABORHD ABOD Blood Type			A, B, O, AB Rh Negative, Rh Positive	Agglutination			
Antibody Titer				Negative	Agglutination	Alloimmunization Committee		
Antigen Testing	ABTT AGID Antigen typing			Positive or Negative for each antigen	Agglutination			
Baby Type and DAT (Direct Antiglobulin Test)				ABORH A, B, O, AB Rh Negative, Rh Positive	Agglutination			
Cold Agglutinin Titer	HEELS Heelstick Evaluation COLD			DAT Negative -32	Agglutination			
Cord Blood Evaluation				ABORH A, B, O, AB Rh Negative, Rh Positive	Agglutination			
Cordocentesis	CORDB			DAT Negative	Agglutination			
Direct Antiglobulin Test (DAT)	CORDD DATO DAT			ABORH A, B, O, AB Rh Negative, Rh Positive	Agglutination			
Rhoig Evaluation	RHEV Rhoam Evaluation			ABORH A, B, O, AB Rh Negative, Rh Positive Fetal Screen Negative	Agglutination			
Transfusion Reaction Battery				See Transfusion Reaction Path Interpretation	Agglutination			
Type and Screen	TXN			ABORH A, B, O, AB Rh Negative, Rh Positive	Agglutination			
Type and Screen - Not for Transfusion				ANTIBODY SCREEN Negative	Agglutination			
Type and Screen - Preadmission	TYSC			ABORH A, B, O, AB Rh Negative, Rh Positive ANTIBODY SCREEN Negative	Agglutination			
BMT Panel 1	XMPO				Flow cytometry immunophenotyping			
CD19 CAR T	BMTIR CD14 CAR19T CD19 probe CAR 19 CART19 Chemere antigen receptor T detection				Flow cytometry immunophenotyping	Known value, circulating CAR T can only be found in infused patients		CD19PERCENTL: 0.1-100.0 % CD19PERTOT: 0.1-100.0 % CAR19PERCTL: 0.1-100.0 % CAR19PERTOT: 0.0-100.0 % CAR19PERCD: 0.1-100.0 %
CD45N	CD45N CD3 Enumeration				Flow cytometry immunophenotyping			

<p>Immunodeficiency Batt Plus NK</p>		<p>AlphaBeta 47.0-80.0%</p> <p>CD13 No established reference range</p> <p>CD14 No established reference range</p> <p>CD19 2.0-21.0%</p> <p>CD2 70.0-92.0%</p> <p>CD20 2.0-21.0%</p> <p>CD3 59.0-92.0%</p> <p>CD4 32.0-62.0%</p> <p>CD45RA No established reference range</p> <p>CD45RO No established reference range</p> <p>CD56/16 3.0-25.0%</p> <p>CD8 11.0-40.0%</p> <p>GammaDelta 0.0-.7.0%</p> <p>HLA DR No established reference range</p>	<p>AlphaBeta 50-2,158 ABS/mm³</p> <p>CD13 No established reference range</p> <p>CD14 No established reference range</p> <p>CD19 17-750 ABS/mm³</p> <p>CD2 581-3,284 ABS/mm³</p> <p>CD20 17-750 ABS/mm³</p> <p>CD3 490-3,284 ABS/mm³</p> <p>CD4 266-2,213 ABS/mm³</p> <p>CD45RA No established reference range</p> <p>CD45RO No established reference range</p> <p>CD56/16 25-893 ABS/mm³</p> <p>CD8 91-1,425 ABS/mm³</p> <p>GammaDelta 0-135 ABS/mm³</p> <p>HLA DR No established reference range</p>		<p>Flow cytometry immunophenotyping</p>			
<p>Immunophenotyping</p>	<p>PBIDB2</p>	<p>CD10 No established reference range</p> <p>CD13 No established reference range</p> <p>CD14 No established reference range</p> <p>CD19 2.0-21.0%</p> <p>CD2 70.0-92.0%</p> <p>CD20 2.0-21.0%</p> <p>CD23 No established reference range</p> <p>CD3 59.0-92.0%</p> <p>CD4 32.0-62.0%</p> <p>CD5 No established reference range</p> <p>CD56/16 3.0-25.0%</p> <p>CD7 No established reference range</p> <p>CD8 11.0-40.0%</p> <p>HLA DR No established reference range</p> <p>KAPPA No established reference range</p> <p>LAMBDA No established reference range</p>	<p>CD10 No established reference range</p> <p>CD13 No established reference range</p> <p>CD14 No established reference range</p> <p>CD19 17-750 ABS/mm³</p> <p>CD2 581-3,284 ABS/mm³</p> <p>CD20 17-750 ABS/mm³</p> <p>CD23 No established reference range</p> <p>CD3 490-3,284 ABS/mm³</p> <p>CD4 266-2,213 ABS/mm³</p> <p>CD5 No established reference range</p> <p>CD56/16 25-893 ABS/mm³</p> <p>CD7 No established reference range</p> <p>CD8 91-1,425 ABS/mm³</p> <p>HLA DR No established reference range</p> <p>KAPPA No established reference range</p> <p>LAMBDA No established reference range</p>		<p>Flow cytometry immunophenotyping</p>			
	<p>GPP LNDATP FLOW PBFP BMFP</p>							

<p>TCRV Beta by Flow Cytometry, Blood</p>	<p>TXT TCRVB TCR PBTCR TRBAT</p>	<p>CD13/14 No established reference range</p> <p>CD19 2.0-21.0% 17-750 ABS/mm3</p> <p>CD2 70.0-92.0% 581-3,284 ABS/mm3</p> <p>CD26 No established reference range</p> <p>CD3 59.0-62.0% 490-3,284 ABS/mm3</p> <p>CD4 32.0-62.0% 266-2,213 ABS/mm3</p> <p>CD56/16 3.0-25.0% 25-893 ABS/mm3</p> <p>CD7 No established reference range</p> <p>CD8 11.0-40.0% 91-1,425 ABS/mm3</p> <p>Vb1 1.89-11.70%</p> <p>Vb11 0.25-5.11%</p> <p>Vb12 1.00-4.76%</p> <p>Vb13.1 1.62-8.16%</p> <p>Vb13.2 0.90-5.28%</p> <p>Vb13.6 0.94-8.80%</p> <p>Vb14 1.33-8.03%</p> <p>Vb16 0.42-1.96%</p> <p>Vb17 2.28-12.61%</p> <p>Vb18 0.58-5.23%</p> <p>Vb2 4.03-23.48%</p> <p>Vb20 0.00-9.73%</p> <p>Vb21.3 1.08-5.97%</p> <p>Vb22 1.99-9.89%</p> <p>Vb23 0.26-4.76%</p> <p>Vb3 0.52-15.71%</p> <p>Vb4 0.79-3.26%</p> <p>Vb5.1 3.19-14.93%</p> <p>Vb5.2 0.49-4.98%</p> <p>Vb5.3 0.37-2.98%</p> <p>Vb7.1 0.64-20.01%</p> <p>Vb7.2 0.05-5.45%</p> <p>Vb8 2.26-29.47%</p> <p>Vb9 1.10-9.30%</p> <p>CD4CD3 DUAL (T HELPER)%</p>		<p>Flow cytometry immunophenotyping</p>	<p>Beckman Coulter-verified</p>		
<p>Transplant Battery</p>				<p>Flow cytometry immunophenotyping</p>	<p>OSU Flow Lab established</p>		

Transplant Battery Plus	PBSOT Transplant battery plus CD20	<p>CD19 2.0-21.0% 17-750 ABS/mm3</p> <p>CD2 70.0-92.0% 584-3,284 ABS/mm3</p> <p>CD20 2.0-21.0% 17-750 ABS/mm3</p> <p>CD3 59.0-92.0% 490-3,284 ABS/mm3</p> <p>CD4 32.0-62.0% 266-2,213 ABS/mm3</p> <p>CD8 11.0-40.0% 91-1,428 ABS/mm3</p>		Flow cytometry immunophenotyping	OSU Flow Lab established		
Acid Fast Culture	AFB/CX	Smear: No acid fast bacilli seen Culture: No growth		Organisms growing in pure culture are identified to the species level whenever possible.	Smear Culture Susceptibility (if appropriate)		SMEAR: No acid Fast Bacilli Seen Acid Fast Bacilli, Few Acid fast Bacilli, Rare Acid Fast Bacilli, Moderate Acid Fast Bacilli, Heavy
Acid Fast Culture, Tissue		No acid fast bacilli seen.		Organisms growing in pure culture are identified to the species level whenever possible.	Culture		SMEAR: No acid Fast Bacilli Seen Acid Fast Bacilli, Few Acid fast Bacilli, Rare Acid Fast Bacilli, Moderate Acid Fast Bacilli, Heavy
Acinetobacter Culture, MCU Only	ACINSC	No growth		Organisms growing in pure culture are identified to the species level whenever possible.	Surveillance Culture		Negative, Positive
Actinomyces, Screen	ACTBCN				Gram Stain		
Anaerobe Culture	ANACX	No growth			Culture Susceptibility (if appropriate)		
Bacterial Culture and Direct Smear, Lesion, Tissue, Device	General Cult Routine culture and smear	No growth			Smear Culture Susceptibility (if appropriate)	Validation	
Bacterial Vaginosis Panel	BVAG BV	Negative		This test was performed using Transcription Mediated Amplification for the detection of ribosomal RNA from bacteria associated with bacterial vaginosis.	Transcription-mediated amplification (TMA)	Package Insert Clinical Data	Negative, Positive
BAL Consult							BALQC: Acceptable, Unacceptable
Beta Strep, Vaginal Screen	GBSP Group B Streptococcus Testing by PCR	Negative		This test was performed using a real-time PCR assay. Results should be interpreted in conjunction with other clinical and laboratory findings.	Concentration in LIM broth for ~ 18 hours followed by real-time polymerase chain reaction (PCR) testing for Group B Streptococcus (GBS) DNA sequence	Validation	PCR: Negative, Positive
Beta Strep, Vaginal Screen, Reflex Susceptibility for Penicillin Allergy	GBS REFLEX GBS Streptococcus agalactiae	Negative		This test was performed using a real-time PCR assay. Results should be interpreted in conjunction with other clinical and laboratory findings.	Concentration in LIM broth for ~ 18 hours followed by real-time polymerase chain reaction (PCR) testing for Group B Streptococcus (GBS) DNA sequence	Validation	PCR: Negative, Positive
BK Virus DNA PCR, Quant, Urine	BKBU BKBP	<200 IU/mL (<2.30 Log IU/mL)		This test was performed using a real-time PCR assay. The dynamic range for this assay is 200-100,000,000 IU/mL. (2.30-8.00 Log IU/mL).	Real-time polymerase chain reaction (RT-PCR)	Package Insert	BKBU/T: 200-100,000,000 IU/mL. (2.30-8.00 Log IU/mL) BKBP/T: 200-100,000,000 IU/mL. (2.30-8.00 Log IU/mL)
BK Virus DNA Qn, PCR, Plasma	BKBP	<21.5 IU/mL (<1.33 Log IU/mL)		This test was performed using a real-time PCR assay. The dynamic range for this assay is 21.5-100,000,000 IU/mL. (1.33 to 8.00 Log IU/mL).	Real-time polymerase chain reaction (RT-PCR)	Package Insert	BKBP/P: 21.5-100,000,000.0 IU/mL BKVP/RALP: 1.33-8.00 Log IU/mL
Blood Culture	BLDCX	No growth			Culture Susceptibility (if appropriate)		
Blood Culture, AFB, Mycobacteria	AFBBCX Acid fast blood culture	Negative A final negative report will be issued after 42 days of incubation.		A positive result may support the diagnosis of mycobacteremia.	PCR (if appropriate) Susceptibility (if appropriate)		
Blood Culture, Pediatric	PEDI	No growth			Culture Susceptibility (if appropriate)		
Body Fluid Culture and Direct Smear	BFCX Sterile fluid culture	No growth			Smear Culture Susceptibility (if appropriate)		
C difficile 2 Step	CDSTEP	Negative			Polymerase chain reaction (PCR) Toxin A/B: Rapid Membrane Enzyme Immunoassay	Package Insert	C DIFFICILE BY PCR: Negative, Positive C DIFFICILE TOXIN A/B: Negative, Positive
Candida Auris Screen by PCR	CAAUPCR	Not Detected		This test was performed using a real-time PCR assay.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Literature	Not Detected, Detected
Candida/Trichomonas Panel	CVTV	Not Detected (for all targets)		This test was performed using Transcription Mediated Amplification for the detection of ribosomal RNA from microorganisms associated with vulvovaginal candidiasis and trichomoniasis.	Transcription-mediated amplification (TMA)	Package Insert Clinical Data	Not Detected, Detected
CAPD Fluid Bacterial Culture	BFCX	No growth			Gram Stain Culture Susceptibility (if appropriate)		
Chlamydia & Gonorrhea Amplified Probe	Chlam and Go Chlamydia trachomatis & Neisseria gonorrhoeae NAAT Testing	Not Detected (for all targets)			Transcription-mediated amplification (TMA)	Package Insert	Not Detected, Detected
Chlamydia Amplified Probe		Not Detected		This test was performed using Transcription Mediated Amplification for the detection of Chlamydia trachomatis and/or Neisseria gonorrhoeae nucleic acid. This assay is not intended for the evaluation of suspected sexual abuse or other medical indications.	Transcription-mediated amplification (TMA)	Package Insert	Not Detected, Detected
CMV by PCR, Quantitative, Blood	CMVPCR CMV Viral Load CMV PCR	<35.4 IU/mL (<1.54 Log IU/mL)		This test was performed using a real-time CMV PCR assay. The dynamic range for this assay is 34.5-10,000,000 IU/mL. (1.54-7.00 Log IU/mL). Results should be interpreted in conjunction with other clinical and laboratory findings.	Real-time polymerase chain reaction (RT-PCR)	Package Insert	-10,000,000 IU/mL. (1.54-7.00 Log IU/mL) 5-10,000,000 IU/mL. (1.54-7.00 Log IU/mL)
EBV by PCR, Quantitative, Blood	EBVPCR EBV Viral Load EBV PCR	<35 IU/mL (<1.54 Log IU/mL)		This test was performed using a real-time PCR assay. The dynamic range for this assay is 35 IU/mL-100,000,000 IU/mL. (1.54-8.00 Log IU/mL).	Real-time polymerase chain reaction (RT-PCR)	Package Insert	100,000,000 IU/mL. (1.54-8.00 Log IU/mL) 100,000,000 IU/mL. (1.54-8.00 Log IU/mL)
Fungus Culture	FUNCX	No growth		Positive cultures of yeast and filamentous fungi are reported with the organism identification.	Culture Susceptibility (if appropriate)		
Fungus Culture (Skin, Hair, Nails)	SINXCX	No growth		Positive cultures of yeast and filamentous fungi are reported with the organism identification.	Culture		
Fungus Smear	CALCOFLUOR Calcofluor White fluorescent stain				Smear		
Genital Culture, Bacterial	GENICX Vaginal Culture Cervical Culture Urethral Culture	No growth		Isolation of anaerobes in significant numbers from well-collected specimens indicates infection with the identified organism.	Culture Gram Stain (if appropriate) Susceptibility (if appropriate)		
Gonorrhea Amplified Probe	NG Chlamydia trachomatis & Neisseria gonorrhoeae NAAT Testing	Not Detected		This test was performed using Transcription Mediated Amplification for the detection of Chlamydia trachomatis and/or Neisseria gonorrhoeae nucleic acid. This assay is not intended for the evaluation of suspected sexual abuse or other medical indications.	Transcription-mediated amplification (TMA)	Package Insert	Not Detected, Detected

H. Pylori Urea Breath Test	UBTB UBT for H. pylori	Negative		Infrared Spectrophotometry	Package Insert	Negative, Positive
Hepatitis B DNA	HBDNA HBV Viral Load	HBV QUANT <10 IU/mL	This test was performed using a real time HBV PCR assay. The dynamic range for this assay is 10-1,000,000,000 IU/mL (1.00-9.00 Logs). Results should be interpreted in conjunction with other clinical and laboratory findings.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Validation	HBDNA: ≥10 IU/mL HBVNA1: 1.00-9.00 IU/mL
Hepatitis C by PCR, Quant	HCV QCP HCV Viral Load	HBV QUANT LOG HCV QUANT ≤12 IU/mL	This test was performed using a real time PCR assay. The dynamic range for this assay is 12-100,000,000 IU/mL.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Validation	HEPCQ: 12-100,000,000 IU/mL HEPCQ-1: 1.08-8.00 IU/mL
HIV Viral Load RNA PCR Quant	HVRNAB HIV Viral Load	HIV QUANT ≤40 copies/mL HIV QUANT LOG ≤1.60 copies/mL	This test was performed using the Abbott Real Time HIV-1 PCR assay. The dynamic range for this assay is 40-10,000,000 copies/mL (1.60-7.00 Logs). This test is designed to provide information for monitoring disease progression and is not intended for use in the diagnosis of HIV-1 infection or Acquired Immune Deficiency Syndrome (AIDS).	Real-time polymerase chain reaction (RT-PCR)	Package Insert Validation	HVRNA: 40-10,000,000 copies/mL HVRNAL: 1.60-7.00 copies/mL
HSV by PCR, Fluid/Lesion	HSVPCR Herpes Simplex Virus 1 and 2 Testing by PCR	Not Detected (for all targets)	This test was performed using a real time PCR assay.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Literature	HSV1: Not Detected, Detected HSV2: Not Detected, Detected
Immunocompromised Respiratory Panel	KRISP BioFire RP2.1 ATYPNE Atypical	Not Detected (for all targets)	NP Results should be used in conjunction with other clinical and laboratory findings. This result does not rule out co-infections with pathogens that are not screened for by the Respiratory Panel (RP). This RP assay was performed using a Film Array multiplex nucleic acid assay. BAL Results should be used in conjunction with other clinical and laboratory findings. This result does not rule out co-infections with pathogens that are not screened for by the Respiratory Panel (RP). This RP assay was performed using a Film Array multiplex nucleic acid assay. This test was developed and its performance characteristics determined by The Clinical Microbiology Laboratory at The Ohio State University Wexner Medical Center. It has not been cleared or approved by the FDA. The laboratory is required under CLIA as qualified to perform high-complexity testing. This test is used for clinical purposes. It should not be regarded as investigational or for research.	Multiplex polymerase chain reaction (PCR)	Package Insert Literature	All Targets: Not Detected, Detected
Influenza A/B Rapid Molecular	FLUABM Rapid Flu	Not Detected (for all targets)	This test utilizes isothermal nucleic acid amplification technology for the differential qualitative detection of influenza A and influenza B viral nucleic acids.	Isothermal nucleic acid amplification	Abbott ID NOW Influenza A & B 2 package insert	All Targets: Not Detected, Detected, Indeterminate
Influenza A/B, RSV By PCR	RVP-PCR Flu PCR RSV PCR	Not Detected (for all targets)	This test was performed using a multiplex real time PCR assay. This result does not rule out co-infections with pathogens that were not screened for by this test. Results should be used in conjunction with other clinical and laboratory findings.	Polymerase chain reaction (PCR)	Validation	All Targets: Not Detected, Detected, Indeterminate
Lactoferrin, Qualitative, Stool	WBCSTL Fecal Leukocytes Stool for WBC's	Negative		Immunochromatographic	Package Insert	Negative, Positive
Legionella Culture	LEGCCX	No growth	Identification of Legionella species from respiratory specimens provides a definitive diagnosis of Legionnaires disease.	Culture		
Lower Respiratory Culture, Bacterial	RESPCX Respiratory Culture	No growth		Gram Stain Culture Susceptibility (if appropriate)		
M Tuberculosis Complex by PCR	MTRCOM Tb PCR MTB	Not Detected	This test should always be performed in conjunction with AFB (acid fast bacilli) smear and culture. This test was performed using a Multiplex PCR Test.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Reference Materials	Not Detected, Detected
Macroscopic Arthropod	Macroscopic Arthropod	Negative, Artifact (not arthropod)		Macroscopic exam		Negative, Bedbug, Centipede, Flea, Flea, Tunga sp, Flea, Ctenocephalides sp, Flea, Xenopsylla sp, Flea, Pulex sp, Fly larva (myiasis causing), Kissing Bug (Reduviid, Triatomine), Louse, Louse, Nit, Louse, Pediculus humanus, Louse, Phthirus pubis, Mite, Demodex sp, Mite, not Scabies, Mite-Sarcoptes scabiei, Scorpion, Spider, Spider-Black Widow, Spider-Brown Recluse, Hand tick, Hand tick, not Ixodes, Soft Tick, Tick, Tick-bodies sp, Tick-Dermacentor sp, Tick-Amblyomma sp, Tick-Rhipicephalus sp, Artifact (not arthropod, Arthropod, Arthropod or worm, other than a human parasite, Arthropod not known to transmit human pathogens.
Maternal High Risk Sta HIV	MHRHIV		RHV12: Non Reactive RPDP24: Non Reactive	Qualitative immunoassay / immunochromatographic test for simultaneous and qualitative detection of free HIV-1 p24 antigen and antibodies to HIV-1 and HIV-2.	Package Insert	RHV12: Non Reactive, Reactive RPDP24: Non Reactive, Presumptive Reactive, Reactive
Meningitis/Encephalitis Panel CSF	CSFMFP BioFire ME Panel	Not Detected (for all targets)	A negative result does not exclude the possibility of CNS infection and should not be used as the sole basis for diagnosis, treatment, or other management decisions. Negative results may occur when the concentration of organisms (i.e., viruses), or yeast in the specimen is below the limit of detection. The ME panel does not distinguish between latent and active herpesvirus infections (CMV, HHV-6). This test was performed using a film array method for the detection of: <i>Escherichia coli</i> , K1, <i>Haemophilus influenzae</i> , <i>Listeria monocytogenes</i> , <i>Neisseria meningitidis</i> , <i>Serpentinococcus agalactiae</i> , <i>Serpentinococcus pneumoniae</i> , Cytomegalovirus, Enterovirus, Herpes Simplex virus 1 and 2, Human Herpesvirus 6, Human parvovirus, Varicella zoster virus, and <i>Cryptosporidium parvum</i> oocysts.	Multiplex polymerase chain reaction (PCR)	Package Insert	All Targets: Not Detected, Detected
Molecular Enteric Panel, Stool	STLB Gastrointestinal Panel	Negative (for all targets)	This test was performed using a real time PCR assay. Results should be interpreted in conjunction with clinical findings. A positive result does not necessarily indicate the presence of viable organisms. Positive results do not rule out co-infection with other organisms that are not detected by this assay. For Shiga toxin, this assay detects Shiga toxin 1/ Shiga toxin 2 genes (found in Shiga toxin-producing E. coli as well as Shigella dysenteriae). This test should not be used as a test of cure.	Real-time polymerase chain reaction (RT-PCR)	Package Insert	All Targets: Negative, Positive
Molecular Stool Parasite Panel	SCPPI OAP	Negative (for all targets)	This test was performed using a real time PCR assay. Results should be interpreted in conjunction with clinical findings. A positive result does not necessarily indicate the presence of viable organisms. Positive results do not rule out co-infection with other organisms that are not detected by this assay. This test should not be used as a test of cure.	Real-time polymerase chain reaction (RT-PCR)	Package Insert Reference Materials	All Targets: Negative, Positive
Neisseria-Gonococcus Screen	SCN	No growth		Culture		
Novel Coronavirus PCR - Semi-Private Surveillance	COVID	Not Detected	Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for treatment or other patient management decisions. Optimum specimen types and timing for peak viral levels during infections caused by SARS-CoV-2 has not been determined. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation suggest that SARS-CoV-2 infection is probable, and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative. Collection of a new specimen and re-testing may be necessary if the patient is critically ill or clinically deteriorating.	Real-time polymerase chain reaction (RT-PCR)	Package Insert	Not Detected, Detected

Novel Coronavirus PCR - Nasopharyngeal		Not Detected	Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for treatment or other patient management decisions. Optimum specimen types and timing for peak viral levels during infections caused by SARS-CoV-2 has not been determined. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation suggest that SARS-CoV-2 infection is probable, and diagnostic tests for other causes of illness (e.g. other respiratory illness) are negative. Collection of a new specimen and re-testing may be necessary if the patient is critically ill or clinically deteriorating.	Real-time polymerase chain reaction (RT-PCR)	Package Insert	Not Detected, Detected
	COVID					
Pinworm Exam	Pinworm Exam	Negative		Microscopic exam		Negative, Enterobius vermicularis (Pinworm) Eggs Seen, Enterobius vermicularis (Pinworm) Adult Worm(s) Seen, Enterobius vermicularis (Pinworm) adult worm(s) and eggs seen
Pleiomonas/Aeromonas Screen, Stool	SCT	Negative for Aeromonas species Negative for Pleiomonas species		Culture		
Quantitative Tissue Culture	BKR QUANTITA	No growth		Susceptibility (if appropriate) Tissue is weighed, serially diluted, and cultured for exact colony count		
Rapid HIV-1/HIV-2 Ab With P24 Antigen	RHHV12		RHHV12: Non Reactive RFPDP24: Non Reactive	Qualitative immunoassay/immunochromatographic test for simultaneous and qualitative detection of free HIV-1 p24 antigen and antibodies to HIV-1 and HIV-2	Package Insert	RHHV12: Non Reactive, Reactive RFPDP24: Non Reactive, Presumptive Reactive, Reactive
Rapid Malaria	MPBR EVDMAL Plingpatium	Negative for malaria antigens	Infection due to malaria cannot be entirely ruled out. Malaria antigen in the sample may be below the detection limit of the test. Results should be interpreted in conjunction with the thick and thin malaria preparation/microscopy.	Immunochromatographic membrane assay	Package Insert	Negative for malaria antigens. Positive rapid malaria test for P. falciparum protein antigen only. Positive rapid malaria test for malaria protein antigen, representing P. vivax or P. malariae or P. ovale or a mix of species. Positive rapid malaria test. Positive for P. falciparum protein antigen
Rapid Strep A, Molecular	STRPAM Rapid Strep A	Negative		Molecular in vitro diagnostic test utilizing isothermal nucleic acid amplification.	Abbott ID NOW Strep A 2 package insert	Negative, Positive
Rectal Screening for Cipro Resistance	BKR CIPROFLO Ciprofloxacin Resistance Screening	Negative results do not preclude infection with Group A Streptococcus and should not be used as the sole basis for treatment. Negative for Ciprofloxacin Resistant Enterobacteriales		Culture Susceptibility (if appropriate)		
SARS-COV-2 Rapid		Not Detected	Negative results do not preclude SARS-CoV-2 infection and should not be used as the sole basis for treatment or other patient management decisions. Optimum specimen types and timing for peak viral levels during infections caused by SARS-CoV-2 has not been determined. The possibility of a false negative result should especially be considered if the patient's recent exposures or clinical presentation suggest that SARS-CoV-2 infection is probable, and diagnostic tests for other causes of illness (e.g. other respiratory illness) are negative. Collection of a new specimen and re-testing may be necessary if the patient is critically ill or clinically deteriorating.	Isothermal nucleic acid amplification	Abbott ID NOW Rapid Covid package insert	Not Detected, Detected
	COVID Rapid COVID					
Screen: MRSA Only	SCREEN:MRSA MRSA Screen	Negative		Culture on select agar		Negative, Positive
Screen: MRSA/MSSA	SCRSB Respiratory Staphylococcus Screen	PCR Negative	This test was performed using a real time PCR assay. Results should be interpreted in conjunction with other clinical and laboratory findings. A positive result does not necessarily indicate the presence of viable organism. This test should not be used as a test of cure.	Real-time polymerase chain reaction (RT-PCR)	Package Insert	Staphylococcus aureus by PCR: Negative, Positive Methicillin Resistant S. aureus by PCR: Negative, Positive
Screen: VRE	SCREEN:VRE Vancomycin Resistant Enterococcus Screen	Negative		Culture on select agar		Negative, Positive
Screen: Yeast	STRSCB	Negative		Culture		Negative, Positive
Strep Pneumoniae: Antigen, Urine	PNELUMU	Negative		Immunochromatographic membrane assay	Binax NOW Package Insert	Negative, Positive
Upper Respiratory Culture, Bacterial	THRCX Throat Culture	No growth		Culture		
Urine Culture	RESN URN CX	Negative		Culture Susceptibility (if appropriate)		
Vaginal Infection Panel	BVAG CV TV CT NG	BVAG: Negative CV: Not Detected TV: Not Detected CT: Not Detected NG: Not Detected	BV: This test was performed using transcription mediated amplification for the detection of ribosomal RNA from bacteria associated with bacterial vaginosis. CV/TV: This test was performed using Transcription Mediated Amplification for the detection of ribosomal RNA from microorganisms associated with vulvovaginal candidiasis and trichomoniasis. CT/NG: This test was performed using Transcription Mediated Amplification for the detection of Chlamydia trachomatis and/or Neisseria gonorrhoeae nucleic acid. This assay is not intended for the evaluation of suspected sexual abuse or other medical-legal indications.	Transcription Mediated Amplification	See individual analytes	See individual analytes
Varicella Zoster By PCR, Skin	VZVPCR	Not Detected		Real-time polymerase chain reaction (RT-PCR)	Package Insert Literature	Not Detected, Detected
ADAMTS13 Activity and IgG Ab w/ Reflex to Inhibitor	ADISA ADI 3KG von Willebrand Factor Cleaving Protease Anti-ADAMTS13 antibodies	ACTIVITY No established reference range IGG		Technozym enzyme-linked immunosorbent assay (ELISA)	Technozym kit	ACTIVITY: 2-100 % IGG: 6.0-104.0 U/mL
Alternative Activation Pathway	APBBD Bc Complement Alternative complement pathway	695-1,974 ng/mL		Quidel enzyme-linked immunosorbent assay (ELISA)	Biomarker Reference Lab	
Terminal Activation Pathway	TPCSAB SC5b-9 Complement Membrane attack complex (MAC) Terminal complement complex (TCC)	6-598 ng/mL		Quidel enzyme-linked immunosorbent assay (ELISA)	Biomarker Reference Lab	